
Verify®
Accurato
Facile da usare
Risultato rapido
Self Test Multi-droga (Urine)
Oppiacei, Metilendiossimetamfetamina(MDMA),Cocaina, Anfetamina, Metanfetamina, Ketamina, Marijuana
contiene 1 test
RISULTATO IN 5 MINUTI

EC Declaration of Conformity

Manufacturer:
Name: HANGZHOU ALLTEST BIOTECH CO., LTD.
Address: #550, Yinhai Street, Hangzhou Economic & Technological Development Area, Hangzhou -310018, P.R. China

European Representative:
Name: VidaQuick Biotech S.L.
Address: No.132, Rosello Street, Barcelona, Barcelona Province, 08036, Spain

Product Name: Multi-Drug Drugs Rapid Test Panel (Urine)
Cat. No.: DOA-174H

We, HANGZHOU ALLTEST BIOTECH CO., LTD., herewith declare that we are exclusively responsible for this declaration of conformity. We herewith declare that the above mentioned product which is not **intended for IVD use** meets the transposition into national law, the provisions of the following GPSR (EU) 2023/988, Regulation (EC) No 1907/2006 and Regulation (EC) No 1272/200.
All supporting documentations are retained under the premises of the manufacturer.

DIRECTIVES AND REGULATIONS

General applicable directives and regulations Applied:
CLP REGULATION (EC) No 1272/2008 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 16 December 2008
REACH REGULATION (EC) No 1907/2006 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 18 December 2006
GPSR (EU) 2023/988 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 10 May 2023

Standard Applied: EN ISO 13485:2016, EN ISO 14971:2012, EN 13975:2003, EN ISO 18113-1:2011, EN ISO 18113-4:2011, EN 13612:2002/AC:2002, EN ISO 23640:2015, EN ISO 15223-1:2016

Place, Date of issue of DOC: in Hangzhou on 03/09/2025

Signature:_____
Name: Gao Fei (Position: General Manager)

Hangzhou AllTest Biotech Co.,Ltd.
#550, Yinhai Street,
Hangzhou Economic & Technological Development Area,
Hangzhou -310018, P.R. China

杭州奥泰生物技术股份有限公司
地址：杭州市经济技术开发区
银海街550号
邮编：310018

TEL：+86 571 56267891
EMAIL：info@alltests.com.cn
http:// www.alltests.com.cn

电话：+86 571 56267891
邮箱：info@alltests.com.cn
网址：www.alltests.com.cn

Registration File of
Multi-Drug Rapid Test Panel

HANGZHOU ALLTEST BIOTECH CO., LTD

Address: # 550, Yinhai Street, Hangzhou Economic & Technological Development Area,

Hangzhou-310018, P.R. China

Table of Contents

1 Background 3

1.1 Product Name 3
1.2 Intended Use 3
1.3 Background 5
1.4 Principles 20
1.5 Performance and Specification 20
1.6 Composition of Product 22
1.7 Manufacturing Procedure 23
1.8 User 23
1.9 Classification of Product 23
1.10 Product Certification Conformity Assessment Route 23
1.11 Bibliography 23

2 Performance characteristics 24

2.1 Sample correlation 24
2.2 Analytical Sensitivity 29
2.3 Analytical Specificity 52
2.4 Cross Reactivity 136
2.5 Between Day Reproducibility 137
2.6 Between Lot Reproducibility (lot to lot variability) 167
2.7 Accelerated Stability 197
2.8 Precision 355
2.9 Effect of Urinary Specific Gravity 386
2.10 Effect of Urinary pH 436
2.11 Open Pouch Stability 469
Document History Summary 552

1 Background

1.1.1 Product Name

Multi-Drug Rapid Test Panel

1.1.2 Intended Use

The Multi-Drug Rapid Test Panel is a rapid chromatographic immunoassay for the detection of drugs and drug metabolites in urine at the following cut-off concentrations in urine:

Test	**Calibrator**	**Cut-off (ng/mL)**
Amphetamine (AMP 1000)	d-Amphetamine	1,000
Amphetamine (AMP 500)	d-Amphetamine	500
Amphetamine (AMP 300)	d-Amphetamine	300
Barbiturates (BAR 300)	Secobarbital	300
Barbiturates (BAR 200)	Secobarbital	200
Benzodiazepines (BZO 500)	Oxazepam	500
Benzodiazepines (BZO 300)	Oxazepam	300
Benzodiazepines (BZO 200)	Oxazepam	200
Benzodiazepines (BZO 100)	Oxazepam	100
Buprenorphine (BUP 10)	Buprenorphine	10
Buprenorphine (BUP 5)	Buprenorphine	5
Cocaine (COC 300)	Benzoylecgonine	300
Cocaine (COC200)	Benzoylecgonine	200
Cocaine (COC 100)	Benzoylecgonine	100
Cocaine (COC 150)	Benzoylecgonine	150
Marijuana (THC300)	11-nor-Δ9pHC-9 COOH	300
Marijuana (THC 200)	11-nor-Δ9pHC-9 COOH	200
Marijuana (THC 30)	11-nor-Δ9pHC-9 COOH	30
Marijuana (THC 20)	11-nor-Δ9pHC-9 COOH	20
Marijuana (THC150)	11-nor-Δ9pHC-9 COOH	150
Marijuana (THC 50)	11-nor-Δ9pHC-9 COOH	50
Marijuana (THC 25)	11-nor-Δ9pHC-9 COOH	25
Methadone (MTD 300)	Methadone	300
Methadone (MTD 200)	Methadone	200
Methamphetamine (MET 1,000)	d-Methamphetamine	1,000
Methamphetamine (MET 500)	d-Methamphetamine	500
Methamphetamine (MET 300)	d-Methamphetamine	300
Methylenedioxymethamphetamine (MDMA 300)	d,l-Methylenedioxymethamphetamine	300
Methylenedioxymethamphetamine (MDMA 500)	d,l-Methylenedioxymethamphetamine	500
Methylenedioxymethamphetamine (MDMA 1,000)	d,l-Methylenedioxymethamphetamine	1,000
Morphine (MOP 300)	Morphine	300
Morphine (MOP 200)	Morphine	200
Morphine (MOP 100)	Morphine	100
Opiate (OPI 2000)	Morphine	2,000
Methaqualone(MQL300)	Methaqualone	300
Phencyclidine (PCP25)	Phencyclidine	25
Propoxyphene (PPX300)	Propoxyphene	300

Tricyclic Antidepressants (TCA1000)	Nortriptyline	1,000
Tricyclic Antidepressants (TCA500)	Nortriptyline	500
Tramadol (TML 100)	Cis-Tramadol	100
Tramadol (TML 200)	Cis-Tramadol	200
Tramadol (TML300)	Cis-Tramadol	300
Ketamine (KET 1,000)	Ketamine	1,000
Ketamine (KET 500)	Ketamine	500
Ketamine (KET 300)	Ketamine	300
Ketamine (KET 100)	Ketamine	100
Oxycodone (OXY100)	Oxycodone	100
Oxycodone (OXY300)	Oxycodone	300
Cotinine(COT 500)	Cotinine	500
Cotinine(COT 200)	Cotinine	200
Cotinine(COT 100)	Cotinine	100
Cotinine(COT 50)	Cotinine	50
Cotinine(COT 10)	Cotinine	10
2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP 300)	2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine	300
2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP 100)	2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine	100
Fentanyl(FYL200)	Fentanyl	200
Fentanyl(FYL100)	Fentanyl	100
Fentanyl(FYL20)	Fentanyl	20
Fentanyl(FYL10)	Fentanyl	10
Synthetic Marijuana(K2 50)	JWH -018, JWH- 073	50
Synthetic Marijuana(K2 30)	JWH -018, JWH- 073	30
Synthetic Marijuana(K2 25)	JWH -018, JWH- 073	25
6-mono-aceto-morphine (6-MAM10)	6-mono-aceto-morphine	10
Clonazepam (CLO400)	Clonazepam	400
Clonazepam (CLO150)	Clonazepam	150
Diazepam(DIA300)	Diazepam	300
Diazepam(DIA200)	Diazepam	200
Ethyl Glucuronide(ETG300)	Ethyl Glucuronide	300
Ethyl Glucuronide(ETG500)	Ethyl Glucuronide	500
Ethyl Glucuronide(ETG1000)	Ethyl Glucuronide	1000
Acetaminophen (ACE5000)	Acetaminophen	5000
Lysergic Acid Diethylamide (LSD10)	Lysergic Acid Diethylamide	10
Lysergic Acid Diethylamide (LSD20)	Lysergic Acid Diethylamide	20
Lysergic Acid Diethylamide (LSD50)	Lysergic Acid Diethylamide	50
3,4-Methylenedioxyamphetamine (MDA500)	3,4-Methylenedioxyamphetamine	500
3, 4-methylenedioxypyrovalerone (MDPV1000)	3,4-methylenedioxypyrovalerone	1000
3, 4-methylenedioxypyrovalerone (MDPV500)	3,4-methylenedioxypyrovalerone	500
3, 4-methylenedioxypyrovalerone (MDPV300)	3,4-methylenedioxypyrovalerone	300
Methylphenidate (MPD300)	Methylphenidate	300
Methylphenidate (MPD1000)	Methylphenidate	1000
Zolpidem (ZOL50)	Zolpidem	50
Zopiclone(ZOP50)	Zopiclone	50

Methcathinone(MCAT500)	S(-)-Methcathinone	500
7-Aminoclonazepam(7-ACL300)	7-Aminoclonazepam	300
7-Aminoclonazepam(7-ACL200)	7-Aminoclonazepam	200
7-Aminoclonazepam(7-ACL100)	7-Aminoclonazepam	100
Carfentanyl(CFYL500)	Carfentanyl	500
Carfentanyl(CFYL250)	Carfentanyl	250
Caffeine(CAF1000)	Caffeine	1000
Cathine (CAT150)	(+)-Norpseudoephedrine	150
Tropicamide(TRO350)	Tropicamide	350
Mephedrone(MEP100)	Mephedrone	100
Alprazolam(ALP100)	Alprazolam	100
AB-PINACA (ABP10)	AB-PINACA	10
α-Pyrrolidinovalerophenone (α-PVP1000)	α-Pyrrolidinovalerophenone	1,000
α-Pyrrolidinovalerophenone (α-PVP500)	α-Pyrrolidinovalerophenone	500
α-Pyrrolidinovalerophenone (α-PVP300)	α-Pyrrolidinovalerophenone	300
Cannabinol (CNB500)	Cannabinol	500
Meperidine (MPRD100)	Meperidine	100
Pregabalin (PGB50000)	Pregabalin	50,000
Pregabalin (PGB500)	Pregabalin	500
Trazodone(TZD200)	Trazodone	200
UR-144 25	UR-144 5-Pentanoic acid	25
Zaleplon(ZAL100)	Zaleplon	100
Mescaline (MES300)	Mescaline	300
Mescaline (MES100)	Mescaline	100
Gabapentin(GAB2000)	Gabapentin	2,000
Tilidine(TLD50)	Nortilidine	50
Quetiapine(QTP1000)	Quetiapine	1,000
Papaverine(PAP500)	Papaverine	500
Kratom(KRA300)	Mitragynine	300
Carisoprodol(CAR2000)	Carisoprodol	2,000
Fluoxetine(FLX500)	Fluoxetine	500
Risperidone (RPD150)	Risperidone	150
Citalopram (CIT500)	Desmethylcitalopram	500
F-Ketamine (FKET1000)	F-Ketamine	1000
Tapentadol (TAP1000)	Tapentadol	1000
Scopolamine (SCOP500)	Scopolamine	500
Mirtazapine (MTZ500)	Mirtazapine	500
N,N-Dimethyltryptamine (NND1000)	N,N-Dimethyltryptamine	1000

Configurations of the Multi-Drug Rapid Test Panel come with any combination of the above listed drug analytes**.** This assay provides only a preliminary analytical test result. A more specific alternate chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography/mass spectrometry (GC/MS) is the preferred confirmatory method. Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are indicated.

1.1.3 Background

The Multi-Drug Rapid Test Panel is a rapid urine screening test that can be performed without the use of an instrument. The test utilizes monoclonal antibodies to selectively detect elevated levels of specific

drugs in urine.

1.3.1 Amphetamine (AMP)

Amphetamine is a Schedule II controlled substance available by prescription (Dexedrine®) and is also available on the illicit market. Amphetamines are a class of potent sympathomimetic agents with therapeutic applications. They are chemically related to the human body's natural catecholamines: epinephrine and norepinephrine. Acute higher doses lead to enhanced stimulation of the central nervous system (CNS) and induce euphoria, alertness, reduced appetite, and a sense of increased energy and power. Cardiovascular responses to amphetamines include increased blood pressure and cardiac arrhythmias. More acute responses produce anxiety, paranoia, hallucinations, and psychotic behavior. The effects of Amphetamines generally last 2-4 hours following use and the drug has a half-life of 4-24 hours in the body. About 30% of amphetamines are excreted in the urine in unchanged form, with the remainder as hydroxylated and deaminated derivatives.

1.3.2 Barbiturates (BAR)

Barbiturates are CNS depressants. They are used therapeutically as sedatives, hypnotics, and anticonvulsants barbiturates are almost always taken orally as capsules or tablets. The effects resemble those of intoxication with alcohol. Chronic use of barbiturates leads to tolerance and physical dependence.

Short-acting barbiturates taken at 400 mg/day for 2-3 months can produce a clinically significant degree of physical dependence. Withdrawal symptoms experienced during periods of drug abstinence can be severe enough to cause death.

Only a small amount (less than 5%) of most barbiturates are excreted unaltered in the urine.

The approximate detection time limits for barbiturates are:

Short acting (e.g. Secobarbital)	100 mg PO (oral)	4.5 days
Long acting (e.g. Phenobarbital)	400 mg PO (oral)	7 days

1.3.3 Benzodiazepines (BZO)

Benzodiazepines are medications that are frequently prescribed for the symptomatic treatment of anxiety and sleep disorders. They produce their effects via specific receptors involving a neurochemical called gamma aminobutyric acid (GABA). Because they are safer and more effective, Benzodiazepines have replaced Barbiturates in the treatment of both anxiety and insomnia. Benzodiazepines are also used as sedatives before some surgical and medical procedures, and for the treatment of seizure disorders and alcohol withdrawal.

Risk of physical dependence increases if Benzodiazepines are taken regularly (e.g., daily) for more than a few months, especially at higher than normal doses. Stopping abruptly can bring on such symptoms as trouble sleeping, gastrointestinal upset, feeling unwell, loss of appetite, sweating, trembling, weakness, anxiety and changes in perception. Only trace amounts (less than 1%) of most Benzodiazepines are excreted unaltered in the urine; most of the concentration in urine is conjugated drug. The detection period for the Benzodiazepines in the urine is 3-7 days.

1.3.4 Buprenorphine (BUP)

Buprenorphine is a potent analgesic often used in the treatment of opioid addiction. The drug is sold under the trade names Subutex™, Buprenex™, Temgesic™ and Suboxone™, which contain Buprenorphine HCl alone or in combination with Naloxone HCl. Therapeutically, Buprenorphine is used as a substitution treatment for opioid addicts. Substitution treatment is a form of medical care offered to opiate addicts (primarily heroin addicts) based on a similar or identical substance to the drug normally used. In substitution therapy, Buprenorphine is as effective as Methadone but demonstrates a lower level of physical dependence. Concentrations of free Buprenorphine and Norbuprenorphine in urine may be less than 1 ng/ml after therapeutic administration, but can range up to 20 ng/ml in abuse situations.10 The plasma half life of Buprenorphine is 2-4 hours.10 While complete elimination of a single dose of the drug can take as long as 6 days, the window of detection for the parent drug in urine is thought to be approximately 3 days.

Substantial abuse of Buprenorphine has also been reported in many countries where various forms of the drug are available. The drug has been diverted from legitimate channels through theft, doctor shopping, and fraudulent prescriptions, and been abused via intravenous, sublingual, intranasal and inhalation routes.

1.3.5 Cocaine (COC)

Cocaine is a potent central nervous system stimulant and a local anesthetic. Initially, it brings about extreme energy and restlessness while gradually resulting in tremors, over-sensitivity and spasms. In large amounts, cocaine causes fever, unresponsiveness, difficulty in breathing and unconsciousness.

Cocaine is often self-administered by nasal inhalation, intravenous injection and free-base smoking. It is excreted in the urine in a short time primarily as benzoylecgonine.3,4 Benzoylecgonine, a major metabolite of cocaine, has a longer biological half-life (5-8 hours) than cocaine (0.5-1.5 hours), and can generally be detected for 24-48 hours after cocaine exposure.

1.3.6 Marijuana (THC)

THC (∆9petrahydrocannabinol) is the primary active ingredient in cannabis (marijuana). When smoked or orally administered, THC produces euphoric effects. Users have impaired shortperm memory and slowed learning. They may also experience transient episodes of confusion and anxiety. Longperm, relatively heavy use may be associated with behavioral disorders. The peak effect of marijuana administered by smoking occurs in 20-30 minutes and the duration is 90-120 minutes after one cigarette. Elevated levels of urinary metabolites are found within hours of exposure and remain detectable for 3-10 days after smoking. The main metabolite excreted in the urine is 11-nor-∆9petrahydrocannabinol-9-carboxylic acid (THC-COOH).

1.3.7 Methadone (MTD)

Methadone is a narcotic analgesic prescribed for the management of moderate to severe pain and for the treatment of opiate dependence (heroin, Vicodin, Percocet, morphine). The pharmacology of oral methadone is very different from IV methadone. Oral methadone is partially stored in the liver for later use. IV methadone acts more like heroin. In most states you must go to a pain clinic or a methadone maintenance clinic to be prescribed methadone.

Methadone is a long acting pain reliever producing effects that last from twelve to forty-eight hours. Ideally, methadone frees the client from the pressures of obtaining illegal heroin, from the dangers of injection, and from the emotional roller coaster that most opiates produce. Methadone, if taken for long periods and at large doses, can lead to a very long withdrawal period. The withdrawals from methadone are more prolonged and troublesome than those provoked by heroin cessation, yet the substitution and phased removal of methadone is an acceptable method of detoxification for patients and therapists.7

1.3.8 Methamphetamine (MET)

Methamphetamine is an addictive stimulant drug that strongly activates certain systems in the brain. Methamphetamine is closely related chemically to Amphetamine, but the central nervous system effects of Methamphetamine are greater. Methamphetamine is made in illegal laboratories and has a high potential for abuse and dependence. The drug can be taken orally, injected, or inhaled. Acute higher doses lead to enhanced stimulation of the central nervous system and induce euphoria, alertness, reduced appetite, and a sense of increased energy and power. Cardiovascular responses to Methamphetamine include increased blood pressure and cardiac arrhythmias. More acute responses produce anxiety, paranoia, hallucinations, psychotic behavior, and eventually, depression and exhaustion. The effects of Methamphetamine generally last 2-4 hours and the drug have a half-life of 9-24 hours in the body. Methamphetamine is excreted in the urine primarily as Amphetamine, and oxidized and deaminated derivatives. However, 10-20% of Methamphetamine is excreted unchanged. Thus, the presence of the parent compound in the urine indicates Methamphetamine use. Methamphetamine is generally detectable in the urine for 3-5 days, depending on urine pH level.

1.3.9 Methylenedioxymethamphetamine (MDMA)

Methylenedioxymethamphetamine (ecstasy) is a designer drug first synthesized in 1914 by a German drug company for the treatment of obesity.5 Those who take the drug frequently report adverse effects, such as increased muscle tension and sweating. MDMA is not clearly a stimulant, although it has, in common with amphetamine drugs, a capacity to increase blood pressure and heart rate. MDMA does produce some perceptual changes in the form of increased sensitivity to light, difficulty in focusing, and blurred vision in some users. Its mechanism of action is thought to be via release of the neurotransmitter serotonin. MDMA may also release dopamine, although the general opinion is that this is a secondary effect of the drug (Nichols and Oberlender, 1990). The most pervasive effect of MDMA, occurring in virtually all people who took a reasonable dose of the drug, was to produce a clenching of the jaws.

1.3.10 Morphine/Opiate (MOP/ OPI)

Morphine/Opiate refers to any drug that is derived from the opium poppy, including the natural products, morphine and codeine, and the semi-synthetic drugs such as heroin. Morphine/Opiate is more general, referring to any drug that acts on the opioid receptor.

Opioid analgesics comprise a large group of substances which control pain by depressing the CNS. Large doses of morphine can produce higher tolerance levels, physiological dependency in users, and may lead to substance abuse. Morphine is excreted unmetabolized, and is also the major metabolic product of codeine and heroin. Morphine is detectable in the urine for several days after an opiate dose.

1.3.11 Phencyclidine (PCP)

Phencyclidine, also known as PCP or Angel Dust, is a hallucinogen that was first marketed as a surgical anesthetic in the 1950's. It was removed from the market because patients receiving it became delirious and experienced hallucinations.

PCP is used in powder, capsule, and tablet form. The powder is either snorted or smoked after mixing it with marijuana or vegetable matter. PCP is most commonly administered by inhalation but can be used intravenously, intra-nasally, and orally. After low doses, the user thinks and acts swiftly and experiences mood swings from euphoria to depression. Self-injurious behavior is one of the devastating effects of PCP.

PCP can be found in urine within 4 to 6 hours after use and will remain in urine for 7 to 14 days, depending on factors such as metabolic rate, user's age, weight, activity, and diet.6 PCP is excreted in the urine as an unchanged drug (4% to 19%) and conjugated metabolites (25% to 30%).

1.3.12 Propoxyphene (PPX)

Propoxyphene (PPX) is a narcotic analgesic compound bearing structural similarity to methadone. As an analgesic, propoxyphene can be from 50-75% as potent as oral codeine. Darvocet™, one of the most common brand names for the drug, contains 50-100 mg of propoxyphene napsylate and 325-650 mg of acetaminophen. Peak plasma concentrations of propoxyphene are achieved from 1 to 2 hours post dose. In the case of overdose, propoxyphene blood concentrations can reach significantly higher levels.

In humans, propoxyphene is metabolized by N-demethylation to yield norpropoxyphene. Norpropoxyphene has a longer half-life (30 to 36 hours) than parent propoxyphene (6 to 12 hours). The accumulation of norpropoxyphene seen with repeated doses may be largely responsible for resultant toxicity.

1.3.13 Tricyclic Antidepressants (TCA)

TCA (Tricyclic Antidepressants) are commonly used for the treatment of depressive disorders. TCA overdoses can result in profound CNS depression, cardiotoxicity and anticholinergic effects. TCA overdose is the most common cause of death from prescription drugs. TCAs are taken orally or sometimes by injection. TCAs are metabolized in the liver. Both TCAs and their metabolites are excreted in urine mostly in the form of metabolites for up to ten days.

1.3.14 Tramadol (TML)

Tramadol(TML) is a quasi-narcotic analgesic used in the treatment of moderate to severe pain. It is a synthetic analog of codeine, but has a low binding affinity to the mu-opioid receptors. Large doses of tramadol can develop tolerance and physiological dependency and lead to its abuse. Tramadol is

extensively metabolized after oral administration. Approximately 30% of the dose is excreted in the urine as unchanged drug, whereas 60% is excreted as metabolites. The major pathways appear to be N- and O-demethylation, glucoronidation or sulfation in the liver.

1.3.15 Ketamine (KET)

Ketamine is a dissociative anesthetic developed in 1963 to replace PCP (Phencyclidine). While Ketamine is still used in human anesthesia and veterinary medicine, it is becoming increasingly abused as a street drug. Ketamine is molecularly similar to PCP and thus creates similar effects including numbness, loss of coordination, sense of invulnerability, muscle rigidity, aggressive / violent behavior, slurred or blocked speech, exaggerated sense of strength, and a blank stare. There is depression of respiratory function but not of the central nervous system, and cardiovascular function is maintained. The effects of Ketamine generally last 4-6 hours following use. Ketamine is excreted in the urine as unchanged drug (2.3%) and metabolites (96.8%).

1.3.16 Methaqualone (MQL)

Methaqualone (Quaalude,Sopor) is a quinazoline derivative that was first synthesized in 1951 and found clinically effective as a sedative and hypnotic in 1956.[10] It soon gained popularity as a drug of abuse and in 1984 was removed from the US market due to extensive misuse. It is occasionally encountered in illicit form, and is also available in Europeon countries in combination with diphenhydramine (Mandrax). Methaqualone is extensively metabolized in vivo principally by hydroxylation at every possible position on the molecule. At least 12 metabolites have been identified in the urine.

1.3.17 Oxycodone (OXY)

Oxycodone is a semi-synthetic opioid with a structural similarity to codeine. The drug is manufactured by modifying thebaine, an alkaloid found in the opium poppy. Oxycodone, like all opiate agonists, provides pain relief by acting on opioid receptors in the spinal cord, brain, and possibly directly in the affected tissues. Oxycodone is prescribed for the relief of moderate to high pain under the well-known pharmaceutical trade names of OxyContin®, Tylox®, Percodan® and Percocet®. While Tylox®, Percodan® and Percocet® contain only small doses of oxycodone hydrochloride combined with other analgesics such as acetaminophen or aspirin, OxyContin consists solely of oxycodone hydrochloride in a time-release form. Oxycodone is known to metabolize by demethylation into oxymorphone and noroxycodone. In a 24-hour urine, 33-61% of a single, 5 mg oral dose is excreted with the primary constituents being unchanged drug (13-19%), conjugated drug (7-29%) and conjugated oxymorphone (13-14%). The window of detection for Oxycodone in urine is expected to be similar to that of other opioids such as morphine.

1.3.18 2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP)

Methadone is an unusual drug in that its primary urinary metabolites (EDDP and EMDP) are cyclic in structure, making them very difficult to detect using immunoassays targeted to the native compound. Exacerbating this problem, there is a subsection of the population classified as “extensive metabolizers” of methadone. In these individuals, a urine specimen may not contain enough parent methadone to yield a positive drug screen even if the individual is in compliance with their methadone maintenance. EDDP represents a better urine marker for methadone maintenance than unmetabolized methadone.

1.3.19 Cotinine (COT)

Cotinine is the first-stage metabolite of nicotine, a toxic alkaloid that produces stimulation of the autonomic ganglia and central nervous system when in humans. Nicotine is a drug to which virtually every member of a tobacco-smoking society is exposed whether through direct contact or second-hand inhalation. In addition to tobacco, nicotine is also commercially available as the active ingredient in smoking replacement therapies such as nicotine gum, transdermal patches and nasal sprays.

In a 24-hour urine, approximately 5% of a nicotine dose is excreted as unchanged drug with 10% as cotinine and 35% as hydroxycotinine; the concentrations of other metabolites are believed to account for less than 5%.[10] While cotinine is thought to be an inactive metabolite, it’s elimination profile is more stable than that of nicotine which is largely urine pH dependent. As a result, cotinine is considered a good biological marker for determining nicotine use. The plasma half-life of nicotine is approximately 60 minutes following

inhalation or parenteral administration.[11] Nicotine and cotinine are rapidly eliminated by the kidney; the window of detection for cotinine in urine at a cutoff level of 200 ng/mL is expected to be up to 2-3 days after nicotine use.

1.3.20 Fentanyl (FYL)

Fentanyl, belongs to powerful narcotics analgesics, and is a μ special opiates receptor stimulant. Fentanyl is one of the varieties that been listed in management of United Nations "Single Convention of narcotic drug in 1961". Among the opiates agents that under international control, Fentanyl is one of the most commonly used to cure moderate to severe pain.[1] After continuous injection of fentanyl, the sufferer will have the performance of protracted opioid abstinence syndrome, such as ataxia and irritability etc,[2,3] which presents the addiction after taking fentanyl in a long time. Compared with drug addicts of amphetamine, drug addicts who take fentanyl mainly have got the possibility of higher infection rate of HIV, more dangerous injection behavior and more lifelong medication overdose.[4]

1.3.21 Synthetic Marijuana (K2)

Synthetic Marijuana or K2 a psychoactive herbal and chemical product that, when consumed, mimics the effects of Marijuana. It is best known by the brand names K2 and Spice, both of which have largely become genericized trademarks used to refer to any synthetic Marijuana product. The studies suggest that synthetic marijuana intoxication is associated with acute psychosis, worsening of previously stable psychotic disorders, and also may have the ability to trigger a chronic (longperm) psychotic disorder among vulnerable individuals such as those with a family history of mental illness.

Elevated levels of urinary metabolites are found within hours of exposure and remain detectable for 72 hours after smoking (depending on usage/dosage).

As of March 1, 2011, five cannabinoids, JWH -018, JWH- 073, CP- 47, JWH- 200 and cannabicyclohexanol are now illegal in the US because these substances have the potential to be extremely harmful and, therefore, pose an imminent hazard to the public safety.

1.3.22 6-mono-aceto-morphine (6-MAM)

6-Monoacetylmorphine (6-MAM) or 6-acetylmorphine (6-AM) is one of three active metabolites of heroin (diacetylmorphine), the others being morphine and the much less active 3-monoacetylmorphine (3-MAM). 6-MAM is rapidly created from heroin in the body, and then is either metabolized into morphine and excreted in the urine. 6-MAM remains in the urine for no more than 24 hours. The presence of 6-MAM guarantees that heroin was in fact used as recently as within the last day. 6-MAM is naturally found in the brain, but in such small quantities that detection of this compound in urine virtually guarantees that heroin has recently been consumed.

1.3.23 Clonazepam (CLO)

Clonazepam is a benzodiazepine drug having anxiolytic, anticonvulsant, muscle relaxant, amnestic, sedative, and hypnotic properties. Clonazepam has an intermediate onset of action, with a peak blood level occurring one to four hours after oral administration. Longperm effects of benzodiazepines include tolerance, benzodiazepine dependence, and benzodiazepine withdrawal syndrome, which occurs in one third of patients treated with clonazepam for longer than four weeks. Benzodiazepines such as clonazepam have a fast onset of action, high effectivity rate, and low toxicity in overdose; however, as with most medications, it may have drawbacks due to adverse or paradoxicaleffects.The detection period for the Benzodiazepines in the urine is 3-7 days.

1.3.24 Diazepam (DIA)

Diazepam, first marketed as Valium, is a medication of the benzodiazepine family that typically produces a calming effect. It is commonly used to treat a range of conditions including anxiety, alcohol withdrawal syndrome, benzodiazepine withdrawal syndrome, muscle spasms, seizures, trouble sleeping, and restless legs syndrome. It may also be used to cause memory loss during certain medical procedures. It can be taken by mouth, inserted into the rectum, injected into muscle, or injected into a vein. When given into a vein, effects begin in one to five minutes and last up to an hour.

1.3.25 Ethyl Glucuronide (ETG)

Ethyl Glucuronide (ETG) is a metabolite of ethyl alcohol which is formed in the body by glucuronidation following exposure to ethanol, such as by drinking alcoholic beverages. It is used as a biomarker to test for ethanol use and to monitor alcohol abstinence in situations where drinking is prohibited, such as in the military, in professional monitoring programs (health professionals, attorneys, airline pilots in recovery from addictions), in schools, in liver transplant clinics, or in recovering alcoholic patients. ETG can be measured in urine up to approximately 80 hours after ethanol is ingested. ETG is a more accurate indicator of the recent exposure to alcohol than measuring for the presence of ethanol itself.

1.3.26 Acetaminophen (ACE)

Acetaminophen is one of the most commonly used drugs, yet it is also an important cause of serious liver injury. Acetaminophen is the generic name of a drug found in many common brand name over-the-counter (OTC) products, such as Tylenol, and Prescription (Rx) products, such as Vicodin and Percocet. Acetaminophen is an important drug, and its effectiveness in relieving pain and fever is widely known. Unlike other commonly used drugs to reduce pain and fever (e.g., nonsteroidalantinflammatory drugs (NSAIDs), such as aspirin, ibuprofen, and naproxen), at recommended doses acetaminophen does not cause adverse effects, such as stomach discomfort and bleeding, and acetaminophen is considered safe when used according to the directions on its OTC or Rx labeling. However, taking more than the recommended amount can cause liver damage, ranging from abnormalities in liver function blood tests, to acute liver failure, and even death. Many cases of overdose are caused by patients inadvertently taking more than the recommended dose (i.e., 4 grams a day) of a particular product, or by taking more than one product containing acetaminophen (e.g., an OTC product and an Rx drug containing acetaminophen). The mechanism of liver injury is not related to acetaminophen itself, but to the production of a toxic metabolite. The toxic metabolite binds with liver proteins, which cause cellular injury. The ability of the liver to remove this metabolite before it binds to liver protein influences the extent of liver injury.

1.3.27 Cathinone (CAT)

Cathinone, also known as benzoylethanamine, or β-keto-amphetamine is a monoamine alkaloid found in the shrub Catha edulis (Khat) and is chemically similar to ephedrine, Cathine, methCathinone and other amphetamines. Cathinone has been found to stimulate the release of dopamine and inhibit the re-uptake of epinephrine, norepinephrine, and serotonin in the Central Nervous System (CNS). These neurotransmitters are all considered monoamines and share the general structure of an aromatic ring and an amine group attached by a two-carbon separator. Internationally, cathinone is a Schedule I drug under the Convention on Psychotropic Substances. Cathinone has been shown to selectively metabolize into R,S-(-)-norephedrine and cathine. The reduction of the ketone group in cathinone will produce cathine.

Cathine, also known as d-norpseudoephedrine, is a psychoactive drug of the phenethylamine and amphetamine chemical classes which acts as a stimulant. Along with cathinone, it is found naturally in Catha edulis (Khat), and contributes to its overall effects. It has approximately 10-14% the potency of amphetamine.Cathine is a Schedule III drug under the Convention on Psychotropic Substances.

Detection of Cathine in urine thus provides evidence of presence of Cathinone and/ or Cathine use, by the person, whose urine is tested.

1.3.28 Lysergic Acid Diethylamide (LSD)

Lysergic acid diethylamide (LSD) is a white powder or a clear, colorless liquid. LSD is manufactured from lysergic acid which occurs naturally in the ergot fungus that grows on wheat and rye. It is a Schedule I controlled substance, available in liquid, powder, tablet (microdots), and capsule form. LSD is recreationally used as a hallucinogen for its ability to alter human perception and mood. LSD is primarily used by oral administration, but can be inhaled, injected, and transdermally applied. LSD is a non-selective 5-HT agonist, may exert its hallucinogenic effect by interacting with 5-HT 2Areceptors as a partial agonist and modulating the NMDA receptor-mediated sensory, perceptual, affective and cognitive processes. LSD mimics 5-HT at 5-HT 1A receptors, producing a marked slowing of the firing rate of serotonergic neurons. LSD has a plasma half-life of 2.5-4 hours. Metabolites of LSD include N-desmethyl-LSD, hydroxy-LSD, 2-oxo-LSD, and

2-oxo-3-hydroxy-LSD .These metabolites are all inactive. LSD use can typically be detected in urine for periods of 2-5 days.

1.3.29 3,4-Methylenedioxyamphetamine (MDA)

3,4-Methylenedioxyamphetamine (MDA), also known as tenamfetamine (INN), or with the street name "Sally" or "Sass" or "Sass-a-frass", is a psychedelic and entactogenic drug of the phenethylamine and amphetamine chemical classes. It is mainly used as a recreational drug, an entheogen, and a tool in use to supplement various types of practices for transcendence, including in meditation, psychonautics, and as an agent in psychedelic psychotherapy. It was first synthesized by G. Mannish and W. Jacobson in 1910. There are about 20 different synthetic routes described in the literature for its preparation.

1.3.30 3, 4-methylenedioxypyrovalerone (MDPV)

3, 4-methylenedioxypyrovalerone (MDPV) is a psychoactive recreational drug with stimulant properties which acts as a norepinephrine-dopamine reuptake inhibitor (NDRI). It was first developed in the 1960s by a team at BoehringerIngelheim. MDPV remained an obscure stimulant until around 2004 when it was reportedly sold as a designer drug. Products labeled as bath salts containing MDPV were previously sold as recreational drugs in gas stations and convenience stores in the United States, similar to the marketing for Spice and K2 as incense.

MDPV is the 3,4-methylenedioxy ring-substituted analog of the compound pyrovalerone, developed in the 1960s, which has been used for the treatment of chronic fatigue and as an anorectic, but caused problems of abuse and dependence. However, despite its structural similarity, the effects of MDPV bear little resemblance to other methylenedioxyphenylalkylamine derivatives such as 3,4-methylenedioxy-N-methylamphetamine (MDMA), instead producing primarily stimulant effects with only mild entactogenic qualities.

MDPV undergoes CYP450 2D6, 2C19, 1A2, and COMT phase 1 metabolism (liver) into methylcatechol and pyrrolidine, which in turn are glucuronated (uridine 5'-diphospho-glucuronosyl-transferase) allowing it to be excreted by the kidneys, with only a small fraction of the metabolites being excreted into the stools. No free pyrrolidine will be detected in the urine.

1.3.31 Methylphenidate (MPD)

Methylphenidate (Ritalin) is a psychostimulant drug approved for treatment of ADHD or attention-deficit hyperactivity disorder, postural orthostatic tachycardia syndrome and narcolepsy. Methylphenidate primarily acts as a norepinephrine-dopamine reuptake inhibitor. Methylphenidate is most active at modulating levels of dopamine and to a lesser extent norepinephrine. Similar to cocaine, methylphenidate binds to and blocks dopamine transporters and norepinephrine transporters. Methylphenidate has both dopamine transporter and norepinephrine transporter binding affinity, with the dextromethylphenidate enantiomers displaying a prominent affinity for the norepinephrine transporter. Methylphenidate may also exert a neuroprotective action against the neurotoxic effects of Parkinson's disease and methamphetamine abuse. Methylphenidate taken orally has a bioavailability of 11-52% with a duration of action around 1-4 hours for instant release, 3–8 hours for sustained release, and 8–12 hours for extended release(Concerta). The half-life of methylphenidate is 2-3 hours, depending on the individual. The peak plasma time is achieved at about 2 hours.

1.3.32 Zolpidem (ZOL)

Zolpidem (brand names Ambien, Ambien CR, Intermezzo, Stilnox, Stilnoct, Sublinox, Hypnogen, Zonadin, Sanval and Zolsana) is a prescription medication used for the treatment of insomnia and some brain disorders. It is a short-acting nonbenzodiazepine hypnotic of the imidazopyridine class.

That potentiates GABA, an inhibitoryneurotransmitter, by binding to GABA receptors at the same location as benzodiazepines. It works quickly, usually within 15 minutes, and has a short half-life of two to three hours.

Zolpidem may be detected in blood or plasma to confirm a diagnosis of poisoning in hospitalized patients, provide evidence in an impaired driving arrest, or to assist in a medico-legal death investigation. Blood or plasma Zolpidem concentrations are usually in a range of 30–300 µg/l in persons receiving the drug

therapeutically, 100–700 µg/l in those arrested for impaired driving, and 1000–7000 µg/l in victims of acute overdosage. Analytical techniques, in general, involve gas or liquid chromatography.[3,4,5]

1.3.33 Zopiclone(ZOP)

Zopiclone is a nonbenzodiazepine hypnotic agent used in the treatment of insomnia. It is a cyclopyrrolone, which increases the normal transmission of the neurotransmitter gamma-aminobutyric acid in the central nervous system, as benzodiazepines do, but in a different way. Zopiclone is indicated for the short-term treatment of insomnia where sleep initiation or sleep maintenance are prominent symptoms. Long-term use is not recommended, as tolerance, dependence, and addiction can occur with prolonged use. Zopiclone is partly extensively metabolized in the liver to form an active N-demethylated derivative (N-desmethylzopiclone) and an inactive zopiclone-N-oxide.

In urine, the N-demethyl and N-oxide metabolites account for 30% of the initial dose. Between 7 and 10% of zopiclone is recovered from the urine, indicating extensive metabolism of the drug before excretion. The terminal elimination half-life of zopiclone ranges from 3.5 to 6.5 hours (5 hours on average). Time to peak plasma concentration is 1 - 2 h, the absorption rate constant is 1.3 h-1 and maximum plasma concentration after administration of 7.5 mg is 131µg/l. Zopiclone is recommended to be taken on a short-term basis, usually a week or less. Daily or continuous use of the drug is not usually advised. When taken alone, it usually is not fatal, but when mixed with alcohol or other drugs such as opioids, or in patients with respiratory, or hepatic disorders, the risk of a serious and fatal overdose increases.

Zopiclone may be measured in blood, plasma, or urine by chromatographic methods. Plasma concentrations are typically less than 100ug/l during therapeutic use, but frequently exceed 100ug/l in automotive vehicle operators arrested for impaired driving ability and may exceed 1000ug/l in acutely poisoned patients.

1.3.34 Methcathinone(MCAT)

Methcathinone is a monoamine alkaloid and psychoactive stimulant, a substituted cathinone. Methcathinone is a highly addictive drug, primarily psychologically addicting and most of the signs of addiction to the drug are emotional or psychological. It has been popularized and continues to be sold under misleading names such as "bath salts", "plant fertilizers" or "research chemicals", but it is actually a powerful psycho-stimulant used as a recreational drug. Effects of this drug typically last from 4 to 6 hours. It is used as a recreational drug due to its potent stimulant and euphoric effects and is considered to be addictive, with both physical and psychological withdrawal occurring if its use is discontinued after prolonged or high-dosage administration. It is usually snorted, but can be smoked, injected, or taken orally. Methcathinone is listed as a Schedule I controlled substance by the Convention on Psychotropic Substances and the United States' Controlled Substances Act, and as such it is not considered to be safe or effective in the treatment, diagnosis, prevention, or cure of any disease, and has no approved medical use. Methcathinone has very strong affinities for the dopamine transporter and the norepinephrine (noradrenaline) transporter. Its affinity for the serotonin transporter is less than that of methamphetamine.

Effects of short term intoxication are similar to those produced by crack cocaine or methamphetamine: stimulation of heart rate and respiration; feeling of euphoria; loss of appetite; increased alertness; pupils may be dilated; body temperature may be slightly elevated. Acute intoxication at higher doses may also result in: insomnia, tremors and muscle twitching, fever, headaches, convulsions, irregular heart rate and respirations, anxiety, restlessness, paranoia, hallucinations and delusions.

1.3.35 Caffeine(CAF)

Caffeine is a central nervous system (CNS) stimulant of the methylxanthine class. It is the world's most widely consumed psychoactive drug. It is found in the seeds, nuts, or leaves of a number of plants native to South America and East Asia and confers on them several survival and reproductive benefits.

Caffeine can produce a mild form of drug dependenc- associated with withdrawal symptoms such as sleepiness, headache, and irritability – when an individual stops using caffeine after repeated daily intake. After intravenous administration of caffeine the urine level of the drug is approximately the same in each of the first 4 hourly specimens. Blood samples taken 10 and 70 minutes after injection of the drug were

analyzed and showed 0.29 and 0.28mg. per 100 cc. respectively. There are to be contrasted with the 1st hour urine which contained 0.73mg.per 100 cc., essentially 3 times that in the blood. After oral administration of caffeine to the horse the concentration of caffeine in the urine rose progressively during the first 3 hours, remained relatively constant through the 8th hours. At 48 hours, a urine specimen contained approximately 0.17mg. per 100 cc. of caffeine. In addition, flu-like symptoms, nausea/vomiting, and muscle pain/stiffness were judged likely to represent valid symptom categories. In experimental studies, the incidence of headache was 50% and the incidence of clinically significant distress or functional impairment was 13%. Typically, onset of symptoms occurred 12-24 h after abstinence, with peak intensity at 20-51 h, and for a duration of 2-9 days. 151% to 3% of caffeine is excreted unchanged in the urine. The rate of caffeine metabolism is variable, with a half-life of 4 to 6h.

1.3.36 Tropicamide(TRO)

Tropicamide is an antimuscarinic drug usually prescribed as an ophthalmic solution to induce short-term mydriasis and cycloplegia. Tropicamide is currently abused (injected intravenously) as an inexpensive recreational deliriant drug.

Misuse of tropicamide typically occurs through IV injection; its effects last from 30 min to 6 h, and It is usually mixed with heroin, methadone, and other opioid drugs to potentiate the " rush" when injected intravenously.Medical effects of tropicamide misuse include slurred speech, persistent mydriasis, unconsciousness/unresponsiveness, hallucinations, kidney pain, dysphoria, “open eye dreams,” hyperthermia, tremors, suicidal feelings, convulsions, psychomotor agitation, tachycardia and headache.

1.3.37 alpha-Pyrrolidinovalerophenone (α-PVP**)**

alpha-Pyrrolidinovalerophenone (also known as α-PVP, A-PVP, alpha-PVP, and Flakka) is a synthetic stimulant substance of the cathinone and pyrrolidine chemical classes.α-PVP may be quantified in blood, plasma or urine to confirm a diagnosis of poisoning in hospitalized patients or to provide evidence in a medicolegal death investigation. It generally comes in the form of either a crystalline powder or crystallized shards which users can ingest to produce powerful but short-lived euphoric stimulant effects which are comparable to those of methamphetamine and cocaine when insufflated or vaporized. α-PVP has been reported to be the cause, or a significant contributory cause of death in suicides and overdoses caused by combinations of drugs. It has also been linked to at least one death where it was combined with pentedrone and caused heart failure.

1.3.38 Pregabalin (PGB)

Pregabalin, also known as β-isobutyl-γ-amino butyric acid (beta-isobutyl-GABA), is a medication used to treat epilepsy, neuropathic pain, fibromyalgia, and generalized anxiety disorder. Common side effects include: sleepiness, confusion, trouble with memory, poor coordination, dry mouth, problem with vision, and weight gain. Potentially serious side effects include angioedema, drug misuse, and an increased suicide risk.

Pregabalin is eliminated from the systemic circulation primarily by renal excretion as unchanged drug. The Pregabalin is predominantly excreted unchanged in urine (≧98%). Pregabalin mean elimination half-life is 6.3 hours. 50% would be expected to have negative urine specimens within 3 days and a total of 5 days would be needed to achieve negative urine specimens in the subject with the maximum urinary concentration measured.

1.3.39 AB-PINACA(ABP)

AB-PINACA is a compound that was first identified as a component of synthetic cannabis products in Japan in 2012.43 It was originally developed by Pfizer in 2009 as an analgesic medication. AB-PINACA acts as a potent agonist for the CB1 receptor (Ki = 2.87 nM, EC50 = 1.2 nM) and CB2 receptor (Ki = 0.88 nM, EC50 = 2.5 nM) and fully substitutes for Δ9-THC in rat discrimination studies, while being 1.5x more potent.

1.3.40 7-aminoclonazepam(7-ACL)

7-aminoclonazepam is the major metabolite of clonazepam. Clonazepam sold under the brandname Klonopin among others, is a medication used to prevent and treat seizures, panic disorder, and for the

movement disorder known as akathisia. It is a type of benzodiazepine. As a major metabolite, 7-aminoclonazepam may be used to monitor use of the parent drug, clonazepam. Clonazepam, marketed as Klonopin and Rivotril, is a long-acting benzodiazepine with anxiolytic, anticonvulsant, muscle relaxant, and hypnotic properties.

1.3.41 Alprazolam(ALP)

Alprazolam, available under the trade name Xanax among others, is a short-acting anxiolytic of the benzodiazepine class. It is commonly used for the treatment of panic disorder, and anxiety disorders, such as generalized anxiety disorder (GAD) or social anxiety disorder (SAD). Alprazolam, like other benzodiazepines, binds to specific sites on the GABAA receptor. It possesses anxiolytic, sedative, hypnotic, skeletal muscle relaxant, anticonvulsant, and amnestic properties.

A mean half-life of alprazolam of 16.3 hours has been observed in healthy elderly subjects (range: 9.0-26.9 hours, n=16) compared to 11.0 hours (range: 6.3-15.8 hours, n=16) in healthy adult subjects.

Alprazolam and its metabolites are excreted primarily in the urine. The pharmacokinetics of alprazolam and two of its major active metabolites (4-hydroxyalprazolam and α-hydroxyalprazolam) are linear, and concentrations are proportional up to the recommended maximum daily dose of 10 mg given once daily. Peak concentrations in the plasma occur in one to two hours following administration. Plasma levels are proportionate to the dose given; over the dose range of 0.5 to 3.0 mg, peak levels of 8.0 to 37ng/ml were observed.

1.3.42 Mephedrone(MEP)

Mephedrone, also known as 4-methylmethcathinone (4-MMC) or 4-methylephedrone is a synthetic stimulant drug of the amphetamine and cathinone classes.Slang names include drone,M-CAT, White Magic and meow meow. It is chemically similar to the cathinone compounds found in the khat plant of eastern Africa.

Mephedrone comes in the form of tablets or a powder, which users can swallow, snort or inject, producing similar effects to MDMA, amphetamines and cocaine. In addition to its stimulant effects, mephedrone produces side effects, of which teeth grinding are the most common. A number of metabolites are possible, however the n-demethyl metabolite of Mephedrone will be 4-Methylcathinone. This metabolite appears to be nearly inactive as a Monoamine Oxydase Inhibitor.On further metabolism of this metabolite one of the possible metabolites is 4-Methylnorephedrine, caused by the reduction of the Keto. A dose of 150mg-250mg is the average, giving a duration of around 2 hours.the duration will lengthen in larger 250mg+ dosages .

1.3.43 Cannabinol (CNB)

Cannabinol (CNB) is a non-psychoactive cannabinoid found only in trace amounts in Cannabis,1 and is mostly found in aged Cannabis. Pharmacologically relevant quantities are formed as a metabolite of tetrahydrocannabinol (CNB).3CNB acts as a partial agonistat the CB1 receptors, but has a higher affinity to CB2 receptors; however, it has lower affinities relative to CNB.4，5 Degraded or oxidized cannabis products, such as low-quality baled cannabis and traditionally produced hashish, are high in CNB, but modern production processes minimize the formation of CNB.

Unlike other cannabinoids, CNB does not stem from cannabigerol (CBG). There is no clinical evidence that CNB breaks down naturally into CNB once the CNB has become decarboxylated and forms delta-9 CNB. CNB is formed by decarboxylation of cannabinolic acid.

1.3.44 Meperidine(MPRD)

Meperidine (also known as Pethidine , Pethidin , Meperidol and Dolantin) a phenylpiperidine derivative, is a synthetic opioid analgesic. Many of its pharmacologic properties and indications are similar to those of morphine, Meperidine is preferred to morphine for obstetric use because its rapid onset of action and shorter duration usually permit greater flexibility in maternal analgesia, possibly with less effect on neonatal respiration.

Like other opioid drugs, pethidine has the potential to cause physical dependence or addiction. It may be

more likely to be abused than other prescription opioids, perhaps because of its rapid onset of action.When compared with oxycodone, hydromorphone, and placebo, pethidine was consistently associated with more euphoria, difficulty concentrating, confusion, and impaired psychomotor and cognitive performance when administered to healthy volunteers. The especially severe side effects unique to pethidine among opioids-serotonin syndrome, seizures, delirium, dysphoria, tremor—are primarily or entirely due to the action of its metabolite, norpethidine.

1.3.45 Trazodone (TZD)

Trazodone is a triazolopyridine derivative which is used to treat major depressive disorder. It possesses antidepressant, and also some anxiolytic and hypnotic activity.

An initial dose of 150 mg/day in divided doses is suggested. The dose may be increased by 50 mg/day every three to four days. The maximum dose for outpatients usually should not exceed 400 mg/day in divided doses. Inpatients (ie, more severely depressed patients) may be given up to but not in excess of 600 mg/day in divided doses. The half-life of trazodone in the initial phase is about 3–6 h and the half-life in the terminal phase is about 5–9 h.TZD is extensively metabolized with only about 1% of the dose excreted unchanged in urine after 24 hr. Around 70 to 75% of 14C-labelled trazodone was found to be excreted in the urine within 72 hours.

1.3.46 Zaleplon (ZAL)

Zaleplon is a sedative-hypnotic, almost entirely used for the management/treatment of insomnia. It is a nonbenzodiazepine hypnotic from the pyrazolopyrimidine class.

Zaleplon has a pharmacological profile similar to benzodiazepines, characterized by an increase in slow wave deep sleep (SWDS) with rapid onset of hypnotic action. Zaleplon is a full agonist for the benzodiazepine α1 receptor located on the GABAA receptor complex in the body, with lower affinity for the α2 and α3 subsites. It selectively enhances the action of GABA similar to, but more selectively than benzodiazepines. Zaleplon, although not a benzodiazepine, maintains a very similar chemical structure nonetheless, known for inducing hypnotic effects by α1 subreceptor sites, anxiolytic and muscle relaxant effects via α2 and α3 subsites, with negligible anticonvulsant properties (via α5 subsite), as zaleplon action is modulated at benzodiazepine receptor sites. The elimination half-life of zaleplon is about 1–1.5 hours. The absorption rate of zaleplon is rapid and the onset of therapeutic effects is typically breached within 5–15 minutes following ingestion.Zaleplon is primarily metabolised by aldehyde oxidase, and its half-life can be affected by substances which inhibit or induce aldehyde oxidase. Taken orally, zaleplon reaches full concentration in about one hour. It is extensively metabolised into 5-oxozaleplon and 5-oxodesethylzaleplon (the latter via desethylzaleplon), with less than 1% of it excreted intact in urine.

1.3.47 UR-144

UR-144 is a synthetic cannabinoid receptor agonist (SCRA) and has affinity for CB1 and CB2 receptors. It has a high selectivity for the CB2-receptors.

UR-144 is a psychoactive substance and has effects similar to delta-9-tetrahydrocannabinol (THC), though slightly less potent than THC. UR-144 has been detected in herbal products marketed under a variety of names.

In mice, UR-144 is moderately potent in reducing in a time- and dose-dependent manner the locomotor activity (ID50-value 7.8 mg/kg), induces an anti-nociceptive effect, and decreases rectal temperature and ring immobility with potencies several-fold greater than THC. In mice, UR-144 substituted for THC in a THC discrimination study (ED50-value 7.1 to 7.4 µmol/kg intra-peritoneal), an effect antagonized by rimonabant.

1.3.48 Mescaline(MES)

Mescaline (MES) is a derivative of phenethylamine. It is a strong hallucinogen. Can cause hallucinations, self-distortion and mind splitting, leading to depression and pupil dilation, tachycardia, excessive limbs, tremors, nausea, vomiting, and long-term use can lead to organ damage.Hallucinations can last for seven, eight or even 12 hours.The main hazard of using Mescaline is mental disorder.If users experience transient psychosis, they can also engage in violent attacks, suicide, self-harm and other behaviors.

1.3.49 Gabapentin(GAB)

Gabapentin, sold under the brand name Neurontin among others, is a medication which is used to treat epilepsy (specifically partial seizures), neuropathic pain, hot flashes, and restless legs syndrome.
Common side effects of gabapentin include sleepiness and dizziness.Serious side effects include an increased risk of suicide, aggressive behavior, and drug reaction with eosinophilia and systemic symptoms. In 2009 the U.S. Food and Drug Administration issued a warning of an increased risk of suicidal thoughts and behaviors in patients taking some anticonvulsant drugs, including gabapentin,modifying the packaging inserts to reflect this.
The oral bioavailability of gabapentin enacarbil (as gabapentin) is greater than or equal to 68%, across all doses assessed (up to 2,800 mg), with a mean of approximately 75%. Gabapentin undergoes little or no metabolism.The Tmax of the instant-release (IR) formulation of gabapentin enacarbil (as active gabapentin) is about 2.1 to 2.6 hours across all doses (350–2,800 mg) with single administration and 1.6 to 1.9 hours across all doses (350–2,100 mg) with repeated administration.

1.3.50 Tilidine (TLD)

Tilidine, or tilidate (brand names: Tilidin, Valoron and Valtran) is a synthetic opioid painkiller, used mainly in Germany, Switzerland, South Africa and Belgium for treatment of moderate to severe pain, both acute and chronic. Its onset of pain relief after oral administration is about 10-15 minutes and peak relief from pain occurs about 25-50 minutes after oral administration.
It usually comes in its hydrochloride hemihydrate salt form; in this form it is highly soluble in water, ethanol and dichloromethane and appears as a white/almost white crystalline powder.[3] Its storage is restricted by its sensitivity to degradation by light and oxygen, hence necessitating its storage in amber bottles and at temperatures below 30 degrees Celsius, respectively.
Tilidine is a prodrug from which the active metabolite nortilidine is formed by demethylation. The pharmacokinetics of tilidine (T), nortilidine (NT) and bisnortilidine (BNT) were studied in nine healthy subjects following single intravenous (10 min infusion) and oral 50 mg T-HCl dose as well as following multiple 50 mg T-HCl oral doses. Systemic availability of the parent substance was 6% and of the active metabolite NT 99%. The terminal half-life of NT was 3.3 h following single oral administration, 4.9 h following intravenous administration and 3.6 h following multiple dosing. Following intravenous infusion, concentrations of unchanged substance were found which were 30 times higher than following oral administration. BNT was eliminated with half-lives of 5 h after oral administration and 6.9 h after intravenous administration. Renal elimination of unchanged substance was 1.6% of the dose following intravenous administration and less than 0.1% of the dose following oral administration. Approximately 3% were recovered in urine as NT and 5% as BNT following both routes of administration.

1.3.51 Fluoxetine (FLX)

Fluoxetine, also known by trade names Prozac and Sarafem, among others, is an antidepressant of the selective serotonin reuptake inhibitor (SSRI)class. It is used for the treatment of major depressive disorder, obsessive–compulsive disorder(OCD), bulimia nervosa, panic disorder and premenstrual dysphoric disorder. It may decrease the risk of suicide in those over the age of 65. It has also been used to treat premature ejaculation. Fluoxetine is taken by mouth.
Common side effects include trouble sleeping, sexual dysfunction, loss of appetite, dry mouth, rash and abnormal dreams. Serious side effects include serotonin syndrome, mania, seizures, an increased risk of suicidal behavior in people under 25 years old and an increased risk of bleeding. If stopped suddenly, a withdrawal syndromemay occur with anxiety, dizzinessand changes in sensation. It is unclear if it is safe in pregnancy. If already on the medication, it may be reasonable to continue during breastfeeding. Its mechanism of action is not entirely clear but believed to be related to increasing serotonin activity in the brain.

1.3.52 Quetiapine(QTP)

Quetiapine, sold under the trade name Seroquel among others, is an atypical antipsychotic used for the treatment of schizophrenia, bipolar disorder, and major depressive disorder. It is also used as a sleep aid due to its sedating effect, but this use is not recommended. It is taken by mouth.

Common side effects include sleepiness, constipation, weight gain, and dry mouth. Other side effects include low blood pressure with standing, seizures, a prolonged erection, high blood sugar, tardive dyskinesia, and neuroleptic malignant syndrome. In older people with dementia, its use increases the risk of death. Use in the later part of pregnancy may result in a movement disorder in the baby for some time after birth. Quetiapine is believed to work by blocking a number of receptors including serotonin and dopamine.

1.3.53 Papaverine(PAP)

Papaverine (latin Papaverine,"popy")is an opiu alkaloid antispasmodic drug,used primarily in the treatment of visceral spasm and vasospasm(especially those involving the intestines,heart,or brain),and occasionally in the treatment of eretile dysfunction.It is used in the treatment of acute mesenteric ischemia.While it is found in the opium poppy,Papaverine differs in both structure and pharmacological action from the analgesic (morphine-related)opium alkaloids(opiates).

Papaverine is found as a contaminantin some hroin and can be used by forensic laboratories in heroin profiling to identify its source. The metabolites can aslo be found in the urine of herin users,allowing street heroin to be distinguished from pharmaceutical.

Papaverine (4-(3', 4'-dimethoxybenzyl)-6, 7-dimethoxy-quinoline, Mw 339), as one of benzyl isoquinoline alkaloids, was used clinically as a bronchodilator to relaxes various smooth muscles, smooth musculature of the larger blood vessels, especially coronary, systemic peripheral and pulmonary arteries to increase cerebral blood flow .

By the in vitro metabolic experiment, there were five metabolites in liver microsomal incubation solution and two metabolites in intestinal flora incubation solution.

1.3.54 Kratom(KRA)

Mitragynine (MG) and its major metabolites 7-hydroxyMitragynine (7-OH-MG) are two of the major components of the plant extract Kratom, which is a tree planted in Southeast Asia. Kratom has long been used by opioid-dependent individuals as an alternative to their unavailable opioid of choice and chronic pain medication, as a stealth-to-urine drug screening opiate substitute while in opioid recovery treatment and recreationally, alone or as a booster.In this study, a direct infusion method was utilized and electrospray ionization triple quadrupole mass spectrometer was used as the detector for data acquisition. Pharmacokinetic study was conducted to investigate the effect of Mitragynine and 7-hydroxyMitragynine and major fragments of both compounds were proposed.

1.3.55 Carisoprodol(CAR)

Carisoprodol, marketed under the brand name Soma among others, is a medication used for musculoskeletal pain. Use is only approved for up to three weeks. Effects generally begin within half an hour and last for up to six hours. It is taken by mouth.

Common side effects include headache, dizziness, and sleepiness. Serious side effect may include addiction, allergic reactions, and seizures. In people with a sulfa allergy certain formulations may result in problems. Safety during pregnancy and breastfeeding is not clear. Meprobamate and other muscle-relaxing drugs often were subjects of misuse in the 1950s and 60s. Overdose cases were reported as early as 1957, and have been reported on several occasions since then.

Carisoprodol is metabolized by the liver and excreted by the kidneys so this drug must be used with caution with patients that have impaired hepatic or renal function. Because of potential for more severe side effects, this drug is on the list to avoid for elderly people.

1.3.56 Carfentanyl (CFYL)

Carfentanyl is an analog of the synthetic opioid analgesic fentanyl. It is 10,000 times more potent than morphine, making it among the most potent commercially used opioids. Carfentanil was first synthesized in 1974. It is marketed under the trade name Wildnil as a general anaesthetic agent for large animals. Side effects of carfentanil are similar to those of fentanyl, which include itching, nausea and respiratory depression, which can be life-threatening. Carfentanil is classified as Schedule II under the Controlled Substances Act in the United States with a DEA ACSCN of 9743.

1.3.57 Risperidone (RPD)

Risperidone is a benzisoxazole derivative with serotonergic and dopaminergic receptor antagonist properties. The drug has been used clinically since 1990 as an antipsychotic agent. It is supplied as the free base in 0.25-6 mg tablets and a 1 mg/mL solution for oral administration; daily oral dose for adults usually range from 2-6 mg. Adverse reactions include dizziness, somnolence, nausea, orthostatic hypotension, anxiety, headache and extrapyramidal symptoms. A single labeled oral risperidone dose is eliminated in urine (70%) and feces (14%) over a one week period.

1.3.58 Citalopram (CIT)

Citalopram, an antidepressant of the selective serotonin reuptake inhibitor (SSRI) type, is only available as a racemic drug. Its metabolism occurs, partially, by N-demethylation to demethylcitalopram (DCIT) and didemethylcitalopram (DDCIT), but also by deamination to a propionic acid derivative (CIT-PROP) and by N-oxidation to CIT-N-oxide (NO-CIT). CIT has equal affinity for 5-HTT and dopamine transporters [2]. Studies have shown that compared with healthy subjects, the binding potential of thalamus and brainstem binding area in patients with depression who take 20-60 mg citalopram every day is reduced by 50%.

It is supplied as the free base in 20 mg tablets oral administration; daily oral dose for adults usually range from 20-60 mg. The common adverse reactions were nausea, dry mouth and drowsiness. The rare adverse reactions were agitation, income and anxiety. The oral bioavailability of citalopram is about 80%. 15% of citalopram is excreted by kidney, and 12%-23% of citalopram in daily dose is excreted in original form from urine.

1.3.59 F-Ketamine (FKET)

F-Ketamine is generally used as the replacement of ketamine with hallucinogenic properties. When abused, people are immersed in anesthesia situation with distorting perceptions of sight and sound and feeling out of control. Currently, there is limited research and using regulation of F-Ketamine. Therefore, a large number of people abuse F-Ketamine instead of ketamine. But nowadays, the detection of F-Ketamine is restricted to high performance liquid chromatography (HPLC), gas chromatograph-mass spectrometer (GC-MS), thin layer chromatography (TLC), etc. which require not only expensive equipment but professional analysis. At the same time, high purity of test material and complex operation are necessary as well. Therefore, it is a tendency that the test methods of FKET become fast and convenient.

1.3.60 Tapentadol (TAP)

Tapentadol is used to help relieve moderate to severe short-term pain such as pain from an injury or after surgery. It belongs to a class of drugs known as opioid analgesics. It works in the brain to change how the body feels and responds to pain. Tapentadol is mainly metabolized in the liver and is excreted by the kidneys in urine as well as in feces. The major pathway of metabolism is conjugation with glucuronic acid to produce glucuronides; tapentadol-O-glucuronide is the major metabolite. For the most part, shorter-term (15 weeks) therapy with tapentadol ER 100–250 mg twice daily provided more effective pain relief and significantly improved functional and health.

The most common side effects of TAP are related to gastrointestinal tract (nausea and vomiting) and the nervous system (dizziness, headache and somnolence). Due to its mu-opioid receptor agonism, gastrointestinal and central nervous system related adverse effects may also exist. Convulsion, hypersensitivity and impaired gastric emptying were rarely reported.

1.3.61 Scopolamine (SCOP)

Scopolamine is a tertiary amine antifungal drug.often used in sedative anesthesia, cough and asthma.It is supplied as the free base in 0.2-0.3 mg tablets and 0.25% ~ 1% for eye treatment;Adverse reactions include tachycardia, vertigo, restlessness, tremor, fatigue and dyskinesia, drowsiness, restlessness, memory impairment, hallucinations and confusion, urinary retention, conjunctivitis, eczema dermatitis, contact dermatitis, and acute angle closure glaucoma.

1.3.62 Mirtazapine (MTZ)

Mirtazapine is a norepinephrine and specific serotonin antidepressant, which can act on central

presynaptic α2 receptor antagonist and enhance adrenergic nerve conduction. The drug has been used clinically since 1990s as an antidepressants. The effective dose for adults is usually 15-45 mg/d. Adverse reactions include increased appetite, weight gain, sedationdizziness, somnolence, nausea, orthostatic hypotension,and mania. A single labeled oral mirtazapine dose is eliminated in urine (85%) and feces (15%) over a one week period.

1.3.63 N,N-Dimethyltryptamine (NND)

N, N-Dimethyltryptamine is a powerful and illegal hallucinogenic drug. While there is an open debate on whether or not N, N-Dimethyltryptamine is addictive, this drug poses dangerous physical and psychological effects for frequent users. Additionally, individuals who frequently abuse N, N-Dimethyltryptamine often develop a dependence on the drug, proving the dangers of N, N-Dimethyltryptamine abuse.

N, N-Dimethyltryptamine produces a high referred to as a psychedelic trip. When smoked, a N, N-Dimethyltryptamine trip usually begins instantly and lasts less than an hour. However, when drunk in the form of Ayahuasca tea, hallucinations begin after about 30 minutes and may last from 4-6 hours. Some users report experiencing mild, lingering hallucinogenic effects for several days after consuming or smoking N, N-Dimethyltryptamine. On a physiological level, N, N-Dimethyltryptamine causes an array of adverse side effects.


1.1.4 Principles

1.1.5 Principles (For DOA Tests)

The Multi-Drug Rapid Test Panel is an immunoassay based on the principle of competitive binding. Drugs which may be present in the urine specimen compete against their respective drug conjugate for binding sites on their specific antibody.

During testing, a urine specimen migrates upward by capillary action. A drug, if present in the urine specimen below its cut-off concentration, will not saturate the binding sites of its specific antibody. The antibody will then react with the drug-protein conjugate and a visible colored line will show up in the test line region. The presence of drug above the cut-off concentration will saturate all the binding sites of the antibody. Therefore, the colored line will not form in the test line region.

A drug-positive urine specimen will not generate a colored line in the test line region because of drug competition, while a drug-negative urine specimen will generate a line in the test line region because of the absence of drug competition. To serve as a procedural control, a colored line will always appear at the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.



As shown in illustration above, the specimen (A) migrates via capillary action along the membrane to react with the colored conjugate (B) Antidepressants present in the specimen binds to the conjugate, forming a colored antibody-antigen complex. Antidepressants antigen immobilized in the test zone of the membrane captures the test region (C). The formation of a visible color line in the test region indicates a negative result (C). The absence of a color line in the test zones suggests a positive result. In the control zone of the membrane, immobilized reagents capture colored conjugate regardless of test specimen composition. The resulting visible colored band (D) confirms control line.

1.2 Performance and Specification

The Multi-Drug Rapid Test Panel is a rapid chromatographic immunoassay for the detection of drugs and drug metabolites in urine. The test utilizes a combination of Antidepressants antigen to selectively detect

levels of Antidepressants in urine. At the level of claimed sensitivity, The Multi-Drug Rapid Test Panel shows no cross-reactivity interference from the drugs of Acetophenetidin, Bilirubin, and Isoxsuprine and so on.

1.2.1 Precautions

For self-testing use.Do not use after the expiration date.

The test should remain in the sealed pouch until use.

All specimens should be considered potentially hazardous and handled in the same manner as an infectious agent.

The used test should be discarded according to local regulations.

1.2.2 Storage and Stability

Store as packaged in the sealed pouch at room temperature or refrigerated (2-30℃). The test is stable through the expiration date printed on the sealed pouch. The test must remain in the sealed pouch until use. **DO NOT FREEZE.** Do not use beyond the expiration date.

1.2.3 Specimen Collection and Preparation

1.5.3.1 Urine Assay

The urine specimen should be collected in a clean and dry container. Urine specimens exhibiting visible precipitates should be centrifuged, filtered, or allowed to settle to obtain a clear specimen for testing.

1.5.3.2 Specimen Storage

Urine specimens may be stored at 2-8℃ for up to 48 hours prior to testing. For prolonged storage, specimens may be frozen and stored below -20℃. Frozen specimens should be thawed and mixed before testing.Specimens should not be frozen and thawed repeatedly.

1.2.4 Standard testing procedure

Allow the test, urine specimen, and/or controls to reach room temperature (15-30ºC) prior to testing.

1. Bring the pouch to room temperature before opening it. Remove the test panel from the sealed pouch and use it within one hour.
2. Remove the cap.
3. With the arrow pointing toward the urine specimen, immerse the test panel vertically in the urine specimen for at least 10 to 15 seconds. Immerse the dipstick to at least the level of the wavy lines, but not above the arrow on the test panel.
4. Replace the cap and place the test panel on a non-absorbent flat surface.
5. Start the timer and wait for the colored line(s) to appear.
6. The drug strip result should be read at 5 minutes. Do not interpret the result after 10 minutes.

1.2.5 Interpretation of results

NEGATIVE:* Two distinct colored lines appear. One line should be in the control line region (C) and another line should be in the test line region (T).

*NOTE: The intensity of the color in the test line region (T) may vary depending on the concentration of drug present in the specimen. Therefore, any shade of color in the test line region (T) should be considered negative.

POSITIVE: One colored line appears in the control line region (C). No apparent colored line appears in the test line region (T).

INVALID: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test. If the problem persists, discontinue using the test immediately and contact your local distributor.

1.2.6 Quality Control

A procedural control is included in the test. A color line appearing in the control region (C) is considered an internal procedural control. It confirms sufficient specimen volume and correct procedural technique.

Control standards are not supplied with this test; however it is recommended that positive and negative

controls be tested as good laboratory testing practices to confirm the test procedure and to verify proper test performance.

1.2.7 Limitations

The Multi-Drug Rapid Test Panel provides only a qualitative, preliminary analytical result. A secondary analytical method must be used to obtain a confirmed result. Gas chromatography/mass spectrometry (GC/MS) is the preferred confirmatory method.[1, 10]

There is a possibility that technical or procedural errors, as well as interfering substances in the urine specimen may cause erroneous results.

Adulterants, such as bleach and/or alum, in urine specimens may produce erroneous results regardless of the analytical method used. If adulteration is suspected, the test should be repeated with another urine specimen.

A positive result does not indicate level or intoxication, administration route or concentration in urine.

A negative result may not necessarily indicate drug-free urine. Negative results can be obtained when drug is present but below the cut-off level of the test.

This test does not distinguish between drugs of abuse and certain medications.

A positive test result may be obtained from certain foods or food supplements. Alcohol in the atmosphere, such as spray from perfumes, deodorizers, glass cleaners etc. can affect the Alcohol Rapid Tests. Therefore, adequate measures should be taken to avoid undue interference from such atmospheric agents in the testing area.

The test is only for detection of presence/ absence of alcohol in the urine, which may result from habitual drinking or medications and does not discriminate the two.

1.2.8 Reagents

Each test line contains anti-drug mouse monoclonal antibody and corresponding drug-protein conjugates. The control line contains goat anti-rabbit IgG polyclonal antibodies and rabbit IgG.

1.3 Composition of Product

- **MATERIALS**
  - **Absorbent Pad**: Filter paper AN3

    Thickness: 0.587mm

    Water absorption: 991.2g/m^2

    Composition: 100% Cotton fiber
  - **Label Pad**: 6613 Polyester film

    Weight: 99.86g/m^2

    Thickness: 16.7mils

    Composition: Polyester film
  - **Sample Pad**: Fiberglass

    Weight: 50.12g/ m^2

    Thickness: 12.53mils

    Composition: Fiberglass
  - **Nitrocellulose Membrane**: Sartorius CN140

    Thickness of membrane layer (μm): 120-160

    Capillary wicking rate along the roll (mm/5min): ＞52
  - **Backing Card**: Polystyrene (PS)

    Thickness of card: 0.455mm

    Composition: Polystyrene
- **MAJOR ANTIBODY**
  - **Control Zone**: Goat anti-Rabbit IgG (450~550ng/test)
  - **Test Zone**: Drug-BSA (300~600ng/test)
  - **Conjugate pad**: Mouse anti-Drug antibody (100~500ng/test) and Rabbit IgG (150~200ng/test)

1.4 Manufacturing Procedure

a) Dispense the colored drug antibodies and Rabbit IgG Gold conjugate on the label pad.
b) Use the sprayer to dispense capture reagent and Goat anti Rabbit IgG onto the membrane.
c) Assemble the membrane, label pad, absorbant pad, sample pad on the plastic backing.
d) Cut the plastic backing into strips of selected size.
e) Lay the strip into the panel, pack the panel and desiccant packet into a pouch and seal the pouch.
f) Test the panel according to the QC procedure and release the finished product.

1.5 User

Layuser

1.6 Classification of Product

Non-listed Devices of IVDD 98/79/EC

1.7 Product Certification Conformity Assessment Route

IVDD 98/79/EC Annex III

1.11 Bibliography

1. Hawks RL, CN Chiang. *Urine Testing for Drugs of Abuse*. National Institute for Drug Abuse (NIDA), Research Monograph 73, 1986.
2. Tietz NW. Textbook of Clinical Chemistry. W.B. Saunders Company. 1986; 1735.
3. Stewart DJ, Inaba T, Lucassen M, Kalow W. Clin. Pharmacol. Ther. April 1979; 25 ed: 464, 264-8.
4. Ambre J. J. Anal. Toxicol.1985; 9:241.
5. Winger, Gail, A Handbook of Drug and Alcohol Abuse, Third Edition, Oxford Press, 1992, page 146.
6. Robert DeCresce. *Drug Testing in the workplace,* 1989 page 114.
7. Glass, IB. The International Handbook of Addiction Behavior. Routledge Publishing, New York, NY. 1991; 216
8. B. Cody, J.T., "Specimen Adulteration in drug urinalysis. Forensic Sci. Rev., 1990, 2:63.
9. C. Tsai, S.C. et.al., J. Anal. Toxicol. 1998; 22 (6): 474
10. Baselt RC. Disposition of Toxic Drugs and Chemicals in Man. 6th Ed. Biomedical Publ., Foster City, CA 2002.
11. Hardman JG, Limbird LE. Goodman and Gilman's: The Pharmacological Basis for Therapeutics. 10th Edition. McGraw Hill Medical Publishing, 2001; 208-209.

2 Performance characteristics

2.1 Sample correlation

Evaluate the panel with approximately 250 specimens per drug type previously collected from subjects presenting for Drug Screen Testing which were confirmed by GC/MS. These specimens were randomized and tested using the Multi-Drug Rapid Test Panel. Specimens were rated as either positive or negative at 5minutes. The test results are shown in table below.

Table: Specimen Correlation

Method		GC/MS		% agreement with GC/MS
Multi-Drug Rapid Test Cup		Positive	Negative	
AMP 1,000	Positive	103	3	98.1%
	Negative	2	142	97.9%
AMP 500	Positive	110	2	99.1%
	Negative	1	137	98.6%
AMP 300	Positive	116	2	99.1%
	Negative	1	131	98.5%
BAR 300	Positive	98	2	96.1%
	Negative	4	146	98.6%
BAR 200	Positive	101	3	95.3%
	Negative	5	141	97.9%
BZO 500	Positive	112	3	98.2%
	Negative	2	133	97.8%
BZO 300	Positive	121	1	98.4%
	Negative	2	126	99.2%
BZO 200	Positive	127	2	99.2%
	Negative	1	120	98.4%
BZO 100	Positive	128	3	99.2%
	Negative	1	118	97.5%
BUP 10	Positive	105	0	99.1%
	Negative	1	144	>99.9%
BUP 5	Positive	105	0	99.1%
	Negative	1	144	>99.9%
COC 300	Positive	111	3	98.2%
	Negative	2	134	97.8%
COC 200	Positive	40	0	>99.9%
	Negative	0	60	>99.9%
COC 150	Positive	116	4	98.3%
	Negative	2	128	97.0%
COC 100	Positive	117	4	99.2%
	Negative	1	128	97.0%
THC 150	Positive	86	4	94.5%
	Negative	5	155	97.5%
THC 50	Positive	92	3	97.9%
	Negative	2	153	98.1%
THC 25	Positive	95	4	96.9%
	Negative	3	148	97.4%
THC 300	Positive	85	3	95.5%
	Negative	4	158	98.1%
THC 20	Positive	92	1	94.8%
	Negative	5	152	99.3%
THC	Positive	85	4	93.4%

Method		GC/MS		% agreement with GC/MS
Multi-Drug Rapid Test Cup		Positive	Negative	
200	Negative	6	155	97.4%
THC 30	Positive	94	3	97.9%
	Negative	2	151	98.1%
MTD 300	Positive	89	2	98.9%
	Negative	1	158	98.8%
MTD 200	Positive	91	2	98.9%
	Negative	1	156	98.7%
MET 1,000	Positive	76	5	96.2%
	Negative	3	166	97.1%
MET 500	Positive	83	5	97.6%
	Negative	2	160	97.0%
MET 300	Positive	88	4	97.8%
	Negative	2	156	97.5%
MDMA 1,000	Positive	99	1	98.0%
	Negative	2	148	99.3%
MDMA 500	Positive	102	1	98.1%
	Negative	2	145	99.3%
MDMA 300	Positive	103	1	98.1%
	Negative	2	144	99.3%
MOP 300	Positive	95	7	95.0%
	Negative	5	143	95.3%
MOP 200	Positive	95	6	95.0%
	Negative	5	144	96.0%
MOP 100	Positive	98	5	97.0%
	Negative	3	144	96.6%
MQL 300	Positive	79	11	89.8%
	Negative	9	151	93.2%
OPI 2000	Positive	117	8	96.7%
	Negative	4	121	93.8%
PCP 25	Positive	85	5	92.4%
	Negative	7	153	96.8%
PPX 300	Positive	97	9	96.0%
	Negative	4	140	94.0%
TCA 1000	Positive	91	13	94.8%
	Negative	5	141	91.6%
TCA 500	Positive	93	12	94.9%
	Negative	5	140	92.1%
TML 100	Positive	82	12	88.2%
	Negative	11	145	92.4%
TML 200	Positive	82	6	88.2%
	Negative	11	151	96.2%
TML 300	Positive	81	6	88.0%
	Negative	11	152	96.2%
KET 1,000	Positive	77	3	97.5%
	Negative	2	168	98.2%
KET 500	Positive	81	3	97.6%
	Negative	2	164	98.2%
KET 300	Positive	89	4	96.7%
	Negative	3	154	97.5%
KET 100	Positive	97	4	96.0%
	Negative	4	145	97.3%

Method		GC/MS		% agreement with GC/MS
Multi-Drug Rapid Test Cup		Positive	Negative	
OXY 100	Positive	84	1	97.7%
	Negative	2	163	99.4%
OXY 300	Positive	83	1	96.5%
	Negative	3	163	99.4%
COT 500	Positive	67	5	95.7%
	Negative	3	123	96.1%
COT 200	Positive	88	4	96.7%
	Negative	3	155	97.5%
COT 100	Positive	93	3	97.9%
	Negative	2	152	98.1%
COT 50	Positive	89	4	96.7%
	Negative	3	154	97.5%
COT 10	Positive	90	3	97.8%
	Negative	2	155	98.1%
EDDP 300	Positive	92	1	97.9%
	Negative	2	155	99.4%
EDDP 100	Positive	95	5	96.9%
	Negative	3	147	96.7%
FYL 20	Positive	29	1	96.7%
	Negative	1	92	98.9%
FYL 10	Positive	30	2	>99.9%
	Negative	0	91	97.8%
FYL 200	Positive	32	1	97.0%
	Negative	1	70	98.6%
FYL 100	Positive	33	1	97.1%
	Negative	1	78	98.7%
FYL 300	Positive	32	2	97.0%
	Negative	1	70	98.6%
K2 50	Positive	78	3	97.5%
	Negative	2	167	98.2%
K2 30	Positive	82	2	97.6%
	Negative	2	164	98.8%
K2 25	Positive	82	3	97.6%
	Negative	2	163	98.2%
6-MAM 10	Positive	42	2	97.7%
	Negative	1	105	98.1%
MDA 500	Positive	103	3	98.1%
	Negative	2	142	97.9%
ETG 300	Positive	79	1	98.8%
	Negative	1	169	99.4%
ETG 500	Positive	83	1	97.6%
	Negative	2	164	99.4%
ETG 1,000	Positive	81	1	95.3%
	Negative	4	164	99.4%
CLO 400	Positive	101	1	97.1%
	Negative	3	145	99.3%
CLO 150	Positive	103	2	99.0%
	Negative	1	144	98.6%
DIA 300	Positive	121	1	98.4%
	Negative	2	126	99.2%
DIA 200	Positive	121	1	98.4%

Method		GC/MS		% agreement with GC/MS
Multi-Drug Rapid Test Cup		Positive	Negative	
ACE 5000	Positive	29	1	93.5%
	Negative	2	68	98.6%
LSD 10	Positive	33	1	94.3%
	Negative	2	64	98.5%
LSD 20	Positive	33	1	94.3%
	Negative	2	64	98.5%
LSD 50	Positive	32	1	94.1%
	Negative	2	65	98.5%
MDPV 1000	Positive	28	1	93.3%
	Negative	2	69	98.6%
MDPV 500	Positive	27	1	93.1%
	Negative	2	59	98.3%
MDPV 300	Positive	30	2	93.8%
	Negative	2	68	97.1%
MPD 300	Positive	35	1	94.6%
	Negative	2	62	98.4%
MPD 1000	Positive	35	1	94.6%
	Negative	2	62	98.4%
ZOL 50	Positive	20	2	90.9%
	Negative	2	66	97.1%
ZOP 50	Positive	19	2	86.4%
	Negative	3	69	97.2%
MCAT 500	Positive	20	4	90.9%
	Negative	2	76	95.0%
7-ACL 300	Positive	32	1	94.1%
	Negative	2	43	97.7%
7-ACL 200	Positive	35	1	94.6%
	Negative	2	40	97.6%
7-ACL 100	Positive	36	1	94.7%
	Negative	2	39	97.5%
CFYL 500	Positive	36	1	94.7%
	Negative	2	72	98.6%
CFYL 250	Positive	36	1	94.7%
	Negative	2	72	98.6%
CAF 1000	Positive	21	3	91.3%
	Negative	2	66	95.7%
CAT 150	Positive	19	2	90.5%
	Negative	2	73	97.3%
TRO 350	Positive	23	2	92.0%
	Negative	2	64	97.0%
MEP 100	Positive	19	2	90.5%
	Negative	2	64	97.0%
ALP 100	Positive	20	2	90.9%
	Negative	2	74	97.4%
ABP 10	Positive	23	2	92.0%
	Negative	2	68	97.1%
α-PVP 300	Positive	35	3	92.1%
	Negative	3	59	95.2%
α-PVP 500	Positive	34	3	91.9%
	Negative	3	60	95.2%
α-PVP 1000	Positive	35	2	92.1%

Method		GC/MS		% agreement with GC/MS
Multi-Drug Rapid Test Cup		Positive	Negative	
	Negative	3	60	96.8%
CNB 500	Positive	23	1	95.8%
	Negative	1	40	97.6%
MPRD 100	Positive	19	3	95.0%
	Negative	1	49	94.2%
PGB 50000	Positive	20	2	90.9%
	Negative	2	73	97.3%
PGB 500	Positive	20	2	95.2%
	Negative	1	52	96.3%
TZD 200	Positive	26	3	92.9%
	Negative	2	73	96.1%
UR-144 25	Positive	34	1	97.1%
	Negative	1	62	98.4%
ZAL 100	Positive	20	1	95.2%
	Negative	1	38	97.4%
MES 100	Positive	23	1	95.8%
	Negative	1	40	97.6%
MES 300	Positive	23	1	95.8%
	Negative	1	40	97.6%
GAB 2000	Positive	24	1	92.3%
	Negative	2	65	98.5%
TLD 50	Positive	36	1	97.3%
	Negative	1	57	98.3%
QTP 1000	Positive	34	1	97.1%
	Negative	1	59	98.3%
PAP 500	Positive	31	1	96.8%
	Negative	1	49	98.0%
KAR 300	Positive	22	1	95.7%
	Negative	1	59	98.3%
CAR 2000	Positive	19	3	95.0%
	Negative	1	49	94.2%
FLX 500	Positive	33	2	97.1%
	Negative	1	57	96.6%
RPD150	Positive	28	3	93.3%
	Negative	2	63	95.5%
CIT500	Positive	28	3	93.3%
	Negative	2	63	95.5%
FKET1000	Positive	29	2	96.7%
	Negative	1	64	97.0%
TAP1000	Positive	34	1	94.4%
	Negative	2	55	98.2%
SCOP500	Positive	29	1	93.5%
	Negative	2	68	98.6%
MTZ500	Positive	28	3	93.3%
	Negative	2	65	95.6%
NND1000	Positive	29	2	96.7%
	Negative	1	64	97.0%

Conclusion: From the data above, it obtained statistically similar positive agreement, negative agreement and overall agreement rates as GC/MS data.

2.2 Analytical Sensitivity

The analytical sensitivity was determined by spiking negative urine specimens with intact standard at free urine (negative),-50% Cut-off, -25% Cut-off, Cut-off, +25% Cut-off, +50% Cut-off and +300% Cut-off. The standards were randomized and coded. The results were confirmed by GC/MS. A total of 10 replicates for each standard were tested. The specimens were tested with visual interpretations occurring at 5 minutes after specimen application. Results are presented in table below:

Table: Analytical Sensitivity Summary

AMP 1,000 Standard	AMP14030001p-T	AMP14030002p-T	AMP14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
1,000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1,250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3,000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

AMP 500 Standard	AMP14030007p-T	AMP14030008p-T	AMP14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

AMP 300 Standard	AMP14030013p-T	AMP14030014p-T	AMP14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BAR 300 Standard	BAR14030001p-T	BAR14030002p-T	BAR14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BAR 200 Standard	BAR14030007p-T	BAR14030008p-T	BAR14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BZO 500 Standard	BZO14030019p-T	BZO14030020p-T	BZO14030021p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BZO 300 Standard	BZO14030001p-T	BZO14030002p-T	BZO14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BZO 200 Standard	BZO14030007p-T	BZO 14030008p-T	BZO14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BZO 100 Standard	BZO14030013p-T	BZO14030014p-T	BZO14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BUP 10 Standard	BUP14030001p-T	BUP14030002p-T	BUP14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
7.5ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
10ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
12.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-

15ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

BUP 5 Standard	BUP14030007p-T	BUP14030008p-T	BUP14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
2.5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
3.75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
5ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
6.25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
7.5ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
15ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COC 300 Standard	COC14030001p-T	COC14030002p-T	COC14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COC 100 Standard	COC14030007p-T	COC14030008p-T	COC14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COC 150 Standard	COC14030013p-T	COC14030014p-T	COC14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
112.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
150ng/ml	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
187.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
225ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COC 200 Standard	COC14030019p-T	COC14030020p-T	COC14030021p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

THC 150 Standard	THC14030001p-T	THC14030002p-T	THC14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
112.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
150ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
187.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
225ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

THC 50 Standard	THC14030007p-T	THC14030008p-T	THC14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
37.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
50ng/ml	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
62.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

THC 25 Standard	THC14030013p-T	THC14030014p-T	THC14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
12.5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
18.75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
25ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
31.25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
37.5ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

THC 300 Standard	THC14030019p-T	THC14030020p-T	THC14030021p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

THC 20 Standard	THC14030037p-T	THC14030038p-T	THC14030039p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
10ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
15ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
20ng/ml	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
60ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

THC 200 Standard	THC14030025p-T	THC14030026p-T	THC14030027p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

THC 30 Standard	THC14030031p-T	THC14030032p-T	THC14030033p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
15ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
22.5ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
30ng/ml	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
37.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
45ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
90ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MTD 300 Standard	MTD14030001p-T	MTD14030002p-T	MTD14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MTD 200 Standard	MTD14030007p-T	MTD14030008p-T	MTD14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MET 1,000 Standard	MET14030001p-T	MET14030002p-T	MET14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
1,000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1,250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3,000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MET 500 Standard	MET14030007p-T	MET14030008p-T	MET14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MET 300 Standard	MET14030013p-T	MET14030014p-T	MET14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MDMA 1,000 Standard	MDMA14030007p-T	MDMA14030008p-T	MDMA 14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
1,000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1,250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3,000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MDMA 500 Standard	MDMA14030001p-T	MDMA14030002p-T	MDMA14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MDMA 300 Standard	MDMA14030013p-T	MDMA14030014p-T	MDMA14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MOP 300 Standard	MOP14030001p-T	MOP14030002p-T	MOP14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+

375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MOP 100 Standard	MOP14030007p-T	MOP14030008p-T	MOP14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MOP 200 Standard	MOP14030013p-T	MOP14030014p-T	MOP14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MQL 300 Standard	MQL14030001p-T	MQL14030002p-T	MQL14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

OPI 2000 Standard	OPI14030001p-T	OPI14030002p-T	OPI14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
1,000ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
1,500ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
2,000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
2,500ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
3,000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
6,000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

PCP 25 Standard	PCP14030001p-T	PCP14030002p-T	PCP14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
12.5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
18.75ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
25ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
31.25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
37.5ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

PPX 300 Standard	PPX14030001p-T	PPX14030002p-T	PPX14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TCA 1000 Standard	TCA14030001p-T	TCA14030002p-T	TCA14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
1,000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1,250ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3,000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TCA 500 Standard	TCA14030007p-T	TCA14030008p-T	TCA14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TML 100 Standard	TML14030001p-T	TML14030002p-T	TML14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TML 200 Standard	TML14030007p-T	TML14030008p-T	TML14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
400ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TML 300 Standard	TML14030013p-T	TML14030014p-T	TML14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

KET 1,000 Standard	KET14030001p-T	KET14030002p-T	KET14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
1,000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1,250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3,000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

KET 500 Standard	KET14030007p-T	KET14030008p-T	KET14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

KET 300 Standard	KET14030013p-T	KET14030014p-T	KET14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

KET 100 Standard	KET 14030019p-T	KET 14030020p-T	KET 14030021p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

OXY 100 Standard	OXY14030001p-T	OXY 14030002p-T	OXY 14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+

100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

OXY 300 Standard	OXY14030007p-T	OXY14030008p-T	OXY14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COT 500 Standard	COT14030013p-T	COT 14030014p-T	COT 14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
325ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COT 200 Standard	COT14030001p-T	COT 14030002p-T	COT 14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COT 100 Standard	COT 14030007p-T	COT 14030008p-T	COT 14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COT 50 Standard	COT14030019p-T	COT 14030020p-T	COT 14030021p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
32.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
50ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+
62.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

COT 10 Standard	COT14030025p-T	COT 14030026p-T	COT 14030027p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
7.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
10ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
12.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
15ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

EDDP 300 Standard	EDDP14030001p-T	EDDP 14030002p-T	EDDP 14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

EDDP 100 Standard	EDDP14030007p-T	EDDP14030008p-T	EDDP14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

FYL 20 Standard	FYL18040001p-T	FYL18040002p-T	FYL18040003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
10ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
15ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
20ng/ml	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
60ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

FYL 10 Standard	FYL18040007p-T	FYL18040008p-T	FYL18040009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
7.5ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
10ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
12.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
15ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

FYL 100 Standard	FYL18040013p-T	FYL18040014p-T	FYL18040015p-T

Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

FYL 200 Standard	**FYL18040019p-T**	**FYL18040020p-T**	**FYL18040021p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

FYL 300 Standard	**FYL18040025p-T**	**FYL18040026p-T**	**FYL18040027p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

Synthetic Marijuana (K2) 50 Standard	**SMA14030001p-T**	**SMA14030002p-T**	**SMA14030003p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
37.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
50ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
62.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

K2 30 Standard	**SMA14030007p-T**	**SMA14030008p-T**	**SMA14030009p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
15ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
22.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
30ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+
37.5ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
45ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
90ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

Synthetic Marijuana ((K2) 25 Standard	**SMA14030013p-T**	**SMA14030014p-T**	**SMA14030015p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
12.5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
18.75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
25ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+

31.25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
37.5ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

6-MAM Standard	MAM14030001p-T	MAM14030002p-T	MAM14030003p-T
Free urine	-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-
5ng/ml	-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-
7.5ng/ml	-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+
10ng/ml	-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+
12.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
15ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CLO 400 Standard	CLO14030001p-T	CLO14030002p-T	CLO14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
200ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
300ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+
400ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+
500ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1200ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CLO 150 Standard	CLO14030007p-T	CLO14030008p-T	CLO14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
112.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
150ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+
187.5ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
225ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ETG 300 Standard	ETG14030013p-T	ETG14030014p-T	ETG14030015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ETG 500 Standard	ETG14030001p-T	ETG14030002p-T	ETG14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ETG 1000 Standard	ETG14030007p-T	ETG14030008p-T	ETG14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
1500ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

DIA 300 Standard	DIA14030001p-T	DIA14030002p-T	DIA14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

DIA 200 Standard	DIA14030007p-T	DIA14030008p-T	DIA14030009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,+
600ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MDA 500 Standard	MDA14030001p-T	MDA14030002p-T	MDA14030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1,500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ACE5000 Standard	ACE14060001p-T	ACE14060002p-T	ACE14060003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
2500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
3750ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
5000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
6250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
7500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
15000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

LSD 10 Standard	LSD14060013p-T	LSD14060014p-T	LSD14060015p-T

Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
7.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
10ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
12.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
15ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

LSD 20 Standard	**LSD14060001p-T**	**LSD14060002p-T**	**LSD14060003p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
10ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
15ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
20ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
60ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

LSD 50 Standard	**LSD14060007p-T**	**LSD14060008p-T**	**LSD14060009p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
37.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
50ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
62.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MDPV1000 Standard	**MDPV14080001p-T**	**MDPV14080002p-T**	**MDPV14080003p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MDPV500 Standard	**MDPV14080007p-T**	**MDPV14080008p-T**	**MDPV14080009p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
325ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MDPV300 Standard	**MDPV15090001p-T**	**MDPV15090002p-T**	**MDPV15090003p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MPD300 Standard	MPD14060001p-T	MPD14060002p-T	MPD14060003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
325ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MPD1000 Standard	MPD14060007p-T	MPD14060008p-T	MPD14060009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ZOL 50 Standard	ZOL15010001p-T	ZOL15010002p-T	ZOL15010003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
37.5ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
50ng/ml	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
62.5ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ZOP 50 Standard	ZOP16080001p-T	ZOP16080002p-T	ZOP16080003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
37.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
50ng/ml	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
62.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MCAT 500 Standard	MCAT16080001p-T	MCAT16080002p-T	MCAT16080003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,-,+,+,+,+

625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

7-ACL 300 Standard	ACL14100001p-T	ACL14100002p-T	ACL14100003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150 ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225 ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300 ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375 ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
450 ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900 ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

7-ACL 100 Standard	ACL14100007p-T	ACL14100008p-T	ACL14100009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50 ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75 ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100 ng/ml	-,-,-,-,+,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125 ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
150 ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300 ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

7-ACL 200 Standard	ACL14100013p-T	ACL14100014p-T	ACL14100015p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100 ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150 ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
200 ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250 ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
300 ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
600 ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CFYL 500 Standard	CFYL15020001p-T	CFYL15020002p-T	CFYL15020003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
500ng/mL	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CFYL 250 Standard	CFYL15020007p-T	CFYL15020008p-T	CFYL15020009p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
125ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
187.5ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
250ng/mL	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
312.5ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
375ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CAF 1000 Standard	CAF15080001p-T	CAF15080002p-T	CAF15080003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,-,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CAT 150 Standard	CAT14080001p-T	CAT14080002p-T	CAT14080003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
112.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
150ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,-,+,+,+,+
187.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
225ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TRO 350 Standard	TRO15050001p-T	TRO15050002p-T	TRO15050003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
175ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
262.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
350ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
437.5ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
525ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1050ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MEP 100 Standard	MEP15070001p-T	MEP15070002p-T	MEP15070003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,-,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ALP 100 Standard	ALP16040001p-T	ALP16040002p-T	ALP16040003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,-,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

AB-PINACA 10 Standard	ABP17010001p-T	ABP17010002p-T	ABP17010003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
7.5ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
10ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
12.5ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
15ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
30ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

α-PVP1000 Standard	**APVP15050001p-T**	**APVP15050002p-T**	**APVP15050003p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

α-PVP500 Standard	**APVP15050007p-T**	**APVP15050008p-T**	**APVP15050009p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

α-PVP300 Standard	**APVP15050013p-T**	**APVP15050014p-T**	**APVP15050015p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CNB 500 Standard	**CNB18060001p-T**	**CNB18060002p-T**	**CNB18060003p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
325ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MPRD 100 Standard	**MPRD18060001p-T**	**MPRD18060002p-T**	**MPRD18060003p-T**
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,+,-	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+

125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

PGB 50000 Standard	PGB16010001p-T	PGB16010002p-T	PGB16010003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ug/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
37.5ug/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
50ug/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
62.5ug/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-
75ug/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
150ug/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

PGB 500 Standard	PGB19040001p-T	PGB19040002p-T	PGB19040003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TZD Standard	TZD18070001p-T	TZD18070002p-T	TZD18070003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
100ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,-
200ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

UR-144 Standard	UR18060001p-T	UR18060002p-T	UR18060003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
12.5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
18.75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
25ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
31.25ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
37.5ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

ZAL Standard	ZAL18050001p-T	ZAL18050002p-T	ZAL18050003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MES 100 Standard	MES18060001p-T	MES18060002p-T	MES18060003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
50ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
100ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
125ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
300ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MES 300 Standard	MES18090001p-T	MES18090002p-T	MES18090003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,-,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

GAB 2000 Standard	GAB18070001p-T	GAB18070002p-T	GAB18070003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
1000ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
1500ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,-
2000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
2500ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
6000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TLD 50 Standard	TLD18110001p-T	TLD18110002p-T	TLD18110003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
25ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
37.5ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
50ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
62.5ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,-
75ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
150ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

QTP 1000 Standard	QTP19030001p-T	QTP19030002p-T	QTP19030003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

PAP 500 Standard	PAP18110001p-T	PAP18110002p-T	PAP18110003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

375ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

KRA 300 Standard	KRA19040001p-T	KRA19040002p-T	KRA19040003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
225ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
300ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
375ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,-
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
900ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CAR 2000 Standard	CAR18100001p-T	CAR18100002p-T	CAR18100003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
1000ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
1500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,-,+
2000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,-,+,+,+,+
2500ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
6000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

FlX 500 Standard	FLX18110001p-T	FLX18110002p-T	FLX18110003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

RPD 150 Standard	RPD20090001p-T	RPD20090002p-T	RPD20090003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
75ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
112.5ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,-
150ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
187.5ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
225ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
450ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

CIT 500 Standard	CIT20120001p-T	CIT20120002p-T	CIT20120003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

FKET 1000 Standard	FKE21010001p-T	FKE21010002p-T	FKE21010003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,-
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

TAP 1000 Standard	TAP21090001p-T	TAP21090002p-T	TAP21090003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,-
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

SCOP 500 Standard	SCOP21100004-T	SCOP21100005-T	SCOP21100006-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,+
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,+,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

MTZ 500 Standard	MTZ21110007-T	MTZ21110008-T	MTZ21110009-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
250ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
375ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
625ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
750ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

NND 1000 Standard	NND21120001p-T	NND21120002p-T	NND21120003p-T
Free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
500ng/ml	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
750ng/ml	-,-,-,-,-,-,-,-,-,+	-,-,-,-,-,-,-,-,+,+	-,-,-,-,-,-,-,-,-,-
1000ng/ml	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+	-,-,-,-,-,+,+,+,+,+
1250ng/ml	+,+,+,+,+,+,+,+,-,-	+,+,+,+,+,+,+,+,+,-	+,+,+,+,+,+,+,+,+,-
1500ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
3000ng/ml	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

Conclusion: The results indicated that each drug in the Multi-Drug Rapid Test Panel could detect above the cut-off concentration.

2.3 Analytical Specificity

The related compounds of each single drug in panel will be spiked in drug-free urine, diluted sequentially to different concentrations and tested until the lowest concentration that has a negative result. The following compounds have negative result at their respective concentrations. And the cutoff concentrations were showed in the tables with "*". The specimens were tested in 5 replicates with visual interpretations occurring at 5 minutes after specimen application. Results are presented in table below.

Table: Interfering Substances

AMP 1,000:

Analytes	Conc. (ng/ml)	AMP14030001p-T		AMP14030002p-T		AMP14030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
D,L-Amphetamine sulfate	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
L-Amphetamine	12,500	5	0	5	0	5	0
	*25,000	2	3	2	3	2	3
	37,500	0	5	0	5	0	5
(±) 3,4-Methylenedioxy-amphetamine	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Phentermine	500	5	0	5	0	5	0
	*1000	2	3	3	2	2	3
	1500	0	5	0	5	0	5
Maprotiline	25,000	5	0	5	0	5	0
	*50,000	4	1	3	2	2	3
	75,000	0	5	0	5	0	5
Methoxyphenamine	3,000	5	0	5	0	5	0
	*6,000	2	3	3	2	2	3
	9,000	0	5	0	5	0	5
D-Amphetamine	500	5	0	5	0	5	0
	*1,000	2	3	3	2	2	3
	1,500	0	5	0	5	0	5

AMP 500:

Analytes	Conc. (ng/ml)	AMP14030007p-T		AMP14030008p-T		AMP14030009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
D,L-Amphetamine sulfate	75	5	0	5	0	5	0
	*150	3	2	3	2	2	3

	225	0	5	0	5	0	5
L-Amphetamine	6,250	5	0	5	0	5	0
	*12,500	2	3	2	3	2	3
	18,750	0	5	0	5	0	5
(±) 3,4-Methylenedioxy-amphetamine	125	5	0	5	0	5	0
	*250	2	3	3	2	2	3
	375	0	5	0	5	0	5
Phentermine	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Maprotiline	12,500	5	0	5	0	5	0
	*25,000	4	1	3	2	2	3
	37,500	0	5	0	5	0	5
Methoxyphenamine	1,500	5	0	5	0	5	0
	*3,000	2	3	3	2	2	3
	4,500	0	5	0	5	0	5
D-Amphetamine	250	5	0	5	0	5	0
	*500	2	3	2	3	2	3
	750	0	5	0	5	0	5

AMP 300:

Analytes	**Conc. (ng/ml)**	**AMP14030013p-T**		**AMP14030014p-T**		**AMP14030015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
D,L-Amphetamine sulfate	37.5	5	0	5	0	5	0
	*75	3	2	3	2	2	3
	112.5	0	5	0	5	0	5
L-Amphetamine	5,000	5	0	5	0	5	0
	*10,000	2	3	2	3	2	3
	15,000	0	5	0	5	0	5
(±) 3,4-Methylenedioxy-amphetamine	75	5	0	5	0	5	0
	*150	2	3	3	2	2	3
	225	0	5	0	5	0	5
Phentermine	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Maprotiline	6,250	5	0	5	0	5	0

	*15,000	4	1	3	2	2	3
	18,750	0	5	0	5	0	5
Methoxyphenamine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
D-Amphetamine	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5

BAR 300:

Analytes	**Conc. (ng/ml)**	**BAR14030001p-T**		**BAR14030002p-T**		**BAR14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Amobarbital	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
5,5-Diphenylhydantoin	4,000	5	0	5	0	5	0
	*8,000	2	3	3	2	2	3
	12,000	0	5	0	5	0	5
Allobarbital	300	5	0	5	0	5	0
	*600	3	2	3	2	2	3
	900	0	5	0	5	0	5
Barbital	4,000	5	0	5	0	5	0
	*8,000	2	3	2	3	2	3
	12,000	0	5	0	5	0	5
Talbutal	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Cyclopentobarbital	15,000	5	0	5	0	5	0
	*30,000	2	3	3	2	2	3
	45,000	0	5	0	5	0	5
Pentobarbital	4,000	5	0	5	0	5	0
	*8,000	2	3	3	2	3	2
	12,000	0	5	0	5	0	5
Alphenol	300	5	0	5	0	5	0
	*600	2	3	3	2	2	3
	900	0	5	0	5	0	5

Aprobarbital	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Butabarbital	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Butalbital	4,000	5	0	5	0	5	0
	*8,000	2	3	3	2	2	3
	12,000	0	5	0	5	0	5
Butethal	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Phenobarbital	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Secobarbital	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5

BAR 200:

Analytes	**Conc. (ng/ml)**	**BAR14030007p-T**		**BAR14030008p-T**		**BAR14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Amobarbital	1,500	5	0	5	0	5	0
	*3,000	2	3	3	2	2	3
	4,500	0	5	0	5	0	5
5,5-Diphenylhydantoin	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
Allobarbital	200	5	0	5	0	5	0
	*400	1	4	3	2	2	3
	600	0	5	0	5	0	5
Barbital	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5

Talbutal	75	5	0	5	0	5	0
	*150	2	3	3	2	2	3
	225	0	5	0	5	0	5
Cyclopentobarbital	10,000	5	0	5	0	5	0
	*20,000	2	3	3	2	2	3
	30,000	0	5	0	5	0	5
Pentobarbital	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	3	2
	7,500	0	5	0	5	0	5
Alphenol	200	5	0	5	0	5	0
	*400	2	3	3	2	2	3
	600	0	5	0	5	0	5
Aprobarbital	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Butabarbital	75	5	0	5	0	5	0
	*150	2	3	3	2	2	3
	225	0	5	0	5	0	5
Butalbital	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
Butethal	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Phenobarbital	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Secobarbital	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5

BUP10:

Analytes	**Conc. (ng/ml)**	**BUP14030001p-T**		**BUP14030002p-T**		**BUP14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Buprenorphine	5	5	0	5	0	5	0

	*10	3	2	3	2	2	3
	15	0	5	0	5	0	5
Buprenorphine 3-D-Glucuronide	25	5	0	5	0	5	0
	*50	2	3	2	3	2	3
	75	0	5	0	5	0	5
Norbuprenorphine	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5
Norbuprenorphine 3-D-Glucuronide	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5

BUP 5

Analytes	**Conc. (ng/ml)**	**BUP14030007p-T**		**BUP14030008p-T**		**BUP14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Buprenorphine	2.5	5	0	5	0	5	0
	*5	3	2	3	2	2	3
	7.5	0	5	0	5	0	5
Buprenorphine 3-D-Glucuronide	12.5	5	0	5	0	5	0
	*25	2	3	2	3	2	3
	37.5	0	5	0	5	0	5
Norbuprenorphine	12.5	5	0	5	0	5	0
	*25	3	2	3	2	2	3
	37.5	0	5	0	5	0	5
Norbuprenorphine 3-D-Glucuronide	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5

BZO 500:

Analytes	**Conc. (ng/ml)**	**BZO14030019p-T**		**BZO14030020p-T**		**BZO14030021p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Alprazolam	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
a-hydroxyalpra	1,250	5	0	5	0	5	0

zolam	*2,500	2	3	2	3	3	2
	3,750	0	5	0	5	0	5
Bromazepam	750	5	0	5	0	5	0
	*1,500	2	3	3	2	2	3
	2,250	0	5	0	5	0	5
Chlordiazepoxide	750	5	0	5	0	5	0
	*1,500	3	2	2	3	2	3
	2,250	0	5	0	5	0	5
Clobazam	150	5	0	5	0	5	0
	*300	3	2	2	3	2	3
	450	0	5	0	5	0	5
Clonazepam	400	5	0	5	0	5	0
	*800	3	2	2	3	2	3
	1,200	0	5	0	5	0	5
Clorazepate dipotassium	400	5	0	5	0	5	0
	*800	2	3	2	3	2	3
	1,200	0	5	0	5	0	5
Delorazepam	750	5	0	5	0	5	0
	*1,500	2	3	2	3	2	3
	2,250	0	5	0	5	0	5
Desalkylflurazepam	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
Flunitrazepam	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
(±) Lorazepam	2,500	5	0	5	0	5	0
	*5,000	2	3	2	3	2	3
	7,500	0	5	0	5	0	5
RS-Lorazepam glucuronide	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
Midazolam	5,000	5	0	5	0	5	0
	*10,000	3	2	2	3	2	3
	15,000	0	5	0	5	0	5

Nitrazepam	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Norchlordiazepoxide	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
Nordiazepam	750	5	0	5	0	5	0
	*1,500	3	2	2	3	2	3
	2,250	0	5	0	5	0	5
Oxazepam	250	5	0	5	0	5	0
	*500	2	3	2	3	2	3
	750	0	5	0	5	0	5
Temazepam	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Diazepam	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Estazolam	5,000	5	0	5	0	5	0
	*10,000	3	2	3	2	2	3
	15,000	0	5	0	5	0	5
Triazolam	2,500	5	0	5	0	5	0
	*5,000	3	2	3	2	2	3
	7,500	0	5	0	5	0	5

BZO 300:

Analytes	**Conc. (ng/ml)**	**BZO14030001p-T**		**BZO14030002p-T**		**BZO14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Alprazolam	50	5	0	5	0	5	0
	*100	2	3	2	3	2	3
	150	0	5	0	5	0	5
a-hydroxyalprazolam	750	5	0	5	0	5	0
	*1,500	2	3	2	3	3	2
	2,250	0	5	0	5	0	5
Bromazepam	450	5	0	5	0	5	0
	*900	2	3	3	2	2	3

	1,350	0	5	0	5	0	5
Chlordiazepoxide	450	5	0	5	0	5	0
	*900	3	2	2	3	2	3
	1,350	0	5	0	5	0	5
Clobazam	100	5	0	5	0	5	0
	*200	3	2	2	3	2	3
	300	0	5	0	5	0	5
Clonazepam	250	5	0	5	0	5	0
	*500	3	2	2	3	2	3
	750	0	5	0	5	0	5
Clorazepate dipotassium	250	5	0	5	0	5	0
	*500	2	3	2	3	2	3
	750	0	5	0	5	0	5
Delorazepam	450	5	0	5	0	5	0
	*900	2	3	2	3	2	3
	1,350	0	5	0	5	0	5
Desalkylflurazepam	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
Flunitrazepam	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
(±) Lorazepam	1,500	5	0	5	0	5	0
	*3,000	2	3	2	3	2	3
	4,500	0	5	0	5	0	5
RS-Lorazepam glucuronide	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
Midazolam	3,000	5	0	5	0	5	0
	*6,000	3	2	2	3	2	3
	9,000	0	5	0	5	0	5
Nitrazepam	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Norchlordiaze	50	5	0	5	0	5	0

poxide	*100	2	3	2	3	2	3
	150	0	5	0	5	0	5
Nordiazepam	450	5	0	5	0	5	0
	*900	3	2	2	3	2	3
	1,350	0	5	0	5	0	5
Oxazepam	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
Temazepam	50	5	0	5	0	5	0
	*100	2	3	3	2	2	3
	150	0	5	0	5	0	5
Diazepam	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Estazolam	3,000	5	0	5	0	5	0
	*6,000	3	2	3	2	2	3
	9,000	0	5	0	5	0	5
Triazolam	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5

BZO 200:

Analytes	**Conc. (ng/ml)**	**BZO14030007p-T**		**BZO14030008p-T**		**BZO14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Alprazolam	35	5	0	5	0	5	0
	*70	2	3	2	3	2	3
	105	0	5	0	5	0	5
a-hydroxyalprazolam	500	5	0	5	0	5	0
	*1,000	2	3	2	3	3	2
	1,500	0	5	0	5	0	5
Bromazepam	300	5	0	5	0	5	0
	*600	2	3	3	2	2	3
	900	0	5	0	5	0	5
Chlordiazepoxide	300	5	0	5	0	5	0
	*600	3	2	2	3	2	3
	900	0	5	0	5	0	5

Clobazam	60	5	0	5	0	5	0
	*120	3	2	2	3	2	3
	180	0	5	0	5	0	5
Clonazepam	150	5	0	5	0	5	0
	*300	3	2	2	3	2	3
	450	0	5	0	5	0	5
Clorazepate dipotassium	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
Delorazepam	300	5	0	5	0	5	0
	*600	2	3	2	3	2	3
	900	0	5	0	5	0	5
Desalkylflurazepam	60	5	0	5	0	5	0
	*120	2	3	2	3	2	3
	180	0	5	0	5	0	5
Flunitrazepam	60	5	0	5	0	5	0
	*120	2	3	2	3	2	3
	180	0	5	0	5	0	5
(±) Lorazepam	1,000	5	0	5	0	5	0
	*2,000	2	3	2	3	2	3
	3,000	0	5	0	5	0	5
RS-Lorazepam glucuronide	60	5	0	5	0	5	0
	*120	2	3	2	3	2	3
	180	0	5	0	5	0	5
Midazolam	2,000	5	0	5	0	5	0
	*4,000	3	2	2	3	2	3
	6,000	0	5	0	5	0	5
Nitrazepam	60	5	0	5	0	5	0
	*120	2	3	3	2	2	3
	180	0	5	0	5	0	5
Norchlordiazepoxide	35	5	0	5	0	5	0
	*70	2	3	2	3	2	3
	105	0	5	0	5	0	5
Nordiazepam	300	5	0	5	0	5	0
	*600	3	2	2	3	2	3

	900	0	5	0	5	0	5
Oxazepam	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
Temazepam	35	5	0	5	0	5	0
	*70	2	3	3	2	2	3
	105	0	5	0	5	0	5
Diazepam	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Estazolam	2,000	5	0	5	0	5	0
	*4,000	3	2	3	2	2	3
	6,000	0	5	0	5	0	5
Triazolam	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5

BZO 100:

Analytes	**Conc. (ng/ml)**	**BZO14030013p-T**		**BZO14030014p-T**		**BZO14030015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Alprazolam	20	5	0	5	0	5	0
	*40	2	3	2	3	2	3
	60	0	5	0	5	0	5
a-hydroxyalprazolam	250	5	0	5	0	5	0
	*500	2	3	2	3	3	2
	750	0	5	0	5	0	5
Bromazepam	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Chlordiazepoxide	150	5	0	5	0	5	0
	*300	3	2	2	3	2	3
	450	0	5	0	5	0	5
Clobazam	30	5	0	5	0	5	0
	*60	3	2	2	3	2	3
	90	0	5	0	5	0	5
Clonazepam	75	5	0	5	0	5	0

	*150	3	2	2	3	2	3
	225	0	5	0	5	0	5
Clorazepate dipotassium	75	5	0	5	0	5	0
	*150	2	3	2	3	2	3
	225	0	5	0	5	0	5
Delorazepam	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
Desalkylflurazepam	30	5	0	5	0	5	0
	*60	2	3	2	3	2	3
	90	0	5	0	5	0	5
Flunitrazepam	30	5	0	5	0	5	0
	*60	2	3	2	3	2	3
	90	0	5	0	5	0	5
(±) Lorazepam	500	5	0	5	0	5	0
	*1,000	2	3	2	3	2	3
	1,500	0	5	0	5	0	5
RS-Lorazepam glucuronide	30	5	0	5	0	5	0
	*60	2	3	2	3	2	3
	90	0	5	0	5	0	5
Midazolam	1,000	5	0	5	0	5	0
	*2,000	3	2	2	3	2	3
	3,000	0	5	0	5	0	5
Nitrazepam	30	5	0	5	0	5	0
	*60	2	3	3	2	2	3
	90	0	5	0	5	0	5
Norchlordiazepoxide	20	5	0	5	0	5	0
	*40	2	3	2	3	2	3
	60	0	5	0	5	0	5
Nordiazepam	150	5	0	5	0	5	0
	*300	3	2	2	3	2	3
	450	0	5	0	5	0	5
Oxazepam	50	5	0	5	0	5	0
	*100	2	3	2	3	2	3

	150	0	5	0	5	0	5
Temazepam	20	5	0	5	0	5	0
	*40	2	3	3	2	2	3
	60	0	5	0	5	0	5
Diazepam	50	5	0	5	0	5	0
	*100	2	3	3	2	2	3
	150	0	5	0	5	0	5
Estazolam	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Triazolam	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5

COC 300:

Analytes	**Conc. (ng/ml)**	**COC14030001p-T**		**COC14030002p-T**		**COC14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Benzoylecgonine	150	5	0	5	0	5	0
	*300	1	4	3	2	2	3
	450	0	5	0	5	0	5
Cocaine HCl	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
Cocaethylene	10,000	5	0	5	0	5	0
	*20,000	4	1	1	4	2	3
	30,000	0	5	0	5	0	5
Ecgonine	15,000	5	0	5	0	5	0
	*30,000	3	2	3	2	2	3
	45,000	0	5	0	5	0	5

COC 100:

Analytes	**Conc. (ng/ml)**	**COC14030007p-T**		**COC14030008p-T**		**COC14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Benzoylecgonine	50	5	0	5	0	5	0
	*100	2	3	3	2	2	3

	150	0	5	0	5	0	5
Cocaine HCl	40	5	0	5	0	5	0
	*80	3	2	3	2	2	3
	120	0	5	0	5	0	5
Cocaethylene	3,500	5	0	5	0	5	0
	*7,000	2	3	3	2	2	3
	10,500	0	5	0	5	0	5
Ecgonine	5,000	5	0	5	0	5	0
	*10,000	1	4	3	2	2	3
	15,000	0	5	0	5	0	5

COC 150

Analytes	**Conc. (ng/ml)**	**COC14030013p-T**		**COC14030014p-T**		**COC14030015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Benzoylecgonine	75	5	0	5	0	5	0
	*150	2	3	4	1	2	3
	225	0	5	0	5	0	5
Cocaine HCl	60	5	0	5	0	5	0
	*120	3	2	3	2	3	2
	180	0	5	0	5	0	5
Cocaethylene	5,000	5	0	5	0	5	0
	*1,000	4	1	3	2	2	3
	15,000	0	5	0	5	0	5
Ecgonine	7,500	5	0	5	0	5	0
	*15,000	3	2	3	2	2	3
	22,500	0	5	0	5	0	5

COC 200

Analytes	**Conc. (ng/ml)**	**COC14030019p-T**		**COC14030020p-T**		**COC14030021p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Benzoylecgonine	100	5	0	5	0	5	0
	*200	1	4	3	2	2	3

	300	0	5	0	5	0	5
Cocaine HCl	67.5	5	0	5	0	5	0
	*135	3	2	3	2	2	3
	202.5	0	5	0	5	0	5
Cocaethylene	6,750	5	0	5	0	5	0
	*1,3500	4	1	3	2	2	3
	20,250	0	5	0	5	0	5
Ecgonine	10,500	5	0	5	0	5	0
	*20,000	2	3	3	2	2	3
	30,000	0	5	0	5	0	5

KET1000

Analytes	**Conc. (ng/ml)**	**KET14030001p-T**		**KET14030002p-T**		**KET14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Ketamine	500	5	0	5	0	5	0
	*1,000	2	3	3	2	2	3
	1,500	0	5	0	5	0	5
Benzphetamine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
(+) Chlorpheniramine	12,500	5	0	5	0	5	0
	*25,000	3	2	3	2	2	3
	37,500	0	5	0	5	0	5
Clonidine	50,000	5	0	5	0	5	0
	*100,000	2	3	2	3	2	3
	150,000	0	5	0	5	0	5
Dextromethorp	1,000	5	0	5	0	5	0

han	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Disopyramide	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
EDDP	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	3	2
	75,000	0	5	0	5	0	5
Mephentermine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
(1R, 2S) -(-)-Ephedrine	50,000	5	0	5	0	5	0
	*100,000	2	3	3	2	2	3
	150,000	0	5	0	5	0	5
4-Hydroxyphencyclidine	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
Levorphanol	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
MDE	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
Tetrahydrozoline	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
d-Methamphetamine	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3

	75,000	0	5	0	5	0	5
l-Methamphetamine	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
Methoxyphenamine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
(+)3,4-Methylendioxymethamphetamine (MDMA)	50,000	5	0	5	0	5	0
	*100,000	2	3	3	2	2	3
	150,000	0	5	0	5	0	5
d-Norpropoxyphene	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Pentazocine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Phencyclidine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Promazine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Promethazine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Thioridazine	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3

	75,000	0	5	0	5	0	5
Meperidine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5

KET 500:

Analytes	**Conc. (ng/ml)**	**KET14030007p-T**		**KET14030008p-T**		**KET14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Ketamine	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Benzphetamine	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
(+) Chlorpheniramine	6,250	5	0	5	0	5	0
	*12,500	3	2	3	2	2	3
	18,750	0	5	0	5	0	5
Clonidine	25,000	5	0	5	0	5	0
	*50,000	2	3	2	3	2	3
	75,000	0	5	0	5	0	5
Dextromethorphan	500	5	0	5	0	5	0
	*1,000	2	3	3	2	2	3
	1,500	0	5	0	5	0	5
Disopyramide	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
EDDP	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	3	2
	37,500	0	5	0	5	0	5
Mephentermin	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3

e	18,750	0	5	0	5	0	5
(1R, 2S) -(-)-Ephedrine	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
4-Hydroxyphencyclidine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Levorphanol	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
MDE	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Tetrahydrozoline	125	5	0	5	0	5	0
	*250	2	3	3	2	2	3
	325	0	5	0	5	0	5
d-Methamphetamine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
l -Methamphetamine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Methoxyphenamine	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
(+)3,4-Methylendioxymethamphetamine (MDMA)	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	3	2
	75,000	0	5	0	5	0	5

d-Norpropoxyphene	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
Pentazocine	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
Phencyclidine	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
Promazine	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
Promethazine	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5
Thioridazine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Meperidine	6,250	5	0	5	0	5	0
	*12,500	2	3	3	2	2	3
	18,750	0	5	0	5	0	5

KET 300:

Analytes	**Conc. (ng/ml)**	**KET14030013p-T**		**KET14030014p-T**		**KET14030015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Ketamine	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Benzphetamine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5

(+) Chlorpheniramine	3,125	5	0	5	0	5	0
	*6,250	3	2	3	2	2	3
	9,375	0	5	0	5	0	5
Clonidine	15,000	5	0	5	0	5	0
	*30,000	2	3	2	3	2	3
	45,000	0	5	0	5	0	5
Dextromethorphan	300	5	0	5	0	5	0
	*600	2	3	3	2	2	3
	900	0	5	0	5	0	5
Disopyramide	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5
EDDP	7,500	5	0	5	0	5	0
	*15,000	2	3	3	2	3	2
	22,500	0	5	0	5	0	5
Mephentermine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5
(1R, 2S) - (-)-Ephedrine	15,000	5	0	5	0	5	0
	*30,000	2	3	3	2	2	3
	45,000	0	5	0	5	0	5
4-Hydroxyphencyclidine	7,500	5	0	5	0	5	0
	*15,000	2	3	3	2	2	3
	22,500	0	5	0	5	0	5
Levorphanol	7,500	5	0	5	0	5	0
	*15,000	2	3	3	2	2	3
	22,500	0	5	0	5	0	5
MDE	7,500	5	0	5	0	5	0
	*15,000	2	3	3	2	2	3
	22,500	0	5	0	5	0	5

Tetrahydrozoline	75	5	0	5	0	5	0
	*150	2	3	3	2	2	3
	225	0	5	0	5	0	5
d-Methamphetamine	7,500	5	0	5	0	5	0
	*15,000	2	3	3	2	2	3
	22,500	0	5	0	5	0	5
L-Methamphetamine	7,500	5	0	5	0	5	0
	*15,000	2	3	3	2	2	3
	22,500	0	5	0	5	0	5
Methoxyphenamine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5
(+)3,4-Methylendioxymethamphetamine (MDMA)	15,000	5	0	5	0	5	0
	*30,000	3	2	3	2	2	3
	45,000	0	5	0	5	0	5
d-Norpropoxyphene	3,125	5	0	5	0	5	0
	*6,250	2	3	2	3	2	3
	9,375	0	5	0	5	0	5
Pentazocine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5
Phencyclidine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5
Promazine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	3	2
	9,375	0	5	0	5	0	5
Promethazine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5
Thioridazine	7,500	5	0	5	0	5	0

	*15,000	2	3	3	2	2	3
	22,500	0	5	0	5	0	5
Meperidine	3,125	5	0	5	0	5	0
	*6,250	2	3	3	2	2	3
	9,375	0	5	0	5	0	5

KET100

Analytes	**Conc. (ng/ml)**	**KET14030019p-T**		**KET14030020p-T**		**KET14030021p-T**	
		Negative	Positive	Negative	Positive	Negative	**Positive**
Ketamine	50	5	0	5	0	5	0
	*100	2	3	3	2	2	3
	150	0	5	0	5	0	5
Benzphetamine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
(+) Chlorpheniramine	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Clonidine	5,000	5	0	5	0	5	0
	*10,000	2	3	2	3	2	3
	15,000	0	5	0	5	0	5
Dextromethorphan	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Disopyramide	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
EDDP	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	3	2

	7,500	0	5	0	5	0	5
Mephentermine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
(1R, 2S) -(-)-Ephedrine	5,000	5	0	5	0	5	0
	*10,000	2	3	3	2	2	3
	15,000	0	5	0	5	0	5
4-Hydroxyphencyclidine	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
Levorphanol	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
MDE	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
Tetrahydrozoline	25	5	0	5	0	5	0
	*50	2	3	3	2	2	3
	75	0	5	0	5	0	5
d-Methamphetamine	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
L-Methamphetamine	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
Methoxyphenamine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5

(+)3,4-Methylendioxymethamphetamine (MDMA)	5,000	5	0	5	0	5	0
	*10,000	2	3	3	2	2	3
	15,000	0	5	0	5	0	5
d-Norpropoxyphene	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Pentazocine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Phencyclidine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Promazine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Promethazine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Thioridazine	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5
Meperidine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5

MDMA 1000:

Analytes	**Conc. (ng/ml)**	**MDMA14030007p-T**		**MDMA14030008p-T**		**MDMA14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
(±)	500	5	0	5	0	5	0

3,4-Methylenedioxymethamphetamine HCl (MDMA)	*1,000	2	3	3	2	2	3
	1,500	0	5	0	5	0	5
(±) 3,4-Methylenedioxyamphetamine HCl (MDA)	3,000	5	0	5	0	5	0
	*6,000	2	3	3	2	2	3
	9,000	0	5	0	5	0	5
3,4-Methylenedioxyethyl-amphetamine (MDE)	300	5	0	5	0	5	0
	*600	1	4	3	2	2	3
	900	0	5	0	5	0	5

MDMA 500:

Analytes	Conc. (ng/ml)	MDMA14030001p-T		MDMA14030002p-T		MDMA14030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
(±) 3,4-Methylenedioxymethamphetamine HCl (MDMA)	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
(±) 3,4-Methylenedioxyamphetamine HCl (MDA)	1,500	5	0	5	0	5	0
	*3,000	2	3	3	2	2	3
	4,500	0	5	0	5	0	5
3,4-Methylenedioxyethyl-amphetamine (MDE)	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5

MDMA 300:

Analytes	Conc. (ng/ml)	MDMA14030013p-T		MDMA14030014p-T		MDMA14030015p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
(±) 3,4-Methylenedioxymethamphetamine HCl (MDMA)	150	5	0	5	0	5	0
	*300	3	2	4	1	2	3
	450	0	5	0	5	0	5
(±) 3,4-Methylenedioxyamphetamine HCl (MDA)	900	5	0	5	0	5	0
	*1,800	1	4	3	2	2	3
	2,700	0	5	0	5	0	5

3,4-Methylenedioxyethyl-amphetamine (MDE)	90	5	0	5	0	5	0
	*180	2	3	3	2	2	3
	270	0	5	0	5	0	5

MET 1000:

Analytes	**Conc. (ng/ml)**	**MET14030001p-T**		**MET14030002p-T**		**MET14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
ρ-Hydroxymethamphetamine	12,500	5	0	5	0	5	0
	*25,000	2	3	2	3	2	3
	37,500	0	5	0	5	0	5
D-Methamphetamine	500	5	0	5	0	5	0
	*1,000	2	3	3	2	2	3
	1,500	0	5	0	5	0	5
L-Methamphetamine	10,000	5	0	5	0	5	0
	*20,000	2	3	2	3	2	3
	30,000	0	5	0	5	0	5
(±)-3,4-Methylenedioxy-methamphetamine	6,250	5	0	5	0	5	0
	12,500	2	3	2	3	3	2
	25,000	0	5	0	5	0	5
Mephentermine	25,000	5	0	5	0	5	0
	*50,000	3	2	2	3	3	2
	75,000	0	5	0	5	0	5

MET 500:

Analytes	**Conc. (ng/ml)**	**MET14030007p-T**		**MET14030008p-T**		**MET14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
ρ-Hydroxymethamphetamine	6,250	5	0	5	0	5	0
	*12,500	2	3	2	3	2	3
	18,750	0	5	0	5	0	5
D-Methamphetamine	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
L-Methamphetamine	5,000	5	0	5	0	5	0
	*10,000	2	3	2	3	2	3
	15,000	0	5	0	5	0	5
(±)-3,4-Methylenedioxy-methamph	3,125	5	0	5	0	5	0
	6,250	2	3	2	3	3	2

etamine	12,500	0	5	0	5	0	5
Mephentermine	6,250	5	0	5	0	5	0
	*25,000	3	2	2	3	3	2
	18,750	0	5	0	5	0	5

MET 300:

Analytes	**Conc. (ng/ml)**	**MET14030013p-T**		**MET14030014p-T**		**MET14030015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
ρ-Hydroxymethamphetamine	3,750	5	0	5	0	5	0
	*7,500	2	3	2	3	2	3
	11,250	0	5	0	5	0	5
D-Methamphetamine	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
L-Methamphetamine	3,000	5	0	5	0	5	0
	*6,000	2	3	2	3	2	3
	9,000	0	5	0	5	0	5
(±)-3,4-Methylenedioxy-methamphetamine	1,875	5	0	5	0	5	0
	3,750	2	3	2	3	3	2
	7,500	0	5	0	5	0	5
Mephentermine	7,500	5	0	5	0	5	0
	*15,000	3	2	2	3	3	2
	22,500	0	5	0	5	0	5

MOP 300:

Analytes	**Conc. (ng/ml)**	**MOP14030001p-T**		**MOP14030002p-T**		**MOP14030003p-T**	
		Negative	**Positive**	**Negative**	**Positive**	**Negative**	**Positive**
Codeine	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Levorphanol	750	5	0	5	0	5	0
	*1,500	2	3	2	3	2	3
	2,250	0	5	0	5	0	5
Morphine-3-β-D-Glucuronide	400	5	0	5	0	5	0
	*800	2	3	3	2	2	3
	1200	0	5	0	5	0	5
Ethylmorphine	3,000	5	0	5	0	5	0

	*6,000	2	3	3	2	2	3
	9,000	0	5	0	5	0	5
Hydrocodone	25,000	5	0	5	0	5	0
	*50,000	4	1	4	1	2	3
	75,000	0	5	0	5	0	5
Hydromorphone	1,500	5	0	5	0	5	0
	*3,000	2	3	3	2	2	3
	4,500	0	5	0	5	0	5
6-Monoacethylmorphine	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Norcodeine	3,000	5	0	5	0	5	0
	*6,000	2	3	3	2	2	3
	9,000	0	5	0	5	0	5
Normorphone	25,000	5	0	5	0	5	0
	*50,000	2	3	1	4	1	4
	75,000	0	5	0	5	0	5
Oxycodone	15,000	5	0	5	0	5	0
	*30,000	2	3	3	2	2	3
	45,000	0	5	0	5	0	5
Oxymorphone	25,000	5	0	5	0	5	0
	*50,000	4	1	4	1	4	1
	75,000	0	5	0	5	0	5
Procaine	7,500	5	0	5	0	5	0
	*15,000	2	3	3	2	2	3
	22,500	0	5	0	5	0	5
Thebaine	3,000	5	0	5	0	5	0
	*6,000	2	3	3	2	2	3
	9,000	0	5	0	5	0	5
Morphine	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5

MOP 100:

Analytes	**Conc. (ng/ml)**	**MOP14030007p-T**		**MOP14030008p-T**		**MOP14030009p-T**	
		Negative	**Positive**	**Negative**	**Positive**	**Negative**	**Positive**

Codeine	40	5	0	5	0	5	0
	*80	3	2	3	2	2	3
	120	0	5	0	5	0	5
Levorphanol	250	5	0	5	0	5	0
	*500	2	3	2	3	2	3
	750	0	5	0	5	0	5
Morphine-3-β-D-Glucuronide	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Ethylmorphine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Hydrocodone	10,000	5	0	5	0	5	0
	*20,000	4	1	4	1	2	3
	30,000	0	5	0	5	0	5
Hydromorphone	500	5	0	5	0	5	0
	*1,000	2	3	3	2	2	3
	1,500	0	5	0	5	0	5
6-Monoacethylmorphine	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Norcodeine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Normorphone	10,000	5	0	5	0	5	0
	*20,000	2	3	1+	4	1	4
	30,000	0	5	0	5	0	5
Oxycodone	5,000	5	0	5	0	5	0
	*10,000	2	3	3	2	2	3
	15,000	0	5	0	5	0	5
Oxymorphone	10,000	5	0	5	0	5	0
	*20,000	4	1	4	1	4	1
	30,000	0	5	0	5	0	5
Procaine	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3

	7,500	0	5	0	5	0	5
Thebaine	1,000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	3,000	0	5	0	5	0	5
Morphine	50	5	0	5	0	5	0
	*100	2	3	3	2	2	3
	150	0	5	0	5	0	5

MOP 200:

Analytes	Conc. (ng/ml)	MOP14030013p-T		MOP14030014p-T		MOP14030015p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Codeine	80	5	0	5	0	5	0
	*160	3	2	3	2	2	3
	320	0	5	0	5	0	5
Levorphanol	500	5	0	5	0	5	0
	*1000	2	3	2	3	2	3
	2000	0	5	0	5	0	5
Morphine-3-β-D-Glucuronide	300	5	0	5	0	5	0
	*600	2	3	3	2	2	3
	1200	0	5	0	5	0	5
Ethylmorphine	2000	5	0	5	0	5	0
	*4000	2	3	3	2	2	3
	8000	0	5	0	5	0	5
Hydrocodone	20,000	5	0	5	0	5	0
	*40,000	4	1	4	1	2	3
	80,000	0	5	0	5	0	5
Hydromorphone	1000	5	0	5	0	5	0
	*2,000	2	3	3	2	2	3
	4000	0	5	0	5	0	5
6-Monoacethylmorphine	200	5	0	5	0	5	0
	*400	2	3	3	2	2	3
	800	0	5	0	5	0	5
Norcodeine	2,000	5	0	5	0	5	0
	*4,000	2	3	3	2	2	3
	8,000	0	5	0	5	0	5

Normorphone	20,000	5	0	5	0	5	0
	*40,000	2	3	1+	4	1	4
	80,000	0	5	0	5	0	5
Oxycodone	10,000	5	0	5	0	5	0
	*20,000	2	3	3	2	2	3
	40,000	0	5	0	5	0	5
Oxymorphone	20,000	5	0	5	0	5	0
	*40,000	4	1	4	1	4	1
	80,000	0	5	0	5	0	5
Procaine	5000	5	0	5	0	5	0
	*10,000	2	3	3	2	2	3
	20000	0	5	0	5	0	5
Thebaine	2,000	5	0	5	0	5	0
	*4,000	2	3	3	2	2	3
	8,000	0	5	0	5	0	5
Morphine	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5

MQL 300

Analytes	**Conc. (ng/ml)**	**MQL14030001p-T**		**MQL14030002p-T**		**MQL14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Methaqualone	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	600	0	5	0	5	0	5

MTD 300:

Analytes	**Conc. (ng/ml)**	**MTD14030001p-T**		**MTD14030002p-T**		**MTD14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Methadone	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Doxylamine	50,000	5	0	5	0	5	0
	*100,000	2	3	2	3	2	3
	150,000	0	5	0	5	0	5

MTD 200:

Analytes	Conc. (ng/ml)	MTD14030007p-T		MTD14030008p-T		MTD14030009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Methadone	100	5	0	5	0	5	0
	*200	3	2	3	2	3	2
	300	0	5	0	5	0	5
Doxylamine	32,500	5	0	5	0	5	0
	*65,000	2	3	2	3	2	3
	97,500	0	5	0	5	0	5

OPI 2000:

Analytes	Conc. (ng/ml)	OPI14030001p-T		OPI14030002p-T		OPI14030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Codeine	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Ethylmorphine	1,500	5	0	5	0	5	0
	*3,000	2	3	2	3	2	3
	4,500	0	5	0	5	0	5
Hydrocodone	25,000	5	0	5	0	5	0
	*50,000	3	2	3	2	2	3
	75,000	0	5	0	5	0	5
Hydromorphone	7,500	5	0	5	0	5	0
	*15,000	2	3	2	3	2	3
	22,500	0	5	0	5	0	5
Levorphanol	12,500	5	0	5	0	5	0
	*25,000	3	2	3	2	2	3
	37,500	0	5	0	5	0	5
6-Monoacetylmorphine	1,500	5	0	5	0	5	0
	*3,000	2	3	2	3	2	3
	4,500	0	5	0	5	0	5
Morphine 3-β-D-glucuronide	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Morphine	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5

Norcodeine	12,500	5	0	5	0	5	0
	*25,000	2	3	2	3	2	3
	37,500	0	5	0	5	0	5
Normorphone	25,000	5	0	5	0	5	0
	*50,000	3	2	3	2	2	3
	75,000	0	5	0	5	0	5
Oxycodone	12,500	5	0	5	0	5	0
	*25,000	2	3	2	3	2	3
	37,500	0	5	0	5	0	5
Oxymorphone	12,500	5	0	5	0	5	0
	*25,000	3	2	3	2	2	3
	37,500	0	5	0	5	0	5
Procaine	25,000	5	0	5	0	5	0
	*50,000	2	3	2	3	2	3
	75,000	0	5	0	5	0	5
Thebaine	12,500	5	0	5	0	5	0
	*25,000	3	2	3	2	2	3
	37,500	0	5	0	5	0	5

PCP 25:

Analytes	**Conc. (ng/ml)**	**PCP14030001p-T**		**PCP14030002p-T**		**PCP14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Phencyclidine	12.5	5	0	5	0	5	0
	*25	3	2	3	2	3	2
	37.5	0	5	0	5	0	5
4-Hydroxyphencyclidine	6,250	5	0	5	0	5	0
	*12,500	2	3	2	3	2	3
	18,750	0	5	0	5	0	5

PPX 300:

Analytes	**Conc. (ng/ml)**	**PPX14030001p-T**		**PPX14030002p-T**		**PPX14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
D-Propoxyphene	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
D-Norpropoxyphene	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3

	450	0	5	0	5	0	5

THC 150:

Analytes	Conc. (ng/ml)	THC14030001p-T		THC14030002p-T		THC14030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	50,000	5	0	5	0	5	0
	*100,000	3	2	3	2	2	3
	>100,000	0	5	0	5	0	5
11-nor-Δ8pHC-9 COOH	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
11-nor-Δ9 pHC-9 COOH	75	5	0	5	0	5	0
	*150	2	3	2	3	2	3
	225	0	5	0	5	0	5
Δ^8 pHC	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
Δ^9 pHC	25,000	5	0	5	0	5	0
	*50,000	4	1	3	2	2	3
	75,000	0	5	0	5	0	5

THC 50:

Analytes	Conc. (ng/ml)	THC14030007p-T		THC14030008p-T		THC14030009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	17,500	5	0	5	0	5	0
	*35,000	3	2	3	2	2	3
	52,500	0	5	0	5	0	5
11-nor-Δ8pHC-9 COOH	15	5	0	5	0	5	0
	*30	3	2	3	2	2	3
	45	0	5	0	5	0	5
11-nor-Δ9 pHC-9 COOH	25	5	0	5	0	5	0
	*50	2	3	2	3	2	3

	75	0	5	0	5	0	5
Δ8 pHC	8,500	5	0	5	0	5	0
	*17,000	2	3	3	2	2	3
	25,500	0	5	0	5	0	5
Δ9 pHC	8,500	5	0	5	0	5	0
	*17,000	4	1	4	1	2	3
	25,500	0	5	0	5	0	5

THC 25:

Analytes	**Conc. (ng/ml)**	**THC14030013p-T**		**THC14030014p-T**		**THC14030015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	8,750	5	0	5	0	5	0
	*17,500	3	2	3	2	2	3
	26,250	0	5	0	5	0	5
11-nor-Δ^8 p HC-9 COOH	7.5	5	0	5	0	5	0
	*15	3	2	3	2	2	3
	45	0	5	0	5	0	5
11-nor-Δ^9 pHC-9 COOH	12.5	5	0	5	0	5	0
	*25	2	3	2	3	2	3
	37.5	0	5	0	5	0	5
Δ^8 pHC	4,250	5	0	5	0	5	0
	*8,500	2	3	3	2	2	3
	12,750	0	5	0	5	0	5
Δ^9 pHC	4,250	5	0	5	0	5	0

	*8,500	4	1	4	1	2	3
	12,750	0	5	0	5	0	5

THC300

Analytes	Conc. (ng/ml)	THC14030019p-T		THC14030020p-T		THC14030021p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	100,000	5	0	5	0	5	0
	*200,000	3	2	3	2	2	3
	>300,000	0	5	0	5	0	5
11-nor-Δ8-THC-9 COOH	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
11-nor-Δ9 -THC-9 COOH	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
Δ8 -THC	50,000	5	0	5	0	5	0
	*100,000	2	3	3	2	2	3
	150,000	0	5	0	5	0	5
Δ9 -THC	50,000	5	0	5	0	5	0
	*100,000	4	1	3	2	2	3
	150,000	0	5	0	5	0	5

THC 20

Analytes	Conc. (ng/ml)	THC14030037p-T		THC14030038p-T		THC14030039p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	10,000	5	0	5	0	5	0
	*20,000	3	2	3	2	3	2
	>30,000	0	5	0	5	0	5
11-nor-Δ8-THC-9 COOH	10	5	0	5	0	5	0
	*20	3	2	3	2	2	3
	30	0	5	0	5	0	5
11-nor-Δ9 -THC-9 COOH	15	5	0	5	0	5	0
	*30	2	3	3	2	2	3
	45	0	5	0	5	0	5
Δ^8 -THC	5,000	5	0	5	0	5	0
	*10,000	2	3	3	2	2	3

	15,000	0	5	0	5	0	5
Δ^9 -THC	5,000	5	0	5	0	5	0
	*10,000	3	2	3	2	2	3
	15,000	0	5	0	5	0	5

THC 200

Analytes	**Conc. (ng/ml)**	**THC140300025p-T**		**THC140300026p-T**		**THC140300027p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	70,000	5	0	5	0	5	0
	*140,000	3	2	3	2	2	3
	210,000	0	5	0	5	0	5
11-nor-Δ^8-THC-9 COOH	60	5	0	5	0	5	0
	*120	3	2	3	2	2	3
	180	0	5	0	5	0	5
11-nor-Δ^9 -THC-9 COOH	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
Δ^8 -THC	34,000	5	0	5	0	5	0
	*68,000	2	3	3	2	2	3
	102,000	0	5	0	5	0	5
Δ^9 -THC	34,000	5	0	5	0	5	0
	*68,000	4	1	3	2	2	3
	102,000	0	5	0	5	0	5

THC 30

Analytes	**Conc. (ng/ml)**	**THC140300031p-T**		**THC140300032p-T**		**THC140300033p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	10,000	5	0	5	0	5	0
	*20,000	3	2	3	2	2	3
	>30,000	0	5	0	5	0	5
11-nor-Δ^8-THC-9 COOH	10	5	0	5	0	5	0
	*20	3	2	3	2	2	3
	30	0	5	0	5	0	5
11-nor-Δ^9 -THC-9 COOH	15	5	0	5	0	5	0
	*30	2	3	2	3	2	3
	45	0	5	0	5	0	5
Δ^8 -THC	5,000	5	0	5	0	5	0

	*10,000	2	3	3	2	2	3
	15,000	0	5	0	5	0	5
Δ^9 -THC	5,000	5	0	5	0	5	0
	*10,000	4	1	3	2	2	3
	15,000	0	5	0	5	0	5

TCA 1000:

Analytes	Conc. (ng/ml)	TCA14030001p-T		TCA14030002p-T		TCA14030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Nortriptyline	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
Nordoxepine	250	5	0	5	0	5	0
	*500	2	3	2	3	2	3
	750	0	5	0	5	0	5
Trimipramine	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
Amitriptyline	750	5	0	5	0	5	0
	*1,500	2	3	2	3	2	3
	3,000	0	5	0	5	0	5
Promazine	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
Desipramine	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	400	0	5	0	5	0	5
Cyclobenzaprine	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Imipramine	200	5	0	5	0	5	0
	*400	2	3	2	3	2	3
	800	0	5	0	5	0	5
Clomipramine	25,000	5	0	5	0	5	0
	*50,000	3	2	3	2	2	3
	100,000	0	5	0	5	0	5

Doxepine	1,000	5	0	5	0	5	0
	*2,000	2	3	2	3	2	3
	3,000	0	5	0	5	0	5
Maprotiline	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Promethazine	25,000	5	0	5	0	5	0
	*50,000	2	3	2	3	2	3
	100,000	0	5	0	5	0	5
Perphenazine	25,000	5	0	5	0	5	0
	*50,000	3	2	3	2	2	3
	100,000	0	5	0	5	0	5
Dithiaden	5,000	5	0	5	0	5	0
	*10,000	3	2	3	2	3	2
	15,000	0	5	0	5	0	5

TCA 500:

Analytes	**Conc. (ng/ml)**	**TCA14030007p-T**		**TCA14030008p-T**		**TCA14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Nortriptyline	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Nordoxepine	125	5	0	5	0	5	0
	*250	2	3	2	3	2	3
	375	0	5	0	5	0	5
Trimipramine	750	5	0	5	0	5	0
	*1,500	3	2	3	2	2	3
	2250	0	5	0	5	0	5
Amitriptyline	375	5	0	5	0	5	0
	*750	2	3	2	3	2	3
	1125	0	5	0	5	0	5
Promazine	750	5	0	5	0	5	0
	*1,500	3	2	3	2	2	3
	2,250	0	5	0	5	0	5
Desipramine	50	5	0	5	0	5	0
	*100	2	3	2	3	2	3

	150	0	5	0	5	0	5
Cyclobenzaprine	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
Imipramine	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	300	0	5	0	5	0	5
Clomipramine	12,500	5	0	5	0	5	0
	*25,000	3	2	3	2	2	3
	50,000	0	5	0	5	0	5
Doxepine	500	5	0	5	0	5	0
	*1,000	2	3	2	3	2	3
	1,500	0	5	0	5	0	5
Maprotiline	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
Promethazine	12,500	5	0	5	0	5	0
	*25,000	2	3	2	3	2	3
	50,000	0	5	0	5	0	5
Perphenazine	12,500	5	0	5	0	5	0
	*25,000	3	2	3	2	2	3
	50,000	0	5	0	5	0	5
Dithiaden	2,500	5	0	5	0	5	0
	*5,000	3	2	3	2	3	2
	7,500	0	5	0	5	0	5

TML 100:

Analytes	**Conc. (ng/ml)**	**TML14030001p-T**		**TML14030002p-T**		**TML14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
n-Desmethyl-cis -tramadol	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Cis-tramadol	50	5	0	5	0	5	0
	*100	2	3	3	2	2	3
	150	0	5	0	5	0	5
Procyclidine	50,000	5	0	5	0	5	0

	*100,000	2	3	3	2	2	3
	150,000	0	5	0	5	0	5
o-Desmethyl-cis-tramado	5,000	5	0	5	0	5	0
	*10,000	2	3	3	2	2	3
	15,000	0	5	0	5	0	5
Phencyclidine	50,000	5	0	5	0	5	0
	*100,000	2	3	3	2	2	3
	150,000	0	5	0	5	0	5
d,l-O-Desmethyl venlafaxine	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5

TML 200:

Analytes	**Conc. (ng/ml)**	**TML14030007p-T**		**TML14030008p-T**		**TML14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
n-Desmethyl-cis-tramadol	200	5	0	5	0	5	0
	*400	2	3	3	2	2	3
	600	0	5	0	5	0	5
Cis-tramadol	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
Procyclidine	100,000	5	0	5	0	5	0
	*200,000	2	3	3	2	2	3
	300,000	0	5	0	5	0	5
o-Desmethyl-cis-tramado	10,000	5	0	5	0	5	0
	*20,000	2	3	3	2	2	3
	30,000	0	5	0	5	0	5
Phencyclidine	100,000	5	0	5	0	5	0
	*200,000	2	3	3	2	2	3
	300,000	0	5	0	5	0	5
d,l-O-Desmethyl venlafaxine	50,000	5	0	5	0	5	0
	*100,000	2	3	3	2	2	3
	150,000	0	5	0	5	0	5

TML300:

Analytes	**Conc.**	**TML14030013p-T**	**TML14030014p-T**	**TML14030015p-T**

	(ng/ml)	Negative	Positive	Negative	Positive	Negative	Positive
n-Desmethyl-cis-tramadol	300	5	0	5	0	5	0
	*600	2	3	3	2	2	3
	900	0	5	0	5	0	5
Cis-tramadol	150	5	0	5	0	5	0
	*300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Procyclidine	150,000	5	0	5	0	5	0
	*300,000	2	3	3	2	2	3
	450,000	0	5	0	5	0	5
o-Desmethyl-cis-tramado	15,000	5	0	5	0	5	0
	*30,000	2	3	3	2	2	3
	45,000	0	5	0	5	0	5
Phencyclidine	150,000	5	0	5	0	5	0
	*300,000	2	3	3	2	2	3
	450,000	0	5	0	5	0	5
d,l-O-Desmethyl venlafaxine	75,000	5	0	5	0	5	0
	*150,000	2	3	3	2	2	3
	225,000	0	5	0	5	0	5

OXY 100:

Analytes	**Conc. (ng/ml)**	**OXY14030001p-T**		**OXY14030002p-T**		**OXY14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Oxycodone	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Oxymorphone	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	600	0	5	0	5	0	5
Levorphanol	25,000	5	0	5	0	5	0
	*50,000	3	2	3	2	2	3
	100,000	0	5	0	5	0	5
Hydrocodone	12,500	5	0	5	0	5	0
	*25,000	2	3	2	3	2	3
	50,000	0	5	0	5	0	5

Hydromorphone	25,000	5	0	5	0	5	0
	*50,000	3	2	3	2	2	3
	100,000	0	5	0	5	0	5
Naloxone	12,500	5	0	5	0	5	0
	*25,000	2	3	2	3	2	3
	50,000	0	5	0	5	0	5
Naltrexone	12,500	5	0	5	0	5	0
	*25,000	3	2	3	2	2	3
	50,000	0	5	0	5	0	5

OXY 300:

Analytes	**Conc. (ng/ml)**	**OXY14030007p-T**		**OXY14030008p-T**		**OXY14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Oxycodone	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Oxymorphone	450	5	0	5	0	5	0
	*900	2	3	2	3	2	3
	1350	0	5	0	5	0	5
Levorphanol	7,500	5	0	5	0	5	0
	*15,000	3	2	3	2	2	3
	22,500	0	5	0	5	0	5
Hydrocodone	37,500	5	0	5	0	5	0
	*75,000	2	3	2	3	2	3
	112,500	0	5	0	5	0	5
Hydromorphone	75,000	5	0	5	0	5	0
	*150,000	3	2	3	2	2	3
	225,000	0	5	0	5	0	5
Naloxone	37,500	5	0	5	0	5	0
	*75,000	2	3	2	3	2	3
	112,500	0	5	0	5	0	5
Naltrexone	37,500	5	0	5	0	5	0
	*75,000	3	2	3	2	2	3
	112,500	0	5	0	5	0	5

COT 200:

Analytes	Conc. (ng/ml)	COT14030001p-T		COT14030002p-T		COT14030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
(-)-Cotinine	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	400	0	5	0	5	0	5
(-)-Nicotine	2,500	5	0	5	0	5	0
	*5,000	2	3	3	2	2	3
	7,500	0	5	0	5	0	5

COT 100:

Analytes	Conc. (ng/ml)	COT14030007p-T		COT14030008p-T		COT14030009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
(-)-Cotinine	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	200	0	5	0	5	0	5
(-)-Nicotine	1,250	5	0	5	0	5	0
	*2,500	2	3	3	2	2	3
	3,750	0	5	0	5	0	5

COT500:

Analytes	Conc. (ng/ml)	COT14030013p-T		COT14030014p-T		COT14030015p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
(-)-Cotinine	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
(-)-Nicotine	6250	5	0	5	0	5	0
	*12500	2	3	3	2	2	3
	18750	0	5	0	5	0	5

COT50:

Analytes	Conc. (ng/ml)	COT14030019p-T		COT14030020p-T		COT14030021p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
(-)-Cotinine	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5
(-)-Nicotine	625	5	0	5	0	5	0
	*1250	3	2	3	2	2	3

	1875	0	5	0	5	0	5

COT10:

Analytes	Conc. (ng/ml)	COT14030025p-T		COT14030026p-T		COT14030027p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
(-)-Cotinine	5	5	0	5	0	5	0
	*10	3	2	3	2	2	3
	15	0	5	0	5	0	5
(-)-Nicotine	125	5	0	5	0	5	0
	*250	2	3	3	2	2	3
	375	0	5	0	5	0	5

EDDP 300:

Analytes	**Conc. (ng/ml)**	**EDDP14030001p-T**		**EDDP14030002p-T**		**EDDP14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
2-Ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP)	150	5	0	5	0	5	0
	*300	3	2	2	3	2	3
	600	0	5	0	5	0	5

EDDP 100:

Analytes	**Conc. (ng/ml)**	**EDDP14030007p-T**		**EDDP14030008p-T**		**EDDP14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
2-Ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP)	50	5	0	5	0	5	0
	*100	2	3	2	3	2	3
	200	0	5	0	5	0	5

FYL 20

Analytes	**Conc. (ng/ml)**	**FYL18040001p-T**		**FYL18040002p-T**		**FYL18040003p-T**	
		Negative	**Positive**	**Negative**	**Positive**	**Negative**	**Positive**
Fentanyl	10	5	0	5	0	5	0
	*20	3	2	3	2	2	3
	30	0	5	0	5	0	5
Cyclopro Fentanyl	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

Norfentany	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
	>100,000	0	5	0	5	0	5
(±)cis-3-Methyl fentanyl	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Butyl fentanyl	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Valeryl Fentanyl	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Methoxyacetyl -Fentanyl	20	5	0	5	0	5	0
	40	3	2	3	2	2	3
	60	0	5	0	5	0	5
Acetyl Fentanyl	20	5	0	5	0	5	0
	40	3	2	3	2	2	3
	60	0	5	0	5	0	5
Ocfentanil	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
4-Fluoro-isobu tyryl Fentanyl	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
para-Fluorobut yryl fentanil (PBPF)	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
para-Fluorofen tanil	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5

FYL 10

Analytes	**Conc. (ng/ml)**	**FYL18040007p-T**		**FYL18040008p-T**		**FYL18040009p-T**	
		Negative	**Positive**	**Negative**	**Positive**	**Negative**	**Positive**
Fentanyl	5	5	0	5	0	5	0

	*10	3	2	3	2	2	3
	15	0	5	0	5	0	5
Cyclopro Fentanyl	125	5	0	5	0	5	0
	*250	3	2	3	2	2	3
	375	0	5	0	5	0	5
Norfentany	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
	>100,000	0	5	0	5	0	5
(±)cis-3-Methyl fentanyl	125	5	0	5	0	5	0
	*250	3	2	3	2	2	3
	375	0	5	0	5	0	5
Butyl fentanyl	75	5	0	5	0	5	0
	*150	3	2	3	2	2	3
	225	0	5	0	5	0	5
Valeryl Fentanyl	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Methoxyacetyl -Fentanyl	10	5	0	5	0	5	0
	20	3	2	3	2	2	3
	30	0	5	0	5	0	5
Acetyl Fentanyl	10	5	0	5	0	5	0
	20	3	2	3	2	2	3
	30	0	5	0	5	0	5
Ocfentanil	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
4-Fluoro-isobu tyryl Fentanyl	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
para-Fluorobut yryl fentanil (PBPF)	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
para-Fluorofen tanil	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5

FYL 100

Analytes	Conc. (ng/ml)	FYL18040013p-T		FYL18040014p-T		FYL18040015p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Fentanyl	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Cyclopro Fentanyl	1,250	5	0	5	0	5	0
	*2,500	3	2	3	2	2	3
	3,750	0	5	0	5	0	5
Norfentany	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
	>100,000	0	5	0	5	0	5
(±)cis-3-Methyl fentanyl	1,250	5	0	5	0	5	0
	*2,500	3	2	3	2	2	3
	3,750	0	5	0	5	0	5
Butyl fentanyl	750	5	0	5	0	5	0
	*1,500	3	2	3	2	2	3
	2,250	0	5	0	5	0	5
Valeryl Fentanyl	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
Methoxyacetyl -Fentanyl	100	5	0	5	0	5	0
	200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Acetyl Fentanyl	100	5	0	5	0	5	0
	200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Ocfentanil	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
4-Fluoro-isobu tyryl Fentanyl	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
para-Fluorobut	500	5	0	5	0	5	0

yryl fentanil (PBPF)	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
para-Fluorofentanil	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

FYL 200

Analytes	Conc. (ng/ml)	FYL18040019p-T		FYL18040020p-T		FYL18040021p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Fentanyl	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Cyclopro Fentanyl	2,500	5	0	5	0	5	0
	*5,000	3	2	3	2	2	3
	7,500	0	5	0	5	0	5
Norfentany	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
	>100,000	0	5	0	5	0	5
(±)cis-3-Methyl fentanyl	2,500	5	0	5	0	5	0
	*5,000	3	2	3	2	2	3
	7,500	0	5	0	5	0	5
Butyl fentanyl	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
Valeryl Fentanyl	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Methoxyacetyl -Fentanyl	200	5	0	5	0	5	0
	400	3	2	3	2	2	3
	600	0	5	0	5	0	5
Acetyl Fentanyl	200	5	0	5	0	5	0
	400	3	2	3	2	2	3
	600	0	5	0	5	0	5
Ocfentanil	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3

	3,000	0	5	0	5	0	5
4-Fluoro-isobutyryl Fentanyl	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
para-Fluorobutyryl fentanil (PBPF)	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
para-Fluorofentanil	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5

FYL 300

Analytes	**Conc. (ng/ml)**	**FYL18040025p-T**		**FYL18040026p-T**		**FYL18040027p-T**	
		Negative	**Positive**	**Negative**	**Positive**	**Negative**	**Positive**
Fentanyl	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Cyclopro Fentanyl	3,750	5	0	5	0	5	0
	*7,500	3	2	3	2	2	3
	11,250	0	5	0	5	0	5
Norfentany	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
	>100,000	0	5	0	5	0	5
(±)cis-3-Methyl fentanyl	3,750	5	0	5	0	5	0
	*7,500	3	2	3	2	2	3
	11,250	0	5	0	5	0	5
Butyl fentanyl	2,250	5	0	5	0	5	0
	*4,500	3	2	3	2	2	3
	6,750	0	5	0	5	0	5
Valeryl Fentanyl	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
Methoxyacetyl-Fentanyl	300	5	0	5	0	5	0
	600	3	2	3	2	2	3
	900	0	5	0	5	0	5

Acetyl Fentanyl	300	5	0	5	0	5	0
	600	3	2	3	2	2	3
	900	0	5	0	5	0	5
Ocfentanil	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
4-Fluoro-isobutyryl Fentanyl	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
para-Fluorobutyryl fentanil (PBPF)	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
para-Fluorofentanil	750	5	0	5	0	5	0
	*1,500	3	2	3	2	2	3
	2,250	0	5	0	5	0	5

K2 50

Analytes	**Conc. (ng/ml)**	**SMA14030001p-T**		**SMA14030002p-T**		**SMA14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
JWH-018 5-Pentanoic acid	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5
JWH-073 4-butanoic acid	25	5	0	5	0	5	0
	*50	3	2	2	3	2	3
	75	0	5	0	5	0	5
JWH-018 4-Hydroxypentyl	200	5	0	5	0	5	0
	*400	2	3	3	2	2	3
	800	0	5	0	5	0	5
JWH-018 5-Hydroxypentyl	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
JWH-073 4-Hydroxybutyl	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

K2 30

Analytes	Conc. (ng/ml)	SMA14030007p-T		SMA14030008p-T		SMA14030009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
JWH-018 5-Pentanoic acid	15	5	0	5	0	5	0
	*30	3	2	3	2	2	3
	60	0	5	0	5	0	5
JWH-073 4-butanoic acid	15	5	0	5	0	5	0
	*30	3	2	2	3	2	3
	60	0	5	0	5	0	5
JWH-018 4-Hydroxypentyl	125	5	0	5	0	5	0
	*250	2	3	3	2	2	3
	375	0	5	0	5	0	5
JWH-018 5-Hydroxypentyl	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
JWH-073 4-Hydroxybutyl	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5

K2 25

Analytes	Conc. (ng/ml)	SMA14030013p-T		SMA14030014p-T		SMA14030015p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
JWH-018 5-Pentanoic acid metabolite	12.5	5	0	5	0	5	0
	*25	3	2	3	2	2	3
	37.5	0	5	0	5	0	5
JWH-073 4-butanoic acid metabolite	12.5	5	0	5	0	5	0
	*25	2	3	2	3	2	3
	37.5	0	5	0	5	0	5
JWH-018 4-Hydroxypentyl metabolite	100	5	0	5	0	5	0
	*200	2	3	3	2	2	3
	300	0	5	0	5	0	5
JWH-018 5-Hydroxypentyl metabolite	125	5	0	5	0	5	0
	*250	2	3	3	2	2	3
	375	0	5	0	5	0	5

JWH-073 4-Hydroxybuty l metabolite	125	5	0	5	0	5	0
	*250	4	1	4	1	2	3
	375	0	5	0	5	0	5

6-MAM 10

Analytes	**Conc. (ng/ml)**	**MAM14030001p-T**		**MAM14030002p-T**		**MAM14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
6-mono-aceto-morphine	5	5	0	5	0	5	0
	*10	2	3	3	2	2	3
	15	0	5	0	5	0	5
Morphine	50,000	5	0	5	0	5	0
	*100,000	2	3	3	2	2	3
	150,000	0	5	0	5	0	5

CLO 400

Analytes	**Conc. (ng/ml)**	**CLO14030001p-T**		**CLO14030002p-T**		**CLO14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Clonazepam	200	5	0	5	0	5	0
	*400	3	2	3	2	2	3
	600	0	5	0	5	0	5
Alprazolam	100	5	0	5	0	5	0
	*200	2	3	2	3	2	3
	400	0	5	0	5	0	5
a-hydroxyalprazol am	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	4,000	0	5	0	5	0	5
Bromazepam	500	5	0	5	0	5	0
	*1,000	2	3	2	3	2	3
	1,500	0	5	0	5	0	5
Chlordiazepoxide	500	5	0	5	0	5	0

	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
Clobazam	125	5	0	5	0	5	0
	*250	2	3	2	3	2	3
	375	0	5	0	5	0	5
Clorazepate dipotassium	300	5	0	5	0	5	0
	*600	3	2	3	2	2	3
	900	0	5	0	5	0	5
Delorazepam	500	5	0	5	0	5	0
	*1,000	2	3	2	3	2	3
	1,500	0	5	0	5	0	5
Desalkylflurazepam	125	5	0	5	0	5	0
	*250	3	2	3	2	2	3
	375	0	5	0	5	0	5
Diazepam	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
Estazolam	625	5	0	5	0	5	0
	*1250	3	2	3	2	2	3
	1,900	0	5	0	5	0	5
Flunitrazepam	150	5	0	5	0	5	0
	*300	2	3	2	3	2	3
	450	0	5	0	5	0	5
(+-) Lorazepam	625	5	0	5	0	5	0
	*1250	3	2	3	2	2	3
	1,900	0	5	0	5	0	5
RS-Lorazepamglucuronide	125	5	0	5	0	5	0
	*250	3	2	3	2	2	3
	375	0	5	0	5	0	5
Midazolam	2,500	5	0	5	0	5	0
	*5,000	3	2	3	2	2	3
	7,500	0	5	0	5	0	5
Nitrazepam	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	400	0	5	0	5	0	5

Norchlordiazepoxide	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	400	0	5	0	5	0	5
Nordiazepam	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
Oxazepam	175	5	0	5	0	5	0
	*350	3	2	3	2	2	3
	525	0	5	0	5	0	5
Temazepam	75	5	0	5	0	5	0
	*150	3	2	3	2	2	3
	225	0	5	0	5	0	5
Triazolam	2,500	5	0	5	0	5	0
	*5,000	3	2	3	2	2	3
	7,500	0	5	0	5	0	5

CLO150

Analytes	**Conc. (ng/ml)**	**CLO14030007p-T**		**CLO14030008p-T**		**CLO14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Clonazepam	75	5	0	5	0	5	0
	*150	3	2	3	2	2	3
	225	0	5	0	5	0	5
Alprazolam	37	5	0	5	0	5	0
	*75	2	3	2	3	2	3
	110	0	5	0	5	0	5
a-hydroxyalprazolam	370	5	0	5	0	5	0
	*750	3	2	3	2	2	3
	1,100	0	5	0	5	0	5
Bromazepam	200	5	0	5	0	5	0
	*400	2	3	2	3	2	3
	600	0	5	0	5	0	5
Chlordiazepoxide	200	5	0	5	0	5	0
	*400	3	2	3	2	2	3
	600	0	5	0	5	0	5
Clobazam	50	5	0	5	0	5	0

	*100	2	3	2	3	2	3
	150	0	5	0	5	0	5
Clorazepate dipotassium	125	5	0	5	0	5	0
	*250	3	2	3	2	2	3
	375	0	5	0	5	0	5
Delorazepam	200	5	0	5	0	5	0
	*400	2	3	2	3	2	3
	600	0	5	0	5	0	5
Desalkylflurazepam	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Diazepam	60	5	0	5	0	5	0
	*120	2	3	2	3	2	3
	180	0	5	0	5	0	5
Estazolam	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Flunitrazepam	60	5	0	5	0	5	0
	*120	2	3	2	3	2	3
	180	0	5	0	5	0	5
(+-) Lorazepam	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
RS-Lorazepamglucuronide	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Midazolam	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
Nitrazepam	40	5	0	5	0	5	0
	*75	3	2	3	2	2	3
	110	0	5	0	5	0	5
Norchlordiazepoxide	40	5	0	5	0	5	0
	*75	3	2	3	2	2	3
	110	0	5	0	5	0	5

Nordiazepam	200	5	0	5	0	5	0
	*400	3	2	3	2	2	3
	600	0	5	0	5	0	5
Oxazepam	65	5	0	5	0	5	0
	*130	3	2	3	2	2	3
	195	0	5	0	5	0	5
Temazepam	30	5	0	5	0	5	0
	*60	3	2	3	2	2	3
	90	0	5	0	5	0	5
Triazolam	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5

ETG 300

Analytes	**Conc. (ng/ml)**	**ETG14030013p-T**		**ETG14030014p-T**		**ETG14030015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Ethyl- β -D-Glucuronide	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Propyl β-D-glucuronide	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
Morphine 3β-glucuronide	50,000	5	0	5	0	5	0
	*100,000	3	2	3	2	2	3
	150,000	0	5	0	5	0	5
Morphine 6β-glucuronide	50,000	5	0	5	0	5	0
	*100,000	3	2	3	2	2	3
	150,000	0	5	0	5	0	5
Glucuronic Acid	50,000	5	0	5	0	5	0
	*100,000	3	2	3	2	2	3
	150,000	0	5	0	5	0	5
Ethanol	50,000	5	0	5	0	5	0

	*>100,000	3	2	3	2	2	3
Methanol	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3

ETG 500

Analytes	**Conc. (ng/ml)**	**ETG14030001p-T**		**ETG14030002p-T**		**ETG14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Ethyl- β -D-Glucuronide	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
Propyl β-D-glucuronide	25,000	5	0	5	0	5	0
	*50,000	2	3	3	2	2	3
	75,000	0	5	0	5	0	5
Morphine 3β-glucuronide	50,000	5	0	5	0	5	0
	*100,000	3	2	3	2	2	3
	150,000	0	5	0	5	0	5
Morphine 6β-glucuronide	50,000	5	0	5	0	5	0
	*100,000	3	2	3	2	2	3
	150,000	0	5	0	5	0	5
Glucuronic Acid	50,000	5	0	5	0	5	0
	*100,000	3	2	3	2	2	3
	150,000	0	5	0	5	0	5
Ethanol	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
Methanol	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3

ETG1000

Analytes	**Conc. (ng/ml)**	**ETG14030007p-T**		**ETG14030008p-T**		**ETG14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive

Ethyl- β -D-Glucuronide	500	5	0	5	0	5	0
	*1000	2	3	3	2	2	3
	1500	0	5	0	5	0	5
Propyl β-D-glucuronide	50,000	5	0	5	0	5	0
	*100,000	2	3	3	2	2	3
	*>100,000	0	5	0	5	0	5
Morphine 3β-glucuronide	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
Morphine 6β-glucuronide	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
GlucuronicAcid	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
Ethanol	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3
Methanol	50,000	5	0	5	0	5	0
	*>100,000	3	2	3	2	2	3

MDA 500

Analytes	**Conc. (ng/ml)**	**MDA14030001p-T**		**MDA14030002p-T**		**MDA14030003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
(±) 3,4-Methylenedioxyamphetamine	250	5	0	5	0	5	0
	*500	2	3	3	2	2	3
	750	0	5	0	5	0	5
L-Amphetamine	12,500	5	0	5	0	5	0
	*25,000	2	3	3	2	2	3
	37,500	0	5	0	5	0	5
Phentermine	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1500	0	5	0	5	0	5
Maprotiline	25,000	5	0	5	0	5	0
	*50,000	3	2	3	2	2	3

	75,000	0	5	0	5	0	5
Methoxyphenamine	3,000	5	0	5	0	5	0
	*6,000	3	2	3	2	2	3
	9,000	0	5	0	5	0	5
D-Amphetamine	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
D,L-Amphetamine sulfate	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5

ACE 5000

Analytes	**Conc. (ng/ml)**	**ACE14060001p-T**		**ACE14060002p-T**		**ACE14060003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Acetaminophen	2500	5	0	5	0	5	0
	*5000	3	2	3	2	2	3
	7500	0	5	0	5	0	5

LSD10

Analytes	**Conc. (ng/ml)**	**LSD 14060013p-T**		**LSD 14060014p-T**		**LSD 14060015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Lysergic Acid Diethylamide	5	5	0	5	0	5	0
	*10	3	2	3	2	2	3
	15	0	5	0	5	0	5
Fentanyl	7.5	5	0	5	0	5	0
	*15	3	2	3	2	2	3
	22.5	0	5	0	5	0	5

LSD20

Analytes	**Conc. (ng/ml)**	**LSD 14060001p-T**		**LSD 14060002p-T**		**LSD 14060003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Lysergic Acid Diethylamide	10	5	0	5	0	5	0
	*20	3	2	3	2	2	3
	30	0	5	0	5	0	5
Fentanyl	15	5	0	5	0	5	0

	*30	3	2	3	2	2	3
	45	0	5	0	5	0	5

LSD50

Analytes	Conc. (ng/ml)	LSD 14060007p-T		LSD 14060008p-T		LSD 14060009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Lysergic Acid Diethylamide	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5
Fentanyl	37.5	5	0	5	0	5	0
	*75	3	2	3	2	2	3
	112.5	0	5	0	5	0	5

MDPV1000

Analytes	Conc. (ng/ml)	MDPV14080001p-T		MDPV14080002p-T		MDPV14080003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
3,4-methylene dioxypyrovalerone	500	5	0	5	0	5	0
	*1000	3	2	3	2	2	3
	1500	0	5	0	5	0	5

MDPV500

Analytes	Conc. (ng/ml)	MDPV14080007p-T		MDPV14080008p-T		MDPV14080009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
3,4-methylene dioxypyrovalerone	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

MDPV300

Analytes	Conc. (ng/ml)	MDPV15090001p-T		MDPV15090002p-T		MDPV15090003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
3,4-methylene dioxypyrovalerone	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5

MPD 300

Analytes	Conc. (ng/ml)	MPD14060001p-T		MPD14060002p-T		MPD14060003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Methylphenidate (Ritalin)	150	5	0	5	0	5	0
	*300	2	3	4	1	2	3
	450	0	5	0	5	0	5
Ritalinic Acid	500	5	0	5	0	5	0
	*1000	3	2	3	2	2	3
	1500	0	5	0	5	0	5

MPD 1000

Analytes	Conc. (ng/ml)	MPD14060007p-T		MPD14060008p-T		MPD14060009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Methylphenidate (Ritalin)	175	5	0	5	0	5	0
	*350	2	3	3	2	2	3
	525	0	5	0	5	0	5
Ritalinic Acid	500	5	0	5	0	5	0
	*1000	3	2	3	2	2	3
	1500	0	5	0	5	0	5

ZOL 50

Analytes	Conc. (ng/ml)	ZOL15010001p-T		ZOL15010002p-T		ZOL15010003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Zolpidem	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5

DIA300:

Analytes	Conc. (ng/ml)	DIA14030001p-T		DIA14030002p-T		DIA14030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Alprazolam	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
a-hydroxyalprazolam	750	5	0	5	0	5	0
	*1,500	3	2	3	2	2	3
	2,250	0	5	0	5	0	5

Bromazepam	450	5	0	5	0	5	0
	*900	3	2	3	2	2	3
	1,350	0	5	0	5	0	5
Chlordiazepoxide	450	5	0	5	0	5	0
	*900	3	2	3	2	2	3
	1,350	0	5	0	5	0	5
Clobazam	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Clonazepam	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Clorazepate dipotassium	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Delorazepam	450	5	0	5	0	5	0
	*900	3	2	3	2	2	3
	1,350	0	5	0	5	0	5
Desalkylflurazepam	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Diazepam	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	600	0	5	0	5	0	5
Estazolam	3,000	5	0	5	0	5	0
	*6,000	3	2	3	2	2	3
	9,000	0	5	0	5	0	5
Flunitrazepam	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
(+-) Lorazepam	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5
RS-Lorazepamglucuronide	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3

	300	0	5	0	5	0	5
Midazolam	3,000	5	0	5	0	5	0
	*6,000	3	2	3	2	2	3
	9,000	0	5	0	5	0	5
Nitrazepam	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Norchlordiazepoxide	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Nordiazepam	450	5	0	5	0	5	0
	*900	3	2	3	2	2	3
	1,350	0	5	0	5	0	5
Oxazepam	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Temazepam	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Triazolam	1,500	5	0	5	0	5	0
	*3,000	3	2	3	2	2	3
	4,500	0	5	0	5	0	5

DIA 200

Analytes	**Conc. (ng/ml)**	**DIA14030007p-T**		**DIA14030008p-T**		**DIA14030009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Alprazolam	35	5	0	5	0	5	0
	*70	3	2	3	2	2	3
	105	0	5	0	5	0	5
a-hydroxyalprazolam	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5
Bromazepam	300	5	0	5	0	5	0
	*600	3	2	3	2	2	3
	900	0	5	0	5	0	5

Chlordiazepoxide	300	5	0	5	0	5	0
	*600	3	2	3	2	2	3
	900	0	5	0	5	0	5
Clobazam	60	5	0	5	0	5	0
	*120	3	2	3	2	2	3
	180	0	5	0	5	0	5
Clonazepam	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Clorazepatedipotassium	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Delorazepam	300	5	0	5	0	5	0
	*600	3	2	3	2	2	3
	900	0	5	0	5	0	5
Desalkylflurazepam	60	5	0	5	0	5	0
	*120	3	2	3	2	2	3
	180	0	5	0	5	0	5
Diazepam	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Estazolam	2,000	5	0	5	0	5	0
	*4,000	3	2	3	2	2	3
	6,000	0	5	0	5	0	5
Flunitrazepam	60	5	0	5	0	5	0
	*120	3	2	3	2	2	3
	180	0	5	0	5	0	5
(+-) Lorazepam	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5
RS-Lorazepam glucuronide	60	5	0	5	0	5	0
	*120	3	2	3	2	2	3
	180	0	5	0	5	0	5
Midazolam	2,000	5	0	5	0	5	0
	*4,000	3	2	3	2	2	3

	6,000	0	5	0	5	0	5
Nitrazepam	60	5	0	5	0	5	0
	*120	3	2	3	2	2	3
	180	0	5	0	5	0	5
Norchlordiazepoxide	35	5	0	5	0	5	0
	*70	3	2	3	2	2	3
	105	0	5	0	5	0	5
Nordiazepam	300	5	0	5	0	5	0
	*600	3	2	3	2	2	3
	900	0	5	0	5	0	5
Oxazepam	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Temazepam	35	5	0	5	0	5	0
	*70	3	2	3	2	2	3
	105	0	5	0	5	0	5
Triazolam	1,000	5	0	5	0	5	0
	*2,000	3	2	3	2	2	3
	3,000	0	5	0	5	0	5

ZOP 50

Analytes	**Conc. (ng/ml)**	**ZOP16080001p-T**		**ZOP16080002p-T**		**ZOP16080003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Zopiclone-N-oxide	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5
Zopiclone	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5

MCAT 500

Analytes	**Conc. (ng/ml)**	**MCAT16080001p-T**		**MCAT16080002p-T**		**MCAT16080003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
S(-)-Methcathinone HCl	250	5	0	5	0	5	0
	*500	3	2	3	2	3	2

	750	0	5	0	5	0	5
R(+)-Methcathinone HCl	750	5	0	5	0	5	0
	*1500	3	2	3	2	2	3
	2250	0	5	0	5	0	5
Methoxyphenamine	*100000	3	2	3	2	3	2
3-Fluoromethcathinone HCl	750	5	0	5	0	5	0
	*1500	3	2	3	2	2	3
	2250	0	5	0	5	0	5

7-ACL 300

Analytes	**Conc. (ng/ml)**	**ACL14100001p-T**		**ACL14100002p-T**		**ACL14100003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
a-hydroxyalprazo	3000	5	0	5	0	5	0
	*6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
Bromazepam	3000	5	0	5	0	5	0
	*6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
Chlordiazepoxide	3000	5	0	5	0	5	0
	*6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
Clobazam	4500	5	0	5	0	5	0
	*9000	3	2	3	2	2	3
	13500	0	5	0	5	0	5
Clonazepam	1200	5	0	5	0	5	0
	*2400	3	2	3	2	2	3
	3600	0	5	0	5	0	5
Delorazepam	3000	5	0	5	0	5	0
	*6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
Desalkylflurazep	3000	5	0	5	0	5	0
	*6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
Flunitrazepam	1500	5	0	5	0	5	0

	*3000	3	2	3	2	2	3
	4500	0	5	0	5	0	5
RS-Lorazepam	1350	5	0	5	0	5	0
	*2700	3	2	3	2	2	3
	5400	0	5	0	5	0	5
Norchlordiazepoxide	2250	5	0	5	0	5	0
	*4500	3	2	3	2	2	3
	6750	0	5	0	5	0	5
Nordiazepam	7500	5	0	5	0	5	0
	*15000	3	2	3	2	2	3
	22500	0	5	0	5	0	5
Temazepam	4500	5	0	5	0	5	0
	*9000	3	2	3	2	2	3
	13500	0	5	0	5	0	5
7-Aminoclonazepa	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5

7-ACL 100

Analytes	**Conc. (ng/ml)**	**ACL14100007p-T**		**ACL14100008p-T**		**ACL14100009p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
a-hydroxyalprazolam	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5
Bromazepam	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5
Chlordiazepoxide	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5
Clobazam	1500	5	0	5	0	5	0
	*3000	3	2	3	2	2	3
	4500	0	5	0	5	0	5
Clonazepam	400	5	0	5	0	5	0
	*800	3	2	3	2	2	3

	1200	0	5	0	5	0	5
Delorazepam	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5
Desalkylflurazepam	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5
Flunitrazepam	500	5	0	5	0	5	0
	*1000	3	2	3	2	2	3
	1500	0	5	0	5	0	5
RS-Lorazepam	450	5	0	5	0	5	0
	*900	3	2	3	2	2	3
	1350	0	5	0	5	0	5
Norchlordiazepoxide	750	5	0	5	0	5	0
	*1500	3	2	3	2	2	3
	2250	0	5	0	5	0	5
Nordiazepam	2500	5	0	5	0	5	0
	*5000	3	2	3	2	2	3
	7500	0	5	0	5	0	5
Temazepam	1500	5	0	5	0	5	0
	*3000	3	2	3	2	2	3
	4500	0	5	0	5	0	5
7-Aminoclonazepam	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5

7-ACL 200

Analytes	**Conc. (ng/ml)**	**ACL14100013p-T**		**ACL14100014p-T**		**ACL14100015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
a-hydroxyalprazo	2000	5	0	5	0	5	0
	*4000	3	2	3	2	2	3
	6000	0	5	0	5	0	5
Bromazepam	2000	5	0	5	0	5	0
	*4000	3	2	3	2	2	3
	6000	0	5	0	5	0	5

Chlordiazepoxide	2000	5	0	5	0	5	0
	*4000	3	2	3	2	2	3
	6000	0	5	0	5	0	5
Clobazam	3000	5	0	5	0	5	0
	*6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
Clonazepam	800	5	0	5	0	5	0
	*1600	3	2	3	2	2	3
	2400	0	5	0	5	0	5
Delorazepam	2000	5	0	5	0	5	0
	*4000	3	2	3	2	2	3
	6000	0	5	0	5	0	5
Desalkylflurazep	2000	5	0	5	0	5	0
	*4000	3	2	3	2	2	3
	6000	0	5	0	5	0	5
Flunitrazepam	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5
RS-Lorazepam	900	5	0	5	0	5	0
	*1800	3	2	3	2	2	3
	2700	0	5	0	5	0	5
Norchlordiazepoxide	1500	5	0	5	0	5	0
	*3000	3	2	3	2	2	3
	4500	0	5	0	5	0	5
Nordiazepam	5000	5	0	5	0	5	0
	*10000	3	2	3	2	2	3
	15000	0	5	0	5	0	5
Temazepam	3000	5	0	5	0	5	0
	*6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
7-Aminoclonazepa	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5

CFYL 500

Analytes	Conc. (ng/ml)	CFYL15020001p-T		CFYL15020002p-T		CFYL15020003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Carfentanyl	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Fentanyl	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Sufentanil	25000	5	0	5	0	5	0
	*50000	3	2	3	2	2	3
	75000	0	5	0	5	0	5
(±)cis-3-Methylfentanyl	10000	5	0	5	0	5	0
	*20000	3	2	3	2	2	3
	30000	0	5	0	5	0	5
Ramifentanil	5000	5	0	5	0	5	0
	*10000	3	2	3	2	2	3
	15000	0	5	0	5	0	5
Butyl fentanyl	75	5	0	5	0	5	0
	*150	3	2	3	2	2	3
	225	0	5	0	5	0	5

CFYL 250

Analytes	Conc. (ng/ml)	CFYL15020007p-T		CFYL15020008p-T		CFYL15020009p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Carfentanyl	125	5	0	5	0	5	0
	*250	3	2	3	2	2	3
	375	0	5	0	5	0	5
Fentanyl	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5
Sufentanil	12500	5	0	5	0	5	0
	*25000	3	2	3	2	2	3
	37500	0	5	0	5	0	5
(±)cis-3-Methylfentanyl	5000	5	0	5	0	5	0
	*10000	3	2	3	2	2	3

	15000	0	5	0	5	0	5
Ramifentanil	2500	5	0	5	0	5	0
	*5000	3	2	3	2	2	3
	7500	0	5	0	5	0	5
Butyl fentanyl	37.5	5	0	5	0	5	0
	*75	3	2	3	2	2	3
	112.5	0	5	0	5	0	5

CAF 1000

Analytes	**Conc. (ng/ml)**	**CAF15080001p-T**		**CAF15080002p-T**		**CAF15080003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Caffeine	500	5	0	5	0	5	0
	*1000	3	2	3	2	2	3
	1500	0	5	0	5	0	5

CAT 150

Analytes	**Conc. (ng/ml)**	**CAT14080001p-T**		**CAT14080002p-T**		**CAT14080003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
(+)-Norpseudoephedrine HCl	75	5	0	5	0	5	0
	*150	3	2	3	2	2	3
	225	0	5	0	5	0	5
(+)3,4-Methylenedioxyamphetamine (MDA)	50	5	0	5	0	5	0
	100	3	2	3	2	2	3
	150	0	5	0	5	0	5
d/l-Amphetamine	50	5	0	5	0	5	0
	100	3	2	2	3	2	3
	150	0	5	0	5	0	5
p-Hydroxyamphetamine	50	5	0	5	0	5	0
	100	3	2	3	2	3	2
	150	0	5	0	5	0	5
Tryptamine	6250	5	0	5	0	5	0
	12500	3	2	2	3	2	3
	18750	0	5	0	5	0	5
Methoxyphenamine	6250	5	0	5	0	5	0
	12500	2	3	3	2	2	3

	18750	0	5	0	5	0	5

TRO 350

Analytes	Conc. (ng/ml)	TRO15050001p-T		TRO15050002p-T		TRO15050003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Tropicamide	175	5	0	5	0	5	0
	*350	3	2	3	2	2	3
	525	0	5	0	5	0	5

MEP 100

Analytes	Conc. (ng/ml)	MEP15070001p-T		MEP15070002p-T		MEP15070003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Mephedrone	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
R(+)-Methcathinone HCl	750	5	0	5	0	5	0
	1,500	2	3	3	2	2	3
	2,250	0	5	0	5	0	5
S(-)-Methcathinone HCl	250	5	0	5	0	5	0
	500	3	2	2	3	2	3
	750	0	5	0	5	0	5
3-Fluoromethcathinone HCl	750	5	0	5	0	5	0
	1,500	3	2	3	2	2	3
	2,250	0	5	0	5	0	5
4-Fluoromethcathinone HCl	150	5	0	5	0	5	0
	300	2	3	3	2	2	3
	450	0	5	0	5	0	5
Methoxyphenamine	50,000	5	0	5	0	5	0
	100,000	3	2	3	2	3	2
	150,000	0	5	0	5	0	5

ALP 100

Analytes	Conc. (ng/ml)	ALP16040001p-T		ALP16040002p-T		ALP16040003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Alprazolam	50	5	0	5	0	5	0

	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Benzodiazepines	150	5	0	5	0	5	0
	300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Flunitrazepam	100	5	0	5	0	5	0
	200	3	2	3	2	2	3
	300	0	5	0	5	0	5
a-hydroxyalprazolam	750	5	0	5	0	5	0
	1500	3	2	3	2	2	3
	2250	0	5	0	5	0	5
(±) Lorazepam	1500	5	0	5	0	5	0
	3000	3	2	3	2	2	3
	4500	0	5	0	5	0	5
Bromazepam	450	5	0	5	0	5	0
	900	3	2	3	2	2	3
	1350	0	5	0	5	0	5
RS-Lorazepamglucuronid	100	5	0	5	0	5	0
	200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Chlordiazepoxide	450	5	0	5	0	5	0
	900	3	2	3	2	2	3
	1350	0	5	0	5	0	5
Midazolam	3000	5	0	5	0	5	0
	6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5
Clobazam	100	5	0	5	0	5	0
	200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Nitrazepam	100	5	0	5	0	5	0
	200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Clonazepam	250	5	0	5	0	5	0
	500	3	2	3	2	2	3
	750	0	5	0	5	0	5

Norchlordiazepoxide	50	5	0	5	0	5	0
	100	3	2	2	3	2	3
	150	0	5	0	5	0	5
Clorazepatedipotassium	250	5	0	5	0	5	0
	500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Nordiazepam	450	5	0	5	0	5	0
	900	3	2	3	2	2	3
	1350	0	5	0	5	0	5
Delorazepam	450	5	0	5	0	5	0
	900	3	2	2	3	2	3
	1350	0	5	0	5	0	5
Oxazepam	150	5	0	5	0	5	0
	300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Desalkylflurazepam	100	5	0	5	0	5	0
	200	3	2	3	2	2	3
	300	0	5	0	5	0	5
Temazepam	50	5	0	5	0	5	0
	100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Diazepam	150	5	0	5	0	5	0
	300	3	2	3	2	2	3
	450	0	5	0	5	0	5
Triazolam	1500	5	0	5	0	5	0
	3000	3	2	2	3	2	3
	4500	0	5	0	5	0	5
Estazolam	3000	5	0	5	0	5	0
	6000	3	2	3	2	2	3
	9000	0	5	0	5	0	5

ABP10

Analytes	**Conc. (ng/ml)**	**ABP17010001p-T**		**ABP17010002p-T**		**ABP17010003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
AB-PINACA	5	5	0	5	0	5	0

	*10	3	2	3	2	2	3
	15	0	5	0	5	0	5
AB-PINACA 5-Pentanoic	5	5	0	5	0	5	0
	*10	1	4	2	3	2	3
	15	0	5	0	5	0	5
AB-PINACA 5-hydroxypentyl	5	5	0	5	0	5	0
	*10	2	3	2	3	4	1
	15	0	5	0	5	0	5
AB-FUBINACA	5	5	0	5	0	5	0
	*10	4	1	3	2	2	3
	15	0	5	0	5	0	5
AB-PINACA 4-hydroxypentyl	5,000	5	0	5	0	5	0
	*10,000	3	2	2	3	3	2
	15,000	0	5	0	5	0	5
UR-144 5-Pentanoic	2,500	5	0	5	0	5	0
	*5,000	2	3	2	3	2	3
	7,500	0	5	0	5	0	5
UR-144 5-hydroxypentyl	5,000	5	0	5	0	5	0
	*10,000	2	3	3	2	1	4
	15,000	0	5	0	5	0	5
UR-144 4-hydroxypentyl	5,000	5	0	5	0	5	0
	*10,000	4	1	2	3	4	1
	15,000	0	5	0	5	0	5
APINACA 5-hydroxypentyl	5,000	5	0	5	0	5	0
	*10,000	1	4	3	2	1	4
	15,000	0	5	0	5	0	5
ADB-PINACA N-(5-hydroxypentyl)	15	5	0	5	0	5	0
	*30	3	2	3	2	1	4
	45	0	5	0	5	0	5
ADB-PINACA Pentanoic Acid	5	5	0	5	0	5	0
	*10	4	1	3	2	3	2
	15	0	5	0	5	0	5
5-fluoro AB-PINACA N-(4-hydroxypentyl)	15	5	0	5	0	5	0
	*30	2	3	3	2	3	2
	45	0	5	0	5	0	5

5-fluoro AB-PINACA	12.5	5	0	5	0	5	0
	*25	1	4	3	2	1	4
	37.5	0	5	0	5	0	5

APVP1000

Analytes	**Conc. (ng/ml)**	**APVP15050001p-T**		**APVP15050002p-T**		**APVP15050003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
alpha-Pyrrolidinovalerophenone	500	5	0	5	0	5	0
	*1,000	3	2	3	2	2	3
	1,500	0	5	0	5	0	5

APVP500

Analytes	**Conc. (ng/ml)**	**APVP15050001p-T**		**APVP15050002p-T**		**APVP15050003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
alpha-Pyrrolidinovalerophenone	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

APVP300

Analytes	**Conc. (ng/ml)**	**APVP15050013p-T**		**APVP15050014p-T**		**APVP15050015p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
alpha-Pyrrolidinovalerophenone	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5

CNB500

Analytes	**Conc. (ng/ml)**	**CNB18060001p-T**		**CNB18060002p-T**		**CNB18060003p-T**	
		Negative	Positive	Negative	Positive	Negative	Positive
Cannabinol	250	5	0	5	0	5	0
	*500	3	2	3	2	3	2
	750	0	5	0	5	0	5
11-nor-Δ^9 -THC-9 COOH	150	5	0	5	0	5	0
	*300	3	2	3	2	3	2
	450	0	5	0	5	0	5
Δ^9 –THC	50,000	5	0	5	0	5	0

	*100,000	3	2	3	2	3	2
	150,000	0	5	0	5	0	5

MPRD100

Analytes	Conc. (ng/ml)	MPRD18060001p-T		MPRD18060002p-T		MPRD18060003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Normeperidine	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5
Meperidine	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5

PGB50,000

Analytes	Conc. (ug/ml)	PGB16010001p-T		PGB16010002p-T		PGB16010003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Pregabalin	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5

PGB500

Analytes	Conc. (ng/ml)	PGB19040001p-T		PGB19040002p-T		PGB19040003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Pregabalin	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

TZD200

Analytes	Conc. (ng/ml)	TZD18070001p-T		TZD18070002p-T		TZD18070003p-T	
		Negative	**Positive**	**Negative**	**Positive**	**Negative**	**Positive**
Trazodone	100	5	0	5	0	5	0
	*200	3	2	3	2	2	3
	300	0	5	0	5	0	5

UR-144 25

UR-144 Analytes	Conc. (ng/ml)	UR18060001p-T		UR18060002p-T		UR18060003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
UR-144 5-Pentanoic acid	12.5	5	0	5	0	5	0
	*25	3	2	3	2	2	3
	37.5	0	5	0	5	0	5
UR-144 5-Pentanoic acid	5000	5	0	5	0	5	0
	*10000	3	2	3	2	2	3
	15000	0	5	0	5	0	5
UR-144 5-hydroxypentyl	2500	5	0	5	0	5	0
	*5000	3	2	3	2	2	3
	7500	0	5	0	5	0	5
XLR-11 4-hydroxypentyl	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5
5-fluoro AB-Pinaca N-(4-hydroxypentyl)	5000	5	0	5	0	5	0
	*10000	3	2	3	2	2	3
	15000	0	5	0	5	0	5
ADB-PINAC N-(4-hydroxypentyl)	5000	5	0	5	0	5	0
	*>10000	3	2	3	2	2	3
AB-PINACA 4-hydroxypentyl	5000	5	0	5	0	5	0
	*>10000	3	2	3	2	2	3

ZAL100

Analytes	Conc. (ng/ml)	ZAL18050001p-T		ZAL18050002p-T		ZAL18050003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Zaleplon	50	5	0	5	0	5	0
	*100	3	2	3	2	3	2
	150	0	5	0	5	0	5

MES 100

Analytes	Conc. (ng/ml)	MES18060001p-T		MES18060002p-T		MES18060003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive

Mescaline	50	5	0	5	0	5	0
	*100	3	2	3	2	3	2
	150	0	5	0	5	0	5

MES 300

Analytes	Conc. (ng/ml)	MES18090001p-T		MES18090002p-T		MES18090003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Mescaline	150	5	0	5	0	5	0
	*300	3	2	3	2	3	2
	450	0	5	0	5	0	5

GAB 2000

Analytes	Conc. (ng/ml)	GAB18070001p-T		GAB18070002p-T		GAB18070003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Gabapentin	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5

TLD 50

Analytes	Conc. (ng/ml)	TLD18110001p-T		TLD18110002p-T		TLD18110003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Nortilidine	25	5	0	5	0	5	0
	*50	3	2	3	2	2	3
	75	0	5	0	5	0	5
Tilidine	50	5	0	5	0	5	0
	*100	3	2	3	2	2	3
	150	0	5	0	5	0	5

QTP 1000

Analytes	Conc. (ng/ml)	QTP19030001p-T		QTP19030002p-T		QTP19030003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Quetiapine	500	5	0	5	0	5	0
	*1000	3	2	3	2	2	3
	1500	0	5	0	5	0	5
NorQuetiazepine	5,000	5	0	5	0	5	0

	*10,000	3	2	3	2	2	3
	15,000	0	5	0	5	0	5

PAP 500

Analytes	Conc. (ng/ml)	PAP18110001p-T		PAP18110002p-T		PAP18110003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Papaverine	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
Diflunisal	500,000	5	0	5	0	5	0
	*1,000,000	3	2	3	2	2	3
Methortrexate	97,500	5	0	5	0	5	0
	*65,000	3	2	3	2	2	3
	32,500	0	5	0	5	0	5
Methedrone	250,000	5	0	5	0	5	0
	*500,000	3	2	3	2	2	3
	750,000	0	5	0	5	0	5
Pragablin	250,000	5	0	5	0	5	0
	*500,000	3	2	3	2	2	3
	750,000	0	5	0	5	0	5
Phenelzine	4,000	5	0	5	0	5	0
	*8,000	3	2	3	2	2	3
	12,000	0	5	0	5	0	5
Quinine	2,000	5	0	5	0	5	0
	*4,000	3	2	3	2	2	3
	6,000	0	5	0	5	0	5

KRA 300

Analytes	Conc. (ng/ml)	KRA19040001p-T		KRA19040002p-T		KRA19040003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Mitragynine	150	5	0	5	0	5	0
	*300	3	2	3	2	2	3
	450	0	5	0	5	0	5
7-hydroxyMitragyn ine	25,000	5	0	5	0	5	0
	*>50,000	3	2	3	2	2	3

CAR 2000

Analytes	Conc. (ng/ml)	CAR18100004-T		CAR18100005-T		CAR18100006-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Carisoprodol	1000	5	0	5	0	5	0
	*2000	3	2	3	2	2	3
	3000	0	5	0	5	0	5

FLX500

Analytes	Conc. (ng/ml)	FLX18110001p-T		FLX18110002p-T		FLX18110003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Fluoxetine	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

RPD150

Analytes	Conc. (ng/ml)	RPD20090001p-T		RPD20090002p-T		RPD20090003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Risperidone	75	5	0	5	0	5	0
	*150	3	2	3	2	2	3
	225	0	5	0	5	0	5

CIT500

Analytes	Conc. (ng/ml)	CIT20120001p-T		CIT20120002p-T		CIT20120003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Desmethylcitalopram	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

FKET 1000

Analytes	Conc. (ng/ml)	FKE21010001p-T		FKE21010002p-T		FKE21010003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
F-Ketamine	500	5	0	5	0	5	0
	*1000	2	3	3	2	2	3
	1500	0	5	0	5	0	5

TAP 1000

Analytes	Conc. (ng/ml)	TAP21090001p-T		TAP21090002p-T		TAP21090003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Tapentadol	500	5	0	5	0	5	0
	*1000	3	2	3	2	2	3
	1500	0	5	0	5	0	5

SCOP 500

Analytes	Conc. (ng/ml)	SCOP21100004-T		SCOP21100005-T		SCOP21100006-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Scopolamine	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

MTZ 500

Analytes	Conc. (ng/ml)	MTZ21110007-T		MTZ21110008-T		MTZ21110009-T	
		Negative	Positive	Negative	Positive	Negative	Positive
Mirtazapine	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5
N-Desmethylmirtazapine	250	5	0	5	0	5	0
	*500	3	2	3	2	2	3
	750	0	5	0	5	0	5

NND 1000

Analytes	Conc. (ng/ml)	NND21120001p-T		NND21120002p-T		NND21120003p-T	
		Negative	Positive	Negative	Positive	Negative	Positive
N,N-Dimethyltryptamine	500	5	0	5	0	5	0
	*1000	2	3	3	2	2	3
	1500	0	5	0	5	0	5

Conclusion: The cutoff concentrations of each drug related compounds were showed in the tables with "*".

2.4 Cross Reactivity

A study was conducted to determine the cross-reactivity of the test with compounds in either drug-free urine or drug positive urine containing, Amphetamine, Barbiturates, Benzodiazepines, Buprenorphine, Cocaine, Marijuana, Methadone, Methamphetamine, Methylenedioxymethamphetamine, Morphine, Tramadol , Methaqualone, Ketamine ,Phencyclidine, Propoxyphene or Tricyclic Antidepressants and so on. The following compounds show no cross-reactivity when tested with the Multi-Drug Rapid Test Panel at a

concentration of 100 μg/mL. Visual interpretations occurred at 5 minutes after specimen application. Results are presented in table below.

Table: Cross Reactivity

Acetophenetidin	Cortisone	Zomepirac	d-Pseudoephedrine
N-Acetylprocainamide	Creatinine	Ketoprofen	Quinidine
Acetylsalicylic acid	Deoxycorticosterone	Labetalol	Quinine
Aminopyrine	Dextromethorphan	Loperamide	Salicylic acid
Amoxicillin	Diclofenac	Meprobamate	Serotonin
Ampicillin	Diflunisal	Methoxyphenamine	Sulfamethazine
l-Ascorbic acid	Digoxin	Methylphenidate	Sulindac
Apomorphine	Diphenhydramine	Nalidixic acid	Tetracycline
Aspartame	Ethyl-p-aminobenzoate	Naproxen	Tetrahydrocortisone, 3-acetate
Atropine	β-Estradiol	Niacinamide	
Benzilic acid	Estrone-3-sulfate	Nifedipine	Tetrahydrocortisone
Benzoic acid	Erythromycin	Norethindrone	Tetrahydrozoline
Bilirubin	Fenoprofen	Noscapine	Thiamine
d,l-Brompheniramine	Furosemide	d,l-Octopamine	Thioridazine
d,l-Propanolol	Gentisic acid	Oxalic acid	d,l-Tyrosine
Cannabidiol	Hemoglobin	Oxolinic acid	Tolbutamide
Chloral hydrate	Hydralazine	Oxymetazoline	Triamterene
Chloramphenicol	Hydrochlorothiazide	Papaverine	Trifluoperazine
Chlorothiazide	Hydrocortisone	Penicillin-G	Trimethoprim
d,l-Chlorpheniramine	o-Hydroxyhippuric acid	Perphenazine	d,l-Tryptophan
Chlorpromazine	3-Hydroxytyramine	Phenelzine	Uric acid
Cholesterol	d,l-Isoproterenol	Prednisone	Verapamil
Clonidine	Isoxsuprine		

Conclusion: The compounds listed in above table show no cross-reactivity when tested at a concentration of 100 μg/ml in 5 minutes.

2.5 Between Day Reproducibility

The concentration of all tests in the cup at 0ng/ml, -50%Cut-off, +50%Cut-off specimens was run on ten sparate days using the same lot of reagents. The specimens were tested in triplicate in 3 lots of test panel. Results were read visually at 5 minutes after specimen application. Results are presented in table below.

Table: Between Day Results

Day		批号	1	2	3	4	5	6	7	8	9	10
AMP 1,000	0ng/ml	AMP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	AMP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

	1,500ng/ml	AMP14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		AMP14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		AMP14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
AMP 500	0ng/ml	AMP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	AMP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	AMP14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		AMP14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		AMP14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
AMP 300	0ng/ml	AMP14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	AMP14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		AMP14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	AMP14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		AMP14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		AMP14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BAR 300	0ng/ml	BAR14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	BAR14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	BAR14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BAR14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BAR14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BAR 200	0ng/ml	BAR14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	BAR14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BAR14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	BAR14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		BAR14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BAR14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BUP10	0ng/ml	BUP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	5ng/ml	BUP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	BUP14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BUP14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BUP14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BUP5	0ng/ml	BUP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	2.5ng/ml	BUP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BUP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	7.5ng/ml	BUP14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BUP14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BUP14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BZO 300	0ng/ml	BZO14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	BZO14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	BZO14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BZO14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BZO14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BZO 200	0ng/ml	BZO14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	BZO14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	BZO14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BZO14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		BZO14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BZO 100	0ng/ml	BZO14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	BZO14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	BZO14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BZO14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BZO14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
BZO 500	0ng/ml	BZO14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	BZO14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		BZO14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	BZO14030019p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BZO14030020p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		BZO14030021p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
COC 300	0ng/ml	COC14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	COC14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	COC14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
COC 100	0ng/ml	COC14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	COC14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	COC14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

COC 150	0ng/ml	COC14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	COC14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	225ng/ml	COC14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
COC 200	0ng/ml	COC14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	COC14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COC14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	COC14030019p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030020p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COC14030021p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
KET 1000	0ng/ml	KET14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	KET14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	KET14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
KET 500	0ng/ml	KET14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	KET14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	KET14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
KET	0ng/ml	KET14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

300		KET14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	KET14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	KET14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
KET 100	0ng/ml	KET14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	KET14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KET14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	KET14030019p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030020p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KET14030021p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MDMA 1000	0ng/ml	MDMA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	MDMA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	MDMA14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDMA14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDMA14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MDMA 500	0ng/ml	MDMA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	MDMA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	MDMA14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDMA14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDMA14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MDMA 300	150ng/ml	MDMA14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		MDMA14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MDMA14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDMA14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	900ng/ml	MDMA14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDMA14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDMA14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MET 1000	0ng/ml	MET14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	MET14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	MET14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MET14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MET14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MET 500	0ng/ml	MET14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	MET14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	MET14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MET14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MET14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MET 300	0ng/ml	MET14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MET14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MET14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MET14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MET14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MET14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MOP 300	0ng/ml	MOP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

	150ng/ml	MOP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MOP14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MOP14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MOP14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MOP 200	0ng/ml	MOP14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	MOP14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	MOP14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MOP14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MOP14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MOP 100	0ng/ml	MOP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	MOP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MOP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MOP14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MOP14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MOP14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MQL 300	0ng/ml	MQL14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MQL14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MQL14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MQL14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MQL14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MQL14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MQL14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MQL14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MQL14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MTD 300	0ng/ml	MTD14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTD14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTD14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MTD14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		MTD14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTD14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MTD14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MTD14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MTD14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MTD 200	0ng/ml	MTD14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTD14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTD14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	MTD14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTD14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTD14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	MTD14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MTD14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MTD14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
OPI	0ng/ml	OPI14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OPI14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OPI14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1000ng/ml	OPI14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OPI14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OPI14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	3000ng/ml	OPI14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		OPI14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		OPI14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
PCP	0ng/ml	PCP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PCP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PCP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	12.5ng/ml	PCP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PCP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PCP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	37.5ng/ml	PCP14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PCP14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PCP14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
PPX	0ng/ml	PPX14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PPX14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PPX14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	PPX14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PPX14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		PPX14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	PPX14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PPX14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PPX14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
THC 150	0ng/ml	THC14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	THC14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	225ng/ml	THC14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
THC 50	0ng/ml	THC14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25ng/ml	THC14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	THC14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
THC 25	0ng/ml	THC14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	12.5ng/ml	THC14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	37.5ng/ml	THC14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
THC 300	0ng/ml	THC14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	THC14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

	450ng/ml	THC14030019p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030020p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030021p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
THC 20	0ng/ml	THC14030037p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030038p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030039p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	10ng/ml	THC14030037p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030038p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030039p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	30ng/ml	THC14030037p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030038p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030039p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
THC 200	0ng/ml	THC14030025p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030026p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030027p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	THC14030025p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030026p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030027p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	THC14030025p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030026p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030027p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
THC 30	0ng/ml	THC14030031p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030032p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030033p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	THC14030031p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030032p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		THC14030033p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	45ng/ml	THC14030031p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030032p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		THC14030033p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TCA 1000	0ng/ml	TCA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	TCA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	TCA14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		TCA14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TCA14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TCA 500	0ng/ml	TCA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	TCA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TCA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	TCA14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TCA14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TCA14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TML 100	0ng/ml	TML14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	TML14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	TML14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TML14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TML14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TML 200	0ng/ml	TML14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	TML14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	TML14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TML14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TML14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TML 300	0ng/ml	TML14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	TML14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TML14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	TML14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TML14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		TML14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
OXY 100	0ng/ml	OXY14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	OXY14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	OXY14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		OXY14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		OXY14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
OXY 300	0ng/ml	OXY14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	OXY14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		OXY14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	OXY14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		OXY14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		OXY14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
COT 200	0ng/ml	COT14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	COT14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	COT14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
COT 100	0ng/ml	COT14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	COT14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	COT14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

COT 500	0ng/ml	COT14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	COT14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	COT14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
COT 50	0ng/ml	COT14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25ng/ml	COT14030019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	COT14030019p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030020p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030021p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
COT 10	0ng/ml	COT14030025p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030026p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030027p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	5ng/ml	COT14030025p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030026p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		COT14030027p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	COT14030025p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030026p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		COT14030027p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
EDDP 300	0ng/ml	EDDP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		EDDP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		EDDP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	EDDP14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		EDDP14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		EDDP14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	EDDP14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		EDDP14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		EDDP14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
EDDP	0ng/ml	EDDP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

100		EDDP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		EDDP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	EDDP14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		EDDP14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		EDDP14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	EDDP14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		EDDP14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		EDDP14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
FYL 20	0ng/ml	FYL18040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	10ng/ml	FYL18040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	30ng/ml	FYL18040001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
FYL 10	0ng/ml	FYL18040007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	5ng/ml	FYL18040007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	FYL18040007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
FYL 200	0ng/ml	FYL18040019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	FYL18040019p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040020p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040021p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	FYL18040019p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040020p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040021p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
FYL 100	0ng/ml	FYL18040013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		FYL18040015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	FYL18040013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	FYL18040013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
FYL 300	0ng/ml	FYL18040025p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040026p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040027p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	FYL18040025p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040026p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FYL18040027p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	FYL18040025p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040026p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FYL18040027p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
K2 50	0ng/ml	SMA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25ng/ml	SMA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	SMA14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SMA14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SMA14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
K2 30	0ng/ml	SMA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	SMA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	45ng/ml	SMA14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SMA14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SMA14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
K2 25	0ng/ml	SMA14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

	12.5ng/ml	SMA14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SMA14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	37.5ng/ml	SMA14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SMA14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SMA14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
6-MAM10	0ng/ml	MAM14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MAM14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MAM14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	5ng/ml	MAM14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MAM14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MAM14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	MAM14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MAM14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MAM14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
DIA 300	0ng/ml	DIA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	DIA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450n0g/ml	DIA14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		DIA14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		DIA14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
DIA 200	0ng/ml	DIA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	DIA14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		DIA14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	DIA14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		DIA14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		DIA14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CLO 400	0ng/ml	CLO14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CLO14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CLO14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	200ng/ml	CLO14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		CLO14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CLO14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	600ng/ml	CLO14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CLO14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CLO14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CLO 150	0ng/ml	CLO14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CLO14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CLO14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	CLO14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CLO14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CLO14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	225ng/ml	CLO14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CLO14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CLO14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MDA 500	0ng/ml	MDA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	MDA14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDA14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDA14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	MDA14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDA14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDA14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ETG 300	0ng/ml	ETG14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	ETG14030013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	ETG14030013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ETG14030014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ETG14030015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ETG 500	0ng/ml	ETG14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	ETG14030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		ETG14030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	ETG14030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ETG14030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ETG14030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ETG 1000	0ng/ml	ETG14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	ETG14030007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ETG14030009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	ETG14030007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ETG14030008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ETG14030009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ACE 5000	0ng/ml	ACE14060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACE14060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACE14060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	2500ng/ml	ACE14060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACE14060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACE14060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	7500ng/ml	ACE14060001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACE14060002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACE14060003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
LSD 10	0ng/ml	LSD14060013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	5ng/ml	LSD14060013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	LSD14060013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		LSD14060014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		LSD14060015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
LSD 20	0ng/ml	LSD14060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	10ng/ml	LSD14060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

	30ng/ml	LSD14060001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		LSD14060002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		LSD14060003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
LSD 50	0ng/ml	LSD14060007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25ng/ml	LSD14060007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		LSD14060009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	LSD14060007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		LSD14060008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		LSD14060009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MDPV 1000	0ng/ml	MDPV14080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	MDPV14080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	MDPV14080001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDPV14080002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDPV14080003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MDPV 500	0ng/ml	MDPV14080007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	MDPV14080007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV14080009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	MDPV14080007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDPV14080008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDPV14080009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MDPV 300	0ng/ml	MDPV15090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV15090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV15090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MDPV15090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV15090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MDPV15090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MDPV15090001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		MDPV15090002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MDPV15090003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MPD 300	0ng/ml	MPD14080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MPD14080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MPD14080001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MPD14080002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MPD14080003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MPD 1000	0ng/ml	MPD14080007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	MPD14080007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPD14080009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	MPD14080007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MPD14080008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MPD14080009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ZOL 50	0ng/ml	ZOL15010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOL15010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOL15010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25ng/ml	ZOL15010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOL15010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOL15010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	ZOL15010001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ZOL15010002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ZOL15010003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ZOP 50	0ng/ml	ZOP16080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOP16080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOP16080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25ng/ml	ZOP16080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOP16080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZOP16080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	ZOP16080001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ZOP16080002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		ZOP16080003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MCAT 500	0ng/ml	MCAT16080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MCAT16080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MCAT16080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	MCAT16080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MCAT16080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MCAT16080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	MCAT16080001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MCAT16080002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MCAT16080003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
7-ACL 300	0ng/ml	ACL14100001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	ACL14100001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	ACL14100001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACL14100002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACL14100003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
7-ACL 100	0ng/ml	ACL14100007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	ACL14100007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	ACL14100007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACL14100008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACL14100009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
7-ACL 200	0ng/ml	ACL14100013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	ACL14100013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ACL14100015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	ACL14100013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACL14100014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ACL14100015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

CFYL 500	0ng/ml	CFYL15020001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	CFYL15020001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	CFYL15020001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CFYL15020002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CFYL15020003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CFYL 250	0ng/ml	CFYL15020007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	125ng/ml	CFYL15020007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CFYL15020009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	375ng/ml	CFYL15020007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CFYL15020008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CFYL15020009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CAF 1000	0ng/ml	CAF15080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAF15080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAF15080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	CAF15080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAF15080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAF15080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	CAF15080001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CAF15080002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CAF15080003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CAT 150	0ng/ml	CAT14080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAT14080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAT14080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	CAT14080001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAT14080002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAT14080003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	225ng/ml	CAT14080001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CAT14080002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CAT14080003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TRO	0ng/ml	TRO15050001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

350		TRO15050002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TRO15050003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	175ng/ml	TRO15050001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TRO15050002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TRO15050003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	525ng/ml	TRO15050001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TRO15050002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TRO15050003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MEP 100	0ng/ml	MEP15070001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MEP15070002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MEP15070003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	MEP15070001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MEP15070002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MEP15070003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MEP15070001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MEP15070002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MEP15070003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ALP 100	0ng/ml	ALP16040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ALP16040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ALP16040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	ALP16040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ALP16040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ALP16040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	ALP16040001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ALP16040002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ALP16040003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ABP 10	0ng/ml	ABP17010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ABP17010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ABP17010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	5ng/ml	ABP17010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ABP17010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ABP17010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	15ng/ml	ABP17010001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ABP17010002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ABP17010003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
APVP 1000	0ng/ml	APVP15050001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		APVP15050003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	APVP15050001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	APVP15050001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		APVP15050002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		APVP15050003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
APVP 500	0ng/ml	APVP15050007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	APVP15050007p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050008p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050009p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	APVP15050007p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		APVP15050008p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		APVP15050009p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
APVP 300	0ng/ml	APVP15050013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	APVP15050013p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050014p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		APVP15050015p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	APVP15050013p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		APVP15050014p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		APVP15050015p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CNB 500	0ng/ml	CNB18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CNB18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CNB18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	CNB18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CNB18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CNB18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	CNB18060001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CNB18060002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CNB18060003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MPRB 100	0ng/ml	MPRD18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPRD18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPRD18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

	50ng/ml	MPRD18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPRD18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MPRD18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MPRD18060001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MPRD18060002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MPRD18060003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
PGB 50,000	0ng/ml	PGB16010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB16010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB16010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25,000ng/ml	PGB16010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB16010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB16010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75,000ng/ml	PGB16010001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PGB16010002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PGB16010003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
PGB 500	0ng/ml	PGB19040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB19040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB19040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	PGB19040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB19040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PGB19040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	PGB19040001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PGB19040002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PGB19040003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TZD 200	0ng/ml	TZD18070001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TZD18070002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TZD18070003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	100ng/ml	TZD18070001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TZD18070002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TZD18070003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	300ng/ml	TZD18070001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TZD18070002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TZD18070003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
UR-14 4 25	0ng/ml	UR18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		UR18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		UR18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	12.5ng/ml	UR18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		UR18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		UR18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	37.5ng/ml	UR18060001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		UR18060002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		UR18060003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
ZAL 100	0ng/ml	ZAL18050001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZAL18050002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZAL18050003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	ZAL18050001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZAL18050002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		ZAL18050003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	ZAL18050001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ZAL18050002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		ZAL18050003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MES 100	0ng/ml	MES18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	50ng/ml	MES18060001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18060002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18060003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MES18060001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MES18060002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MES18060003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MES 300	0ng/ml	MES18090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	MES18090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MES18090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	450ng/ml	MES18090001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MES18090002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MES18090003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
GAB 2000	0ng/ml	GAB18070001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		GAB18070002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		GAB18070003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1000ng/ml	GAB18070001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		GAB18070002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

		GAB18070003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	3000ng/ml	GAB18070001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		GAB18070002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		GAB18070003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TLD 50	0ng/ml	TLD18110001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TLD18110002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TLD18110003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	25 ng/ml	TLD18110001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TLD18110002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TLD18110003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75 ng/ml	TLD18110001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TLD18110002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TLD18110003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
QTP 1000	0ng/ml	QTP19030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		QTP19030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		QTP19030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	QTP19030001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		QTP19030002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		QTP19030003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	QTP19030001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		QTP19030002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		QTP19030003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
PAP 500	0ng/ml	PAP18110001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PAP18110002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PAP18110003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	PAP18110001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PAP18110002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		PAP18110003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	PAP18110001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PAP18110002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		PAP18110003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
KRA 300	0ng/ml	KRA19040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KRA19040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KRA19040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	150ng/ml	KRA19040001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KRA19040002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		KRA19040003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-

	450ng/ml	KRA19040001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KRA19040002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		KRA19040003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CAR 2000	0ng/ml	CAR18100001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAR18100002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAR18100003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1000ng/ml	CAR18100001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAR18100002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CAR18100003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	3000ng/ml	CAR18100001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CAR18100002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CAR18100003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
FLX 500	0ng/ml	FLX18110001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FLX18110002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FLX18110003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	FLX18110001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FLX18110002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FLX18110003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	FLX18110001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FLX18110002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FLX18110003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
RPD 150	0ng/ml	RPD20090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		RPD20090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		RPD20090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	75ng/ml	RPD20090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		RPD20090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		RPD20090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	225ng/ml	RPD20090001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		RPD20090002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		RPD20090003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
CIT 500	0ng/ml	CIT20120001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CIT20120002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CIT20120003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	CIT20120001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CIT20120002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		CIT20120003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	CIT20120001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		CIT20120002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		CIT20120003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
FKET 1000	0ng/ml	FKE21010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FKE21010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FKE21010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	FKE21010001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FKE21010002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		FKE21010003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	FKE21010001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FKE21010002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		FKE21010003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
TAP 1000	0ng/ml	TAP21090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TAP21090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TAP21090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	TAP21090001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TAP21090002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		TAP21090003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	TAP21090001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TAP21090002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		TAP21090003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
SCOP 500	0ng/ml	SCOP21100004-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SCOP21100005-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SCOP21100006-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	SCOP21100004-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SCOP21100005-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		SCOP21100006-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	SCOP21100004-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SCOP21100005-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		SCOP21100006-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
MTZ 500	0ng/ml	MTZ21110007-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTZ21110008-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTZ21110009-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	250ng/ml	MTZ21110007-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTZ21110008-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		MTZ21110009-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	750ng/ml	MTZ21110007-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		MTZ21110008-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

		MTZ21110009-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
NND 1000	0ng/ml	NND21120001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		NND21120002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		NND21120003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	500ng/ml	NND21120001p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		NND21120002p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
		NND21120003p-T	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-	-,-,-
	1500ng/ml	NND21120001p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		NND21120002p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+
		NND21120003p-T	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+	+,+,+

Conclusion: Test results were consistent over the ten days period.

2.6 Between Lot Reproducibility (lot to lot variability)

Multi-Drug concentration at Drug free urine, -50% Cut-off, and +50% Cut-off urine specimens were run in replicates of ten in three separate lots of products. Results were read visually at 5 after specimen application. The test results are shown in table below.

Table: Between Lot Results

Day			1	2	3	4	5	6	7	8	9	10
AMP 1,000	0ng/ml	AMP14030001p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030002p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	AMP14030001p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030002p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030003p-T	-	-	-	-	-	-	-	-	-	-
	1,500ng/ml	AMP14030001p-T	+	+	+	+	+	+	+	+	+	+
		AMP14030002p-T	+	+	+	+	+	+	+	+	+	+
		AMP14030003p-T	+	+	+	+	+	+	+	+	+	+
AMP 500	0ng/ml	AMP14030007p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030008p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030009p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	AMP14030007p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030008p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030009p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	AMP14030007p-T	+	+	+	+	+	+	+	+	+	+
		AMP14030008p-T	+	+	+	+	+	+	+	+	+	+
		AMP14030009p-T	+	+	+	+	+	+	+	+	+	+
AMP	0ng/ml	AMP14030013p-T	-	-	-	-	-	-	-	-	-	-

300		AMP14030014p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030015p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	AMP14030013p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030014p-T	-	-	-	-	-	-	-	-	-	-
		AMP14030015p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	AMP14030013p-T	+	+	+	+	+	+	+	+	+	+
		AMP14030014p-T	+	+	+	+	+	+	+	+	+	+
		AMP14030015p-T	+	+	+	+	+	+	+	+	+	+
BAR 300	0ng/ml	BAR14030001p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030002p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	BAR14030001p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030002p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	BAR14030001p-T	+	+	+	+	+	+	+	+	+	+
		BAR14030002p-T	+	+	+	+	+	+	+	+	+	+
		BAR14030003p-T	+	+	+	+	+	+	+	+	+	+
BAR 200	0ng/ml	BAR14030007p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030008p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030009p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	BAR14030007p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030008p-T	-	-	-	-	-	-	-	-	-	-
		BAR14030009p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	BAR14030007p-T	+	+	+	+	+	+	+	+	+	+
		BAR14030008p-T	+	+	+	+	+	+	+	+	+	+
		BAR14030009p-T	+	+	+	+	+	+	+	+	+	+
BUP 10	0ng/ml	BUP14030001p-T	-	-	-	-	-	-	-	-	-	-
		BUP14030002p-T	-	-	-	-	-	-	-	-	-	-
		BUP14030003p-T	-	-	-	-	-	-	-	-	-	-
	5ng/ml	BUP14030001p-T	-	-	-	-	-	-	-	-	-	-
		BUP14030002p-T	-	-	-	-	-	-	-	-	-	-
		BUP14030003p-T	-	-	-	-	-	-	-	-	-	-
	15ng/ml	BUP14030001p-T	+	+	+	+	+	+	+	+	+	+
		BUP14030002p-T	+	+	+	+	+	+	+	+	+	+
		BUP14030003p-T	+	+	+	+	+	+	+	+	+	+
BUP 5	0ng/ml	BUP14030007p-T	-	-	-	-	-	-	-	-	-	-
		BUP14030008p-T	-	-	-	-	-	-	-	-	-	-

		BUP14030009p-T	-	-	-	-	-	-	-	-	-	-
	2.5ng/ml	BUP14030007p-T	-	-	-	-	-	-	-	-	-	-
		BUP14030008p-T	-	-	-	-	-	-	-	-	-	-
		BUP14030009p-T	-	-	-	-	-	-	-	-	-	-
	7.5ng/ml	BUP14030007p-T	+	+	+	+	+	+	+	+	+	+
		BUP14030008p-T	+	+	+	+	+	+	+	+	+	+
		BUP14030009p-T	+	+	+	+	+	+	+	+	+	+
BZO 300	0ng/ml	BZO14030001p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030002p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	BZO14030001p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030002p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	BZO14030001p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030002p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030003p-T	+	+	+	+	+	+	+	+	+	+
BZO 200	0ng/ml	BZO14030007p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030008p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030009p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	BZO14030007p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030008p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030009p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	BZO14030007p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030008p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030009p-T	+	+	+	+	+	+	+	+	+	+
BZO 100	0ng/ml	BZO14030013p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030014p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030015p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	BZO14030013p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030014p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030015p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	BZO14030013p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030014p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030015p-T	+	+	+	+	+	+	+	+	+	+
BZO 500	0ng/ml	BZO14030019p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030020p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030021p-T	-	-	-	-	-	-	-	-	-	-

	250ng/ml	BZO14030019p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030020p-T	-	-	-	-	-	-	-	-	-	-
		BZO14030021p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	BZO14030019p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030020p-T	+	+	+	+	+	+	+	+	+	+
		BZO14030021p-T	+	+	+	+	+	+	+	+	+	+
COC 300	0ng/ml	COC14030001p-T	-	-	-	-	-	-	-	-	-	-
		COC14030002p-T	-	-	-	-	-	-	-	-	-	-
		COC14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	COC14030001p-T	-	-	-	-	-	-	-	-	-	-
		COC14030002p-T	-	-	-	-	-	-	-	-	-	-
		COC14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	COC14030001p-T	+	+	+	+	+	+	+	+	+	+
		COC14030002p-T	+	+	+	+	+	+	+	+	+	+
		COC14030003p-T	+	+	+	+	+	+	+	+	+	+
COC 100	0ng/ml	COC14030007p-T	-	-	-	-	-	-	-	-	-	-
		COC14030008p-T	-	-	-	-	-	-	-	-	-	-
		COC14030009p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	COC14030007p-T	-	-	-	-	-	-	-	-	-	-
		COC14030008p-T	-	-	-	-	-	-	-	-	-	-
		COC14030009p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	COC14030007p-T	+	+	+	+	+	+	+	+	+	+
		COC14030008p-T	+	+	+	+	+	+	+	+	+	+
		COC14030009p-T	+	+	+	+	+	+	+	+	+	+
COC 150	0ng/ml	COC14030013p-T	-	-	-	-	-	-	-	-	-	-
		COC14030014p-T	-	-	-	-	-	-	-	-	-	-
		COC14030015p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	COC14030013p-T	-	-	-	-	-	-	-	-	-	-
		COC14030014p-T	-	-	-	-	-	-	-	-	-	-
		COC14030015p-T	-	-	-	-	-	-	-	-	-	-
	225ng/ml	COC14030013p-T	+	+	+	+	+	+	+	+	+	+
		COC14030014p-T	+	+	+	+	+	+	+	+	+	+
		COC14030015p-T	+	+	+	+	+	+	+	+	+	+
COC 200	0ng/ml	COC14030019p-T	-	-	-	-	-	-	-	-	-	-
		COC14030020p-T	-	-	-	-	-	-	-	-	-	-
		COC14030021p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	COC14030019p-T	-	-	-	-	-	-	-	-	-	-

		COC14030020p-T	-	-	-	-	-	-	-	-	-	-
		COC14030021p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	COC14030019p-T	+	+	+	+	+	+	+	+	+	+
		COC14030020p-T	+	+	+	+	+	+	+	+	+	+
		COC14030021p-T	+	+	+	+	+	+	+	+	+	+
KET 1000	0ng/ml	KET14030001p-T	-	-	-	-	-	-	-	-	-	-
		KET14030002p-T	-	-	-	-	-	-	-	-	-	-
		KET14030003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	KET14030001p-T	-	-	-	-	-	-	-	-	-	-
		KET14030002p-T	-	-	-	-	-	-	-	-	-	-
		KET14030003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	KET14030001p-T	+	+	+	+	+	+	+	+	+	+
		KET14030002p-T	+	+	+	+	+	+	+	+	+	+
		KET14030003p-T	+	+	+	+	+	+	+	+	+	+
KET 500	0ng/ml	KET14030007p-T	-	-	-	-	-	-	-	-	-	-
		KET14030008p-T	-	-	-	-	-	-	-	-	-	-
		KET14030009p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	KET14030007p-T	-	-	-	-	-	-	-	-	-	-
		KET14030008p-T	-	-	-	-	-	-	-	-	-	-
		KET14030009p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	KET14030007p-T	+	+	+	+	+	+	+	+	+	+
		KET14030008p-T	+	+	+	+	+	+	+	+	+	+
		KET14030009p-T	+	+	+	+	+	+	+	+	+	+
KET 300	0ng/ml	KET14030013p-T	-	-	-	-	-	-	-	-	-	-
		KET14030014p-T	-	-	-	-	-	-	-	-	-	-
		KET14030015p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	KET14030013p-T	-	-	-	-	-	-	-	-	-	-
		KET14030014p-T	-	-	-	-	-	-	-	-	-	-
		KET14030015p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	KET14030013p-T	+	+	+	+	+	+	+	+	+	+
		KET14030014p-T	+	+	+	+	+	+	+	+	+	+
		KET14030015p-T	+	+	+	+	+	+	+	+	+	+
KET 100	0ng/ml	KET14030019p-T	-	-	-	-	-	-	-	-	-	-
		KET14030020p-T	-	-	-	-	-	-	-	-	-	-
		KET14030021p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	KET14030019p-T	-	-	-	-	-	-	-	-	-	-
		KET14030020p-T	-	-	-	-	-	-	-	-	-	-

		KET14030021p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	KET14030019p-T	+	+	+	+	+	+	+	+	+	+
		KET14030020p-T	+	+	+	+	+	+	+	+	+	+
		KET14030021p-T	+	+	+	+	+	+	+	+	+	+
MDMA 1000	0ng/ml	MDMA14030007p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030008p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030009p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	MDMA14030007p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030008p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030009p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	MDMA14030007p-T	+	+	+	+	+	+	+	+	+	+
		MDMA14030008p-T	+	+	+	+	+	+	+	+	+	+
		MDMA14030009p-T	+	+	+	+	+	+	+	+	+	+
MDMA 500	0ng/ml	MDMA14030001p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030002p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	MDMA14030001p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030002p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	MDMA14030001p-T	+	+	+	+	+	+	+	+	+	+
		MDMA14030002p-T	+	+	+	+	+	+	+	+	+	+
		MDMA14030003p-T	+	+	+	+	+	+	+	+	+	+

MDMA 300	150ng/ml	MDMA14030013p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030014p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030015p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MDMA14030013p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030014p-T	-	-	-	-	-	-	-	-	-	-
		MDMA14030015p-T	-	-	-	-	-	-	-	-	-	-
	900ng/ml	MDMA14030013p-T	+	+	+	+	+	+	+	+	+	+
		MDMA14030014p-T	+	+	+	+	+	+	+	+	+	+
		MDMA14030015p-T	+	+	+	+	+	+	+	+	+	+
MET 1000	0ng/ml	MET14030001p-T	-	-	-	-	-	-	-	-	-	-
		MET14030002p-T	-	-	-	-	-	-	-	-	-	-
		MET14030003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	MET14030001p-T	-	-	-	-	-	-	-	-	-	-
		MET14030002p-T	-	-	-	-	-	-	-	-	-	-
		MET14030003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	MET14030001p-T	+	+	+	+	+	+	+	+	+	+
		MET14030002p-T	+	+	+	+	+	+	+	+	+	+
		MET14030003p-T	+	+	+	+	+	+	+	+	+	+
MET 500	0ng/ml	MET14030007p-T	-	-	-	-	-	-	-	-	-	-
		MET14030008p-T	-	-	-	-	-	-	-	-	-	-
		MET14030009p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	MET14030007p-T	-	-	-	-	-	-	-	-	-	-
		MET14030008p-T	-	-	-	-	-	-	-	-	-	-
		MET14030009p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	MET14030007p-T	+	+	+	+	+	+	+	+	+	+
		MET14030008p-T	+	+	+	+	+	+	+	+	+	+
		MET14030009p-T	+	+	+	+	+	+	+	+	+	+
MET 300	0ng/ml	MET14030013p-T	-	-	-	-	-	-	-	-	-	-
		MET14030014p-T	-	-	-	-	-	-	-	-	-	-
		MET14030015p-T	-	-	-	-	-	-	-	-	-	-

	150ng/ml	MET14030013p-T	-	-	-	-	-	-	-	-	-	-
		MET14030014p-T	-	-	-	-	-	-	-	-	-	-
		MET14030015p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MET14030013p-T	+	+	+	+	+	+	+	+	+	+
		MET14030014p-T	+	+	+	+	+	+	+	+	+	+
		MET14030015p-T	+	+	+	+	+	+	+	+	+	+
MOP 300	0ng/ml	MOP14030001p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030002p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MOP14030001p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030002p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MOP14030001p-T	+	+	+	+	+	+	+	+	+	+
		MOP14030002p-T	+	+	+	+	+	+	+	+	+	+
		MOP14030003p-T	+	+	+	+	+	+	+	+	+	+
MOP 100	0ng/ml	MOP14030007p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030008p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030009p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	MOP14030007p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030008p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030009p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MOP14030007p-T	+	+	+	+	+	+	+	+	+	+
		MOP14030008p-T	+	+	+	+	+	+	+	+	+	+
		MOP14030009p-T	+	+	+	+	+	+	+	+	+	+
MOP 200	0ng/ml	MOP14030013p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030014p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030015p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	MOP14030013p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030014p-T	-	-	-	-	-	-	-	-	-	-
		MOP14030015p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	MOP14030013p-T	+	+	+	+	+	+	+	+	+	+
		MOP14030014p-T	+	+	+	+	+	+	+	+	+	+
		MOP14030015p-T	+	+	+	+	+	+	+	+	+	+
MQL 300	0ng/ml	MQL14030001p-T	-	-	-	-	-	-	-	-	-	-
		MQL14030002p-T	-	-	-	-	-	-	-	-	-	-
		MQL14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MQL14030001p-T	-	-	-	-	-	-	-	-	-	-

		MQL14030002p-T	-	-	-	-	-	-	-	-	-	-
		MQL14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MQL14030001p-T	+	+	+	+	+	+	+	+	+	+
		MQL14030002p-T	+	+	+	+	+	+	+	+	+	+
		MQL14030003p-T	+	+	+	+	+	+	+	+	+	+
MTD 300	0ng/ml	MTD14030001p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030002p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MTD14030001p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030002p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MTD14030001p-T	+	+	+	+	+	+	+	+	+	+
		MTD14030002p-T	+	+	+	+	+	+	+	+	+	+
		MTD14030003p-T	+	+	+	+	+	+	+	+	+	+
MTD 200	0ng/ml	MTD14030007p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030008p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030009p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	MTD14030007p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030008p-T	-	-	-	-	-	-	-	-	-	-
		MTD14030009p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	MTD14030007p-T	+	+	+	+	+	+	+	+	+	+
		MTD14030008p-T	+	+	+	+	+	+	+	+	+	+
		MTD14030009p-T	+	+	+	+	+	+	+	+	+	+
OPI 2000	0ng/ml	OPI14030001p-T	-	-	-	-	-	-	-	-	-	-
		OPI14030002p-T	-	-	-	-	-	-	-	-	-	-
		OPI14030003p-T	-	-	-	-	-	-	-	-	-	-
	1000ng/ml	OPI14030001p-T	-	-	-	-	-	-	-	-	-	-
		OPI14030002p-T	-	-	-	-	-	-	-	-	-	-
		OPI14030003p-T	-	-	-	-	-	-	-	-	-	-
	3000ng/ml	OPI14030001p-T	+	+	+	+	+	+	+	+	+	+
		OPI14030002p-T	+	+	+	+	+	+	+	+	+	+
		OPI14030003p-T	+	+	+	+	+	+	+	+	+	+
PCP 25	0ng/ml	PCP14030001p-T	-	-	-	-	-	-	-	-	-	-
		PCP14030002p-T	-	-	-	-	-	-	-	-	-	-
		PCP14030003p-T	-	-	-	-	-	-	-	-	-	-
	12.5ng/ml	PCP14030001p-T	-	-	-	-	-	-	-	-	-	-
		PCP14030002p-T	-	-	-	-	-	-	-	-	-	-

		PCP14030003p-T	-	-	-	-	-	-	-	-	-	-
	37.5ng/ml	PCP14030001p-T	+	+	+	+	+	+	+	+	+	+
		PCP14030002p-T	+	+	+	+	+	+	+	+	+	+
		PCP14030003p-T	+	+	+	+	+	+	+	+	+	+
PPX 300	0ng/ml	PPX14030001p-T	-	-	-	-	-	-	-	-	-	-
		PPX14030002p-T	-	-	-	-	-	-	-	-	-	-
		PPX14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	PPX14030001p-T	-	-	-	-	-	-	-	-	-	-
		PPX14030002p-T	-	-	-	-	-	-	-	-	-	-
		PPX14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	PPX14030001p-T	+	+	+	+	+	+	+	+	+	+
		PPX14030002p-T	+	+	+	+	+	+	+	+	+	+
		PPX14030003p-T	+	+	+	+	+	+	+	+	+	+
THC 150	0ng/ml	THC14030001p-T	-	-	-	-	-	-	-	-	-	-
		THC14030002p-T	-	-	-	-	-	-	-	-	-	-
		THC14030003p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	THC14030001p-T	-	-	-	-	-	-	-	-	-	-
		THC14030002p-T	-	-	-	-	-	-	-	-	-	-
		THC14030003p-T	-	-	-	-	-	-	-	-	-	-
	225ng/ml	THC14030001p-T	+	+	+	+	+	+	+	+	+	+
		THC14030002p-T	+	+	+	+	+	+	+	+	+	+
		THC14030003p-T	+	+	+	+	+	+	+	+	+	+
THC 50	0ng/ml	THC14030007p-T	-	-	-	-	-	-	-	-	-	-
		THC14030008p-T	-	-	-	-	-	-	-	-	-	-
		THC14030009p-T	-	-	-	-	-	-	-	-	-	-
	25ng/ml	THC14030007p-T	-	-	-	-	-	-	-	-	-	-
		THC14030008p-T	-	-	-	-	-	-	-	-	-	-
		THC14030009p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	THC14030007p-T	+	+	+	+	+	+	+	+	+	+
		THC14030008p-T	+	+	+	+	+	+	+	+	+	+
		THC14030009p-T	+	+	+	+	+	+	+	+	+	+
THC 25	0ng/ml	THC14030013p-T	-	-	-	-	-	-	-	-	-	-
		THC14030014p-T	-	-	-	-	-	-	-	-	-	-
		THC14030015p-T	-	-	-	-	-	-	-	-	-	-
	12.5ng/ml	THC14030013p-T	-	-	-	-	-	-	-	-	-	-
		THC14030014p-T	-	-	-	-	-	-	-	-	-	-
		THC14030015p-T	-	-	-	-	-	-	-	-	-	-

	37.5ng/ml	THC14030013p-T	+	+	+	+	+	+	+	+	+	+
		THC14030014p-T	+	+	+	+	+	+	+	+	+	+
		THC14030015p-T	+	+	+	+	+	+	+	+	+	+
THC 300	0ng/ml	THC14030019p-T	-	-	-	-	-	-	-	-	-	-
		THC14030020p-T	-	-	-	-	-	-	-	-	-	-
		THC14030021p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	THC14030019p-T	-	-	-	-	-	-	-	-	-	-
		THC14030020p-T	-	-	-	-	-	-	-	-	-	-
		THC14030021p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	THC14030019p-T	+	+	+	+	+	+	+	+	+	+
		THC14030020p-T	+	+	+	+	+	+	+	+	+	+
		THC14030021p-T	+	+	+	+	+	+	+	+	+	+
THC 20	0ng/ml	THC14030037p-T	-	-	-	-	-	-	-	-	-	-
		THC14030038p-T	-	-	-	-	-	-	-	-	-	-
		THC14030039p-T	-	-	-	-	-	-	-	-	-	-
	10ng/ml	THC14030037p-T	-	-	-	-	-	-	-	-	-	-
		THC14030038p-T	-	-	-	-	-	-	-	-	-	-
		THC14030039p-T	-	-	-	-	-	-	-	-	-	-
	30ng/ml	THC14030037p-T	+	+	+	+	+	+	+	+	+	+
		THC14030038p-T	+	+	+	+	+	+	+	+	+	+
		THC14030039p-T	+	+	+	+	+	+	+	+	+	+
THC 200	0ng/ml	THC140300025p-T	-	-	-	-	-	-	-	-	-	-
		THC140300026p-T	-	-	-	-	-	-	-	-	-	-
		THC140300027p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	THC140300025p-T	-	-	-	-	-	-	-	-	-	-
		THC140300026p-T	-	-	-	-	-	-	-	-	-	-
		THC140300027p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	THC140300025p-T	+	+	+	+	+	+	+	+	+	+
		THC140300026p-T	+	+	+	+	+	+	+	+	+	+
		THC140300027p-T	+	+	+	+	+	+	+	+	+	+
THC 30	0ng/ml	THC140300031p-T	-	-	-	-	-	-	-	-	-	-
		THC140300032p-T	-	-	-	-	-	-	-	-	-	-
		THC140300033p-T	-	-	-	-	-	-	-	-	-	-
	15ng/ml	THC140300031p-T	-	-	-	-	-	-	-	-	-	-
		THC140300032p-T	-	-	-	-	-	-	-	-	-	-
		THC140300033p-T	-	-	-	-	-	-	-	-	-	-
	45ng/ml	THC140300031p-T	+	+	+	+	+	+	+	+	+	+

		THC140300032p-T	+	+	+	+	+	+	+	+	+	+
		THC140300033p-T	+	+	+	+	+	+	+	+	+	+
TCA 1000	0ng/ml	TCA14030001p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030002p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	TCA14030001p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030002p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	TCA14030001p-T	+	+	+	+	+	+	+	+	+	+
		TCA14030002p-T	+	+	+	+	+	+	+	+	+	+
		TCA14030003p-T	+	+	+	+	+	+	+	+	+	+
TCA 500	0ng/ml	TCA14030007p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030008p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030009p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	TCA14030007p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030008p-T	-	-	-	-	-	-	-	-	-	-
		TCA14030009p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	TCA14030007p-T	+	+	+	+	+	+	+	+	+	+
		TCA14030008p-T	+	+	+	+	+	+	+	+	+	+
		TCA14030009p-T	+	+	+	+	+	+	+	+	+	+
TML 100	0ng/ml	TML14030001p-T	-	-	-	-	-	-	-	-	-	-
		TML14030002p-T	-	-	-	-	-	-	-	-	-	-
		TML14030003p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	TML14030001p-T	-	-	-	-	-	-	-	-	-	-
		TML14030002p-T	-	-	-	-	-	-	-	-	-	-
		TML14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	TML14030001p-T	+	+	+	+	+	+	+	+	+	+
		TML14030002p-T	+	+	+	+	+	+	+	+	+	+
		TML14030003p-T	+	+	+	+	+	+	+	+	+	+
TML 200	0ng/ml	TML14030007p-T	-	-	-	-	-	-	-	-	-	-
		TML14030008p-T	-	-	-	-	-	-	-	-	-	-
		TML14030009p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	TML14030007p-T	-	-	-	-	-	-	-	-	-	-
		TML14030008p-T	-	-	-	-	-	-	-	-	-	-
		TML14030009p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	TML14030007p-T	+	+	+	+	+	+	+	+	+	+
		TML14030008p-T	+	+	+	+	+	+	+	+	+	+

		TML14030009p-T	+	+	+	+	+	+	+	+	+	+
TML 300	0ng/ml	TML14030013p-T	-	-	-	-	-	-	-	-	-	-
		TML14030014p-T	-	-	-	-	-	-	-	-	-	-
		TML14030015p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	TML14030013p-T	-	-	-	-	-	-	-	-	-	-
		TML14030014p-T	-	-	-	-	-	-	-	-	-	-
		TML14030015p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	TML14030013p-T	+	+	+	+	+	+	+	+	+	+
		TML14030014p-T	+	+	+	+	+	+	+	+	+	+
		TML14030015p-T	+	+	+	+	+	+	+	+	+	+
OXY100	0ng/ml	OXY14030001p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030002p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030003p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	OXY14030001p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030002p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	OXY14030001p-T	+	+	+	+	+	+	+	+	+	+
		OXY14030002p-T	+	+	+	+	+	+	+	+	+	+
		OXY14030003p-T	+	+	+	+	+	+	+	+	+	+
OXY300	0ng/ml	OXY14030007p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030008p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030009p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	OXY14030007p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030008p-T	-	-	-	-	-	-	-	-	-	-
		OXY14030009p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	OXY14030007p-T	+	+	+	+	+	+	+	+	+	+
		OXY14030008p-T	+	+	+	+	+	+	+	+	+	+
		OXY14030009p-T	+	+	+	+	+	+	+	+	+	+
COT 200	0ng/ml	COT14030001p-T	-	-	-	-	-	-	-	-	-	-
		COT14030002p-T	-	-	-	-	-	-	-	-	-	-
		COT14030003p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	COT14030001p-T	-	-	-	-	-	-	-	-	-	-
		COT14030002p-T	-	-	-	-	-	-	-	-	-	-
		COT14030003p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	COT14030001p-T	+	+	+	+	+	+	+	+	+	+
		COT14030002p-T	+	+	+	+	+	+	+	+	+	+
		COT14030003p-T	+	+	+	+	+	+	+	+	+	+

COT 100	0ng/ml	COT14030007p-T	-	-	-	-	-	-	-	-	-	-
		COT14030008p-T	-	-	-	-	-	-	-	-	-	-
		COT14030009p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	COT14030007p-T	-	-	-	-	-	-	-	-	-	-
		COT14030008p-T	-	-	-	-	-	-	-	-	-	-
		COT14030009p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	COT14030007p-T	+	+	+	+	+	+	+	+	+	+
		COT14030008p-T	+	+	+	+	+	+	+	+	+	+
		COT14030009p-T	+	+	+	+	+	+	+	+	+	+
COT 500	0ng/ml	COT14030013p-T	-	-	-	-	-	-	-	-	-	-
		COT14030014p-T	-	-	-	-	-	-	-	-	-	-
		COT14030015p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	COT14030013p-T	-	-	-	-	-	-	-	-	-	-
		COT14030014p-T	-	-	-	-	-	-	-	-	-	-
		COT14030015p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	COT14030013p-T	+	+	+	+	+	+	+	+	+	+
		COT14030014p-T	+	+	+	+	+	+	+	+	+	+
		COT14030015p-T	+	+	+	+	+	+	+	+	+	+
COT 50	0ng/ml	COT14030019p-T	-	-	-	-	-	-	-	-	-	-
		COT14030020p-T	-	-	-	-	-	-	-	-	-	-
		COT14030021p-T	-	-	-	-	-	-	-	-	-	-
	25ng/ml	COT14030019p-T	-	-	-	-	-	-	-	-	-	-
		COT14030020p-T	-	-	-	-	-	-	-	-	-	-
		COT14030021p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	COT14030019p-T	+	+	+	+	+	+	+	+	+	+
		COT14030020p-T	+	+	+	+	+	+	+	+	+	+
		COT14030021p-T	+	+	+	+	+	+	+	+	+	+
COT 10	0ng/ml	COT14030025p-T	-	-	-	-	-	-	-	-	-	-
		COT14030026p-T	-	-	-	-	-	-	-	-	-	-
		COT14030027p-T	-	-	-	-	-	-	-	-	-	-
	5ng/ml	COT14030025p-T	-	-	-	-	-	-	-	-	-	-
		COT14030026p-T	-	-	-	-	-	-	-	-	-	-
		COT14030027p-T	-	-	-	-	-	-	-	-	-	-
	15ng/ml	COT14030025p-T	+	+	+	+	+	+	+	+	+	+
		COT14030026p-T	+	+	+	+	+	+	+	+	+	+
		COT14030027p-T	+	+	+	+	+	+	+	+	+	+
EDDP	0ng/ml	EDDP14030001p-T	-	-	-	-	-	-	-	-	-	-

300		EDDP14030002p-T	-	-	-	-	-	-	-	-	-	-
		EDDP14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	EDDP14030001p-T	-	-	-	-	-	-	-	-	-	-
		EDDP14030002p-T	-	-	-	-	-	-	-	-	-	-
		EDDP14030003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	EDDP14030001p-T	+	+	+	+	+	+	+	+	+	+
		EDDP14030002p-T	+	+	+	+	+	+	+	+	+	+
		EDDP14030003p-T	+	+	+	+	+	+	+	+	+	+
EDDP 100	0ng/ml	EDDP14030007p-T	-	-	-	-	-	-	-	-	-	-
		EDDP14030008p-T	-	-	-	-	-	-	-	-	-	-
		EDDP14030009p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	EDDP14030007p-T	-	-	-	-	-	-	-	-	-	-
		EDDP14030008p-T	-	-	-	-	-	-	-	-	-	-
		EDDP14030009p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	EDDP14030007p-T	+	+	+	+	+	+	+	+	+	+
		EDDP14030008p-T	+	+	+	+	+	+	+	+	+	+
		EDDP14030009p-T	+	+	+	+	+	+	+	+	+	+
FYL 20	0ng/ml	FYL18040001p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040002p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040003p-T	-	-	-	-	-	-	-	-	-	-
	10ng/ml	FYL18040001p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040002p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040003p-T	-	-	-	-	-	-	-	-	-	-
	30ng/ml	FYL18040001p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040002p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040003p-T	+	+	+	+	+	+	+	+	+	+
FYL 10	0ng/ml	FYL18040007p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040008p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040009p-T	-	-	-	-	-	-	-	-	-	-
	5ng/ml	FYL18040007p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040008p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040009p-T	-	-	-	-	-	-	-	-	-	-
	15ng/ml	FYL18040007p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040008p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040009p-T	+	+	+	+	+	+	+	+	+	+
FYL 200	0ng/ml	FYL18040019p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040020p-T	-	-	-	-	-	-	-	-	-	-

		FYL18040021p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	FYL18040019p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040020p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040021p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	FYL18040019p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040020p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040021p-T	+	+	+	+	+	+	+	+	+	+
FYL 100	0ng/ml	FYL18040013p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040014p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040015p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	FYL18040013p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040014p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040015p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	FYL18040013p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040014p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040015p-T	+	+	+	+	+	+	+	+	+	+
FYL 300	0ng/ml	FYL18040025p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040026p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040027p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	FYL18040025p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040026p-T	-	-	-	-	-	-	-	-	-	-
		FYL18040027p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	FYL18040025p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040026p-T	+	+	+	+	+	+	+	+	+	+
		FYL18040027p-T	+	+	+	+	+	+	+	+	+	+
K2 50	0ng/ml	SMA14030001p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030002p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030003p-T	-	-	-	-	-	-	-	-	-	-
	25ng/ml	SMA14030001p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030002p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030003p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	SMA14030001p-T	+	+	+	+	+	+	+	+	+	+
		SMA14030002p-T	+	+	+	+	+	+	+	+	+	+
		SMA14030003p-T	+	+	+	+	+	+	+	+	+	+
K2 30	0ng/ml	SMA14030007p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030008p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030009p-T	-	-	-	-	-	-	-	-	-	-

	15ng/ml	SMA14030007p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030008p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030009p-T	-	-	-	-	-	-	-	-	-	-
	45ng/ml	SMA14030007p-T	+	+	+	+	+	+	+	+	+	+
		SMA14030008p-T	+	+	+	+	+	+	+	+	+	+
		SMA14030009p-T	+	+	+	+	+	+	+	+	+	+
K2 25	0ng/ml	SMA14030013p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030014p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030015p-T	-	-	-	-	-	-	-	-	-	-
	12.5ng/ml	SMA14030013p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030014p-T	-	-	-	-	-	-	-	-	-	-
		SMA14030015p-T	-	-	-	-	-	-	-	-	-	-
	37.5ng/ml	SMA14030013p-T	+	+	+	+	+	+	+	+	+	+
		SMA14030014p-T	+	+	+	+	+	+	+	+	+	+
		SMA14030015p-T	+	+	+	+	+	+	+	+	+	+
6-MAM 10	0ng/ml	MAM14030001p-T	-	-	-	-	-	-	-	-	-	-
		MAM14030002p-T	-	-	-	-	-	-	-	-	-	-
		MAM14030003p-T	-	-	-	-	-	-	-	-	-	-
	5ng/ml	MAM14030001p-T	-	-	-	-	-	-	-	-	-	-
		MAM14030002p-T	-	-	-	-	-	-	-	-	-	-
		MAM14030003p-T	-	-	-	-	-	-	-	-	-	-
	15ng/ml	MAM14030001p-T	+	+	+	+	+	+	+	+	+	+
		MAM14030002p-T	+	+	+	+	+	+	+	+	+	+
		MAM14030003p-T	+	+	+	+	+	+	+	+	+	+
DIA300	0ng/ml	DIA14030001p-T	-	-	-	-	-	-	-	-	-	-
		DIA14030002p-T	-	-	-	-	-	-	-	-	-	-
		DIA14030003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	DIA14030001p-T	-	-	-	-	-	-	-	-	-	-
		DIA14030002p-T	-	-	-	-	-	-	-	-	-	-
		DIA14030003p-T	-	-	-	-	-	-	-	-	-	-
	450n0g/ml	DIA14030001p-T	+	+	+	+	+	+	+	+	+	+
		DIA14030002p-T	+	+	+	+	+	+	+	+	+	+
		DIA14030003p-T	+	+	+	+	+	+	+	+	+	+
DIA 200	0ng/ml	DIA14030007p-T	-	-	-	-	-	-	-	-	-	-
		DIA14030008p-T	-	-	-	-	-	-	-	-	-	-
		DIA14030009p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	DIA14030007p-T	-	-	-	-	-	-	-	-	-	-

		DIA14030008p-T	-	-	-	-	-	-	-	-	-	-
		DIA14030009p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	DIA14030007p-T	+	+	+	+	+	+	+	+	+	+
		DIA14030008p-T	+	+	+	+	+	+	+	+	+	+
		DIA14030009p-T	+	+	+	+	+	+	+	+	+	+
CLO 400	0ng/ml	CLO14030001p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030002p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030003p-T	-	-	-	-	-	-	-	-	-	-
	200ng/ml	CLO14030001p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030002p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030003p-T	-	-	-	-	-	-	-	-	-	-
	600ng/ml	CLO14030001p-T	+	+	+	+	+	+	+	+	+	+
		CLO14030002p-T	+	+	+	+	+	+	+	+	+	+
		CLO14030003p-T	+	+	+	+	+	+	+	+	+	+
CLO 150	0ng/ml	CLO14030007p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030008p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030009p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	CLO14030007p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030008p-T	-	-	-	-	-	-	-	-	-	-
		CLO14030009p-T	-	-	-	-	-	-	-	-	-	-
	225ng/ml	CLO14030007p-T	+	+	+	+	+	+	+	+	+	+
		CLO14030008p-T	+	+	+	+	+	+	+	+	+	+
		CLO14030009p-T	+	+	+	+	+	+	+	+	+	+
MDA 500	0ng/ml	MDA14030001p-T	-	-	-	-	-	-	-	-	-	-
		MDA14030002p-T	-	-	-	-	-	-	-	-	-	-
		MDA14030003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	MDA14030001p-T	-	-	-	-	-	-	-	-	-	-
		MDA14030002p-T	-	-	-	-	-	-	-	-	-	-
		MDA14030003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	MDA14030001p-T	+	+	+	+	+	+	+	+	+	+
		MDA14030002p-T	+	+	+	+	+	+	+	+	+	+
		MDA14030003p-T	+	+	+	+	+	+	+	+	+	+
ETG 300	0ng/ml	ETG14030013p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030014p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030015p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	ETG14030013p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030014p-T	-	-	-	-	-	-	-	-	-	-

		ETG14030015p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	ETG14030013p-T	+	+	+	+	+	+	+	+	+	+
		ETG14030014p-T	+	+	+	+	+	+	+	+	+	+
		ETG14030015p-T	+	+	+	+	+	+	+	+	+	+
ETG 500	0ng/ml	ETG14030001p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030002p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	ETG14030001p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030002p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	ETG14030001p-T	+	+	+	+	+	+	+	+	+	+
		ETG14030002p-T	+	+	+	+	+	+	+	+	+	+
		ETG14030003p-T	+	+	+	+	+	+	+	+	+	+
ETG 1000	0ng/ml	ETG14030007p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030008p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030009p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	ETG14030007p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030008p-T	-	-	-	-	-	-	-	-	-	-
		ETG14030009p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	ETG14030007p-T	+	+	+	+	+	+	+	+	+	+
		ETG14030008p-T	+	+	+	+	+	+	+	+	+	+
		ETG14030009p-T	+	+	+	+	+	+	+	+	+	+
ACE 5000	0ng/ml	ACE14060001p-T	-	-	-	-	-	-	-	-	-	-
		ACE14060002p-T	-	-	-	-	-	-	-	-	-	-
		ACE14060003p-T	-	-	-	-	-	-	-	-	-	-
	2500ng/ml	ACE14060001p-T	-	-	-	-	-	-	-	-	-	-
		ACE14060002p-T	-	-	-	-	-	-	-	-	-	-
		ACE14060003p-T	-	-	-	-	-	-	-	-	-	-
	7500ng/ml	ACE14060001p-T	+	+	+	+	+	+	+	+	+	+
		ACE14060002p-T	+	+	+	+	+	+	+	+	+	+
		ACE14060003p-T	+	+	+	+	+	+	+	+	+	+
LSD 10	0ng/ml	LSD14060013p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060014p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060015p-T	-	-	-	-	-	-	-	-	-	-
	10ng/ml	LSD14060013p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060014p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060015p-T	-	-	-	-	-	-	-	-	-	-

	30ng/ml	LSD14060013p-T	+	+	+	+	+	+	+	+	+	+
		LSD14060014p-T	+	+	+	+	+	+	+	+	+	+
		LSD14060015p-T	+	+	+	+	+	+	+	+	+	+
LSD 20	0ng/ml	LSD14060001p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060002p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060003p-T	-	-	-	-	-	-	-	-	-	-
	10ng/ml	LSD14060001p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060002p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060003p-T	-	-	-	-	-	-	-	-	-	-
	30ng/ml	LSD14060001p-T	+	+	+	+	+	+	+	+	+	+
		LSD14060002p-T	+	+	+	+	+	+	+	+	+	+
		LSD14060003p-T	+	+	+	+	+	+	+	+	+	+
LSD 50	0ng/ml	LSD14060007p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060008p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060009p-T	-	-	-	-	-	-	-	-	-	-
	25ng/ml	LSD14060007p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060008p-T	-	-	-	-	-	-	-	-	-	-
		LSD14060009p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	LSD14060007p-T	+	+	+	+	+	+	+	+	+	+
		LSD14060008p-T	+	+	+	+	+	+	+	+	+	+
		LSD14060009p-T	+	+	+	+	+	+	+	+	+	+
MDPV 1000	0ng/ml	MDPV14080001p-T	-	-	-	-	-	-	-	-	-	-
		MDPV14080002p-T	-	-	-	-	-	-	-	-	-	-
		MDPV14080003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	MDPV14080001p-T	-	-	-	-	-	-	-	-	-	-
		MDPV14080002p-T	-	-	-	-	-	-	-	-	-	-
		MDPV14080003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	MDPV14080001p-T	+	+	+	+	+	+	+	+	+	+
		MDPV14080002p-T	+	+	+	+	+	+	+	+	+	+
		MDPV14080003p-T	+	+	+	+	+	+	+	+	+	+
MDPV 500	0ng/ml	MDPV14080007p-T	-	-	-	-	-	-	-	-	-	-
		MDPV14080008p-T	-	-	-	-	-	-	-	-	-	-
		MDPV14080009p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	MDPV14080007p-T	-	-	-	-	-	-	-	-	-	-

		MDPV14080008p-T	-	-	-	-	-	-	-	-	-	-
		MDPV14080009p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	MDPV14080007p-T	+	+	+	+	+	+	+	+	+	+
		MDPV14080008p-T	+	+	+	+	+	+	+	+	+	+
		MDPV14080009p-T	+	+	+	+	+	+	+	+	+	+
MDPV 300	0ng/ml	MDPV15090001p-T	-	-	-	-	-	-	-	-	-	-
		MDPV15090002p-T	-	-	-	-	-	-	-	-	-	-
		MDPV15090003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MDPV15090001p-T	-	-	-	-	-	-	-	-	-	-
		MDPV15090002p-T	-	-	-	-	-	-	-	-	-	-
		MDPV15090003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MDPV15090001p-T	+	+	+	+	+	+	+	+	+	+
		MDPV15090002p-T	+	+	+	+	+	+	+	+	+	+
		MDPV15090003p-T	+	+	+	+	+	+	+	+	+	+
MPD 300	0ng/ml	MPD14080001p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080002p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MPD14080001p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080002p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MPD14080001p-T	+	+	+	+	+	+	+	+	+	+
		MPD14080002p-T	+	+	+	+	+	+	+	+	+	+
		MPD14080003p-T	+	+	+	+	+	+	+	+	+	+
MPD 1000	0ng/ml	MPD14080007p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080008p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080009p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	MPD14080007p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080008p-T	-	-	-	-	-	-	-	-	-	-
		MPD14080009p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	MPD14080007p-T	+	+	+	+	+	+	+	+	+	+
		MPD14080008p-T	+	+	+	+	+	+	+	+	+	+
		MPD14080009p-T	+	+	+	+	+	+	+	+	+	+
ZOL 50	0ng/ml	ZOL15010001p-T	-	-	-	-	-	-	-	-	-	-
		ZOL15010002p-T	-	-	-	-	-	-	-	-	-	-
		ZOL15010003p-T	-	-	-	-	-	-	-	-	-	-
	25ng/ml	ZOL15010001p-T	-	-	-	-	-	-	-	-	-	-
		ZOL15010002p-T	-	-	-	-	-	-	-	-	-	-

		ZOL15010003p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	ZOL15010001p-T	+	+	+	+	+	+	+	+	+	+
		ZOL15010002p-T	+	+	+	+	+	+	+	+	+	+
		ZOL15010003p-T	+	+	+	+	+	+	+	+	+	+
ZOP 50	0ng/ml	ZOP16080001p-T	-	-	-	-	-	-	-	-	-	-
		ZOP16080002p-T	-	-	-	-	-	-	-	-	-	-
		ZOP16080003p-T	-	-	-	-	-	-	-	-	-	-
	25ng/ml	ZOP16080001p-T	-	-	-	-	-	-	-	-	-	-
		ZOP16080002p-T	-	-	-	-	-	-	-	-	-	-
		ZOP16080003p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	ZOP16080001p-T	+	+	+	+	+	+	+	+	+	+
		ZOP16080002p-T	+	+	+	+	+	+	+	+	+	+
		ZOP16080003p-T	+	+	+	+	+	+	+	+	+	+
MCAT 500	0ng/ml	MCAT16080001p-T	-	-	-	-	-	-	-	-	-	-
		MCAT16080002p-T	-	-	-	-	-	-	-	-	-	-
		MCAT16080003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	MCAT16080001p-T	-	-	-	-	-	-	-	-	-	-
		MCAT16080002p-T	-	-	-	-	-	-	-	-	-	-
		MCAT16080003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	MCAT16080001p-T	+	+	+	+	+	+	+	+	+	+
		MCAT16080002p-T	+	+	+	+	+	+	+	+	+	+
		MCAT16080003p-T	+	+	+	+	+	+	+	+	+	+
7-ACL 300	0ng/ml	ACL14100001p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100002p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	ACL14100001p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100002p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	ACL14100001p-T	+	+	+	+	+	+	+	+	+	+
		ACL14100002p-T	+	+	+	+	+	+	+	+	+	+
		ACL14100003p-T	+	+	+	+	+	+	+	+	+	+
7-ACL 100	0ng/ml	ACL14100007p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100008p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100009p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	ACL14100007p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100008p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100009p-T	-	-	-	-	-	-	-	-	-	-

	150ng/ml	ACL14100007p-T	+	+	+	+	+	+	+	+	+	+
		ACL14100008p-T	+	+	+	+	+	+	+	+	+	+
		ACL14100009p-T	+	+	+	+	+	+	+	+	+	+
7-ACL 200	0ng/ml	ACL14100013p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100014p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100015p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	ACL14100013p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100014p-T	-	-	-	-	-	-	-	-	-	-
		ACL14100015p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	ACL14100013p-T	+	+	+	+	+	+	+	+	+	+
		ACL14100014p-T	+	+	+	+	+	+	+	+	+	+
		ACL14100015p-T	+	+	+	+	+	+	+	+	+	+
CFYL 500	0ng/ml	CFYL15020001p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020002p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	CFYL15020001p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020002p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	CFYL15020001p-T	+	+	+	+	+	+	+	+	+	+
		CFYL15020002p-T	+	+	+	+	+	+	+	+	+	+
		CFYL15020003p-T	+	+	+	+	+	+	+	+	+	+
CFYL 250	0ng/ml	CFYL15020007p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020008p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020009p-T	-	-	-	-	-	-	-	-	-	-
	125ng/ml	CFYL15020007p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020008p-T	-	-	-	-	-	-	-	-	-	-
		CFYL15020009p-T	-	-	-	-	-	-	-	-	-	-
	375ng/ml	CFYL15020007p-T	+	+	+	+	+	+	+	+	+	+
		CFYL15020008p-T	+	+	+	+	+	+	+	+	+	+
		CFYL15020009p-T	+	+	+	+	+	+	+	+	+	+
CAF1000	0ng/ml	CAF15080001p-T	-	-	-	-	-	-	-	-	-	-
		CAF15080002p-T	-	-	-	-	-	-	-	-	-	-
		CAF15080003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	CAF15080001p-T	-	-	-	-	-	-	-	-	-	-
		CAF15080002p-T	-	-	-	-	-	-	-	-	-	-
		CAF15080003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	CAF15080001p-T	+	+	+	+	+	+	+	+	+	+

		CAF15080002p-T	+	+	+	+	+	+	+	+	+	+
		CAF15080003p-T	+	+	+	+	+	+	+	+	+	+
CAT150	0ng/ml	CAT14080001p-T	-	-	-	-	-	-	-	-	-	-
		CAT14080002p-T	-	-	-	-	-	-	-	-	-	-
		CAT14080003p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	CAT14080001p-T	-	-	-	-	-	-	-	-	-	-
		CAT14080002p-T	-	-	-	-	-	-	-	-	-	-
		CAT14080003p-T	-	-	-	-	-	-	-	-	-	-
	225ng/ml	CAT14080001p-T	+	+	+	+	+	+	+	+	+	+
		CAT14080002p-T	+	+	+	+	+	+	+	+	+	+
		CAT14080003p-T	+	+	+	+	+	+	+	+	+	+
TRO350	0ng/ml	TRO15050001p-T	-	-	-	-	-	-	-	-	-	-
		TRO15050002p-T	-	-	-	-	-	-	-	-	-	-
		TRO15050003p-T	-	-	-	-	-	-	-	-	-	-
	175ng/ml	TRO15050001p-T	-	-	-	-	-	-	-	-	-	-
		TRO15050002p-T	-	-	-	-	-	-	-	-	-	-
		TRO15050003p-T	-	-	-	-	-	-	-	-	-	-
	525ng/ml	TRO15050001p-T	+	+	+	+	+	+	+	+	+	+
		TRO15050002p-T	+	+	+	+	+	+	+	+	+	+
		TRO15050003p-T	+	+	+	+	+	+	+	+	+	+
MEP 100	0ng/ml	MEP15070001p-T	-	-	-	-	-	-	-	-	-	-
		MEP15070002p-T	-	-	-	-	-	-	-	-	-	-
		MEP15070003p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	MEP15070001p-T	-	-	-	-	-	-	-	-	-	-
		MEP15070002p-T	-	-	-	-	-	-	-	-	-	-
		MEP15070003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MEP15070001p-T	+	+	+	+	+	+	+	+	+	+
		MEP15070002p-T	+	+	+	+	+	+	+	+	+	+
		MEP15070003p-T	+	+	+	+	+	+	+	+	+	+
ALP 100	0ng/ml	ALP16040001p-T	-	-	-	-	-	-	-	-	-	-
		ALP16040002p-T	-	-	-	-	-	-	-	-	-	-
		ALP16040003p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	ALP16040001p-T	-	-	-	-	-	-	-	-	-	-
		ALP16040002p-T	-	-	-	-	-	-	-	-	-	-
		ALP16040003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	ALP16040001p-T	+	+	+	+	+	+	+	+	+	+
		ALP16040002p-T	+	+	+	+	+	+	+	+	+	+

		ALP16040003p-T	+	+	+	+	+	+	+	+	+	+
ABP 10	0ng/ml	ABP17010001p-T	-	-	-	-	-	-	-	-	-	-
		ABP17010002p-T	-	-	-	-	-	-	-	-	-	-
		ABP17010003p-T	-	-	-	-	-	-	-	-	-	-
	5ng/ml	ABP17010001p-T	-	-	-	-	-	-	-	-	-	-
		ABP17010002p-T	-	-	-	-	-	-	-	-	-	-
		ABP17010003p-T	-	-	-	-	-	-	-	-	-	-
	15ng/ml	ABP17010001p-T	+	+	+	+	+	+	+	+	+	+
		ABP17010002p-T	+	+	+	+	+	+	+	+	+	+
		ABP17010003p-T	+	+	+	+	+	+	+	+	+	+
APVP 1000	0ng/ml	APVP15050001p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050002p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	APVP15050001p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050002p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	APVP15050001p-T	+	+	+	+	+	+	+	+	+	+
		APVP15050002p-T	+	+	+	+	+	+	+	+	+	+
		APVP15050003p-T	+	+	+	+	+	+	+	+	+	+
APVP 500	0ng/ml	APVP15050007p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050008p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050009p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	APVP15050007p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050008p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050009p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	APVP15050007p-T	+	+	+	+	+	+	+	+	+	+
		APVP15050008p-T	+	+	+	+	+	+	+	+	+	+
		APVP15050009p-T	+	+	+	+	+	+	+	+	+	+
APVP 300	0ng/ml	APVP15050013p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050014p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050015p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	APVP15050013p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050014p-T	-	-	-	-	-	-	-	-	-	-
		APVP15050015p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	APVP15050013p-T	+	+	+	+	+	+	+	+	+	+
		APVP15050014p-T	+	+	+	+	+	+	+	+	+	+
		APVP15050015p-T	+	+	+	+	+	+	+	+	+	+

CNB 500	0ng/ml	CNB18060001p-T	-	-	-	-	-	-	-	-	-	-
		CNB18060002p-T	-	-	-	-	-	-	-	-	-	-
		CNB18060003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	CNB18060001p-T	-	-	-	-	-	-	-	-	-	-
		CNB18060002p-T	-	-	-	-	-	-	-	-	-	-
		CNB18060003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	CNB18060001p-T	+	+	+	+	+	+	+	+	+	+
		CNB18060002p-T	+	+	+	+	+	+	+	+	+	+
		CNB18060003p-T	+	+	+	+	+	+	+	+	+	+
MPRB 100	0ng/ml	MPRD18060001p-T	-	-	-	-	-	-	-	-	-	-
		MPRD18060002p-T	-	-	-	-	-	-	-	-	-	-
		MPRD18060003p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	MPRD18060001p-T	-	-	-	-	-	-	-	-	-	-
		MPRD18060002p-T	-	-	-	-	-	-	-	-	-	-
		MPRD18060003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MPRD18060001p-T	+	+	+	+	+	+	+	+	+	+
		MPRD18060002p-T	+	+	+	+	+	+	+	+	+	+
		MPRD18060003p-T	+	+	+	+	+	+	+	+	+	+
PGB 50,000	0ng/ml	PGB16010001p-T	-	-	-	-	-	-	-	-	-	-
		PGB16010002p-T	-	-	-	-	-	-	-	-	-	-
		PGB16010003p-T	-	-	-	-	-	-	-	-	-	-
	25,000ng/ml	PGB16010001p-T	-	-	-	-	-	-	-	-	-	-
		PGB16010002p-T	-	-	-	-	-	-	-	-	-	-
		PGB16010003p-T	-	-	-	-	-	-	-	-	-	-
	75,000ng/ml	PGB16010001p-T	+	+	+	+	+	+	+	+	+	+
		PGB16010002p-T	+	+	+	+	+	+	+	+	+	+
		PGB16010003p-T	+	+	+	+	+	+	+	+	+	+
PGB 500	0ng/ml	PGB19040001p-T	-	-	-	-	-	-	-	-	-	-
		PGB19040002p-T	-	-	-	-	-	-	-	-	-	-
		PGB19040003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	PGB19040001p-T	-	-	-	-	-	-	-	-	-	-
		PGB19040002p-T	-	-	-	-	-	-	-	-	-	-
		PGB19040003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	PGB19040001p-T	+	+	+	+	+	+	+	+	+	+
		PGB19040002p-T	+	+	+	+	+	+	+	+	+	+
		PGB19040003p-T	+	+	+	+	+	+	+	+	+	+
TZD 200	0ng/ml	TZD18070001p-T	-	-	-	-	-	-	-	-	-	-

		TZD18070002p-T	-	-	-	-	-	-	-	-	-	-
		TZD18070003p-T	-	-	-	-	-	-	-	-	-	-
	100ng/ml	TZD18070001p-T	-	-	-	-	-	-	-	-	-	-
		TZD18070002p-T	-	-	-	-	-	-	-	-	-	-
		TZD18070003p-T	-	-	-	-	-	-	-	-	-	-
	300ng/ml	TZD18070001p-T	+	+	+	+	+	+	+	+	+	+
		TZD18070002p-T	+	+	+	+	+	+	+	+	+	+
		TZD18070003p-T	+	+	+	+	+	+	+	+	+	+
UR-144 25	0ng/ml	UR18060001p-T	-	-	-	-	-	-	-	-	-	-
		UR18060002p-T	-	-	-	-	-	-	-	-	-	-
		UR18060003p-T	-	-	-	-	-	-	-	-	-	-
	12.5ng/ml	UR18060001p-T	-	-	-	-	-	-	-	-	-	-
		UR18060002p-T	-	-	-	-	-	-	-	-	-	-
		UR18060003p-T	-	-	-	-	-	-	-	-	-	-
	37.5ng/ml	UR18060001p-T	+	+	+	+	+	+	+	+	+	+
		UR18060002p-T	+	+	+	+	+	+	+	+	+	+
		UR18060003p-T	+	+	+	+	+	+	+	+	+	+
ZAL 100	0ng/ml	ZAL18050001p-T	-	-	-	-	-	-	-	-	-	-
		ZAL18050002p-T	-	-	-	-	-	-	-	-	-	-
		ZAL18050003p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	ZAL18050001p-T	-	-	-	-	-	-	-	-	-	-
		ZAL18050002p-T	-	-	-	-	-	-	-	-	-	-
		ZAL18050003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	ZAL18050001p-T	+	+	+	+	+	+	+	+	+	+
		ZAL18050002p-T	+	+	+	+	+	+	+	+	+	+
		ZAL18050003p-T	+	+	+	+	+	+	+	+	+	+
MES 100	0ng/ml	MES18060001p-T	-	-	-	-	-	-	-	-	-	-
		MES18060002p-T	-	-	-	-	-	-	-	-	-	-
		MES18060003p-T	-	-	-	-	-	-	-	-	-	-
	50ng/ml	MES18060001p-T	-	-	-	-	-	-	-	-	-	-
		MES18060002p-T	-	-	-	-	-	-	-	-	-	-
		MES18060003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MES18060001p-T	+	+	+	+	+	+	+	+	+	+
		MES18060002p-T	+	+	+	+	+	+	+	+	+	+
		MES18060003p-T	+	+	+	+	+	+	+	+	+	+
MES 300	0ng/ml	MES18090001p-T	-	-	-	-	-	-	-	-	-	-
		MES18090002p-T	-	-	-	-	-	-	-	-	-	-

		MES18090003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	MES18090001p-T	-	-	-	-	-	-	-	-	-	-
		MES18090002p-T	-	-	-	-	-	-	-	-	-	-
		MES18090003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	MES18090001p-T	+	+	+	+	+	+	+	+	+	+
		MES18090002p-T	+	+	+	+	+	+	+	+	+	+
		MES18090003p-T	+	+	+	+	+	+	+	+	+	+
GAB 2000	0ng/ml	GAB18070001p-T	-	-	-	-	-	-	-	-	-	-
		GAB18070002p-T	-	-	-	-	-	-	-	-	-	-
		GAB18070003p-T	-	-	-	-	-	-	-	-	-	-
	1000ng/ml	GAB18070001p-T	-	-	-	-	-	-	-	-	-	-
		GAB18070002p-T	-	-	-	-	-	-	-	-	-	-
		GAB18070003p-T	-	-	-	-	-	-	-	-	-	-
	3000ng/ml	GAB18070001p-T	+	+	+	+	+	+	+	+	+	+
		GAB18070002p-T	+	+	+	+	+	+	+	+	+	+
		GAB18070003p-T	+	+	+	+	+	+	+	+	+	+
TLD 50	0ng/ml	TLD18110001p-T	-	-	-	-	-	-	-	-	-	-
		TLD18110002p-T	-	-	-	-	-	-	-	-	-	-
		TLD18110003p-T	-	-	-	-	-	-	-	-	-	-
	25 ng/ml	TLD18110001p-T	-	-	-	-	-	-	-	-	-	-
		TLD18110002p-T	-	-	-	-	-	-	-	-	-	-
		TLD18110003p-T	-	-	-	-	-	-	-	-	-	-
	75 ng/ml	TLD18110001p-T	+	+	+	+	+	+	+	+	+	+
		TLD18110002p-T	+	+	+	+	+	+	+	+	+	+
		TLD18110003p-T	+	+	+	+	+	+	+	+	+	+
QTP 1000	0ng/ml	QTP19030001p-T	-	-	-	-	-	-	-	-	-	-
		QTP19030002p-T	-	-	-	-	-	-	-	-	-	-
		QTP19030003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	QTP19030001p-T	-	-	-	-	-	-	-	-	-	-
		QTP19030002p-T	-	-	-	-	-	-	-	-	-	-
		QTP19030003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	QTP19030001p-T	+	+	+	+	+	+	+	+	+	+
		QTP19030002p-T	+	+	+	+	+	+	+	+	+	+
		QTP19030003p-T	+	+	+	+	+	+	+	+	+	+
PAP 500	0ng/ml	PAP18110001p-T	-	-	-	-	-	-	-	-	-	-
		PAP18110002p-T	-	-	-	-	-	-	-	-	-	-
		PAP18110003p-T	-	-	-	-	-	-	-	-	-	-

	250ng/ml	PAP18110001p-T	-	-	-	-	-	-	-	-	-	-
		PAP18110002p-T	-	-	-	-	-	-	-	-	-	-
		PAP18110003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	PAP18110001p-T	+	+	+	+	+	+	+	+	+	+
		PAP18110002p-T	+	+	+	+	+	+	+	+	+	+
		PAP18110003p-T	+	+	+	+	+	+	+	+	+	+
KRA 300	0ng/ml	KRA19040001p-T	-	-	-	-	-	-	-	-	-	-
		KRA19040002p-T	-	-	-	-	-	-	-	-	-	-
		KRA19040003p-T	-	-	-	-	-	-	-	-	-	-
	150ng/ml	KRA19040001p-T	-	-	-	-	-	-	-	-	-	-
		KRA19040002p-T	-	-	-	-	-	-	-	-	-	-
		KRA19040003p-T	-	-	-	-	-	-	-	-	-	-
	450ng/ml	KRA19040001p-T	+	+	+	+	+	+	+	+	+	+
		KRA19040002p-T	+	+	+	+	+	+	+	+	+	+
		KRA19040003p-T	+	+	+	+	+	+	+	+	+	+
CAR 2000	0ng/ml	CAR18100001p-T	-	-	-	-	-	-	-	-	-	-
		CAR18100002p-T	-	-	-	-	-	-	-	-	-	-
		CAR18100003p-T	-	-	-	-	-	-	-	-	-	-
	1000ng/ml	CAR18100001p-T	-	-	-	-	-	-	-	-	-	-
		CAR18100002p-T	-	-	-	-	-	-	-	-	-	-
		CAR18100003p-T	-	-	-	-	-	-	-	-	-	-
	3000ng/ml	CAR18100001p-T	+	+	+	+	+	+	+	+	+	+
		CAR18100002p-T	+	+	+	+	+	+	+	+	+	+
		CAR18100003p-T	+	+	+	+	+	+	+	+	+	+
FLX 500	0ng/ml	FLX18110001p-T	-	-	-	-	-	-	-	-	-	-
		FLX18110002p-T	-	-	-	-	-	-	-	-	-	-
		FLX18110003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	FLX18110001p-T	-	-	-	-	-	-	-	-	-	-
		FLX18110002p-T	-	-	-	-	-	-	-	-	-	-
		FLX18110003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	FLX18110001p-T	+	+	+	+	+	+	+	+	+	+
		FLX18110002p-T	+	+	+	+	+	+	+	+	+	+
		FLX18110003p-T	+	+	+	+	+	+	+	+	+	+
RPD 150	0ng/ml	RPD20090001p-T	-	-	-	-	-	-	-	-	-	-
		RPD20090002p-T	-	-	-	-	-	-	-	-	-	-
		RPD20090003p-T	-	-	-	-	-	-	-	-	-	-
	75ng/ml	RPD20090001p-T	-	-	-	-	-	-	-	-	-	-

		RPD20090002p-T	-	-	-	-	-	-	-	-	-	-
		RPD20090003p-T	-	-	-	-	-	-	-	-	-	-
	225ng/ml	RPD20090001p-T	+	+	+	+	+	+	+	+	+	+
		RPD20090002p-T	+	+	+	+	+	+	+	+	+	+
		RPD20090003p-T	+	+	+	+	+	+	+	+	+	+
CIT 500	0ng/ml	CIT20120001p-T	-	-	-	-	-	-	-	-	-	-
		CIT20120002p-T	-	-	-	-	-	-	-	-	-	-
		CIT20120003p-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	CIT20120001p-T	-	-	-	-	-	-	-	-	-	-
		CIT20120002p-T	-	-	-	-	-	-	-	-	-	-
		CIT20120003p-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	CIT20120001p-T	+	+	+	+	+	+	+	+	+	+
		CIT20120002p-T	+	+	+	+	+	+	+	+	+	+
		CIT20120003p-T	+	+	+	+	+	+	+	+	+	+
FKET 1000	0ng/ml	FKE21010001p-T	-	-	-	-	-	-	-	-	-	-
		FKE21010002p-T	-	-	-	-	-	-	-	-	-	-
		FKE21010003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	FKE21010001p-T	-	-	-	-	-	-	-	-	-	-
		FKE21010002p-T	-	-	-	-	-	-	-	-	-	-
		FKE21010003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	FKE21010001p-T	+	+	+	+	+	+	+	+	+	+
		FKE21010002p-T	+	+	+	+	+	+	+	+	+	+
		FKE21010003p-T	+	+	+	+	+	+	+	+	+	+
TAP 1000	0ng/ml	TAP21090001p-T	-	-	-	-	-	-	-	-	-	-
		TAP21090002p-T	-	-	-	-	-	-	-	-	-	-
		TAP21090003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	TAP21090001p-T	-	-	-	-	-	-	-	-	-	-
		TAP21090002p-T	-	-	-	-	-	-	-	-	-	-
		TAP21090003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	TAP21090001p-T	+	+	+	+	+	+	+	+	+	+
		TAP21090002p-T	+	+	+	+	+	+	+	+	+	+
		TAP21090003p-T	+	+	+	+	+	+	+	+	+	+
SCOP 500	0ng/ml	SCOP21100004-T	-	-	-	-	-	-	-	-	-	-
		SCOP21100005-T	-	-	-	-	-	-	-	-	-	-
		SCOP21100006-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	SCOP21100004-T	-	-	-	-	-	-	-	-	-	-
		SCOP21100005-T	-	-	-	-	-	-	-	-	-	-

		SCOP21100006-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	SCOP21100004-T	+	+	+	+	+	+	+	+	+	+
		SCOP21100005-T	+	+	+	+	+	+	+	+	+	+
		SCOP21100006-T	+	+	+	+	+	+	+	+	+	+
MTZ 500	0ng/ml	MTZ21110007-T	-	-	-	-	-	-	-	-	-	-
		MTZ21110008-T	-	-	-	-	-	-	-	-	-	-
		MTZ21110009-T	-	-	-	-	-	-	-	-	-	-
	250ng/ml	MTZ21110007-T	-	-	-	-	-	-	-	-	-	-
		MTZ21110008-T	-	-	-	-	-	-	-	-	-	-
		MTZ21110009-T	-	-	-	-	-	-	-	-	-	-
	750ng/ml	MTZ21110007-T	+	+	+	+	+	+	+	+	+	+
		MTZ21110008-T	+	+	+	+	+	+	+	+	+	+
		MTZ21110009-T	+	+	+	+	+	+	+	+	+	+
NND 1000	0ng/ml	NND21120001p-T	-	-	-	-	-	-	-	-	-	-
		NND21120002p-T	-	-	-	-	-	-	-	-	-	-
		NND21120003p-T	-	-	-	-	-	-	-	-	-	-
	500ng/ml	NND21120001p-T	-	-	-	-	-	-	-	-	-	-
		NND21120002p-T	-	-	-	-	-	-	-	-	-	-
		NND21120003p-T	-	-	-	-	-	-	-	-	-	-
	1500ng/ml	NND21120001p-T	+	+	+	+	+	+	+	+	+	+
		NND21120002p-T	+	+	+	+	+	+	+	+	+	+
		NND21120003p-T	+	+	+	+	+	+	+	+	+	+

Conclusion: Test results were consistent among the three lots of each test in Multi-Drug Rapid Test Panel.

2.7 Accelerated Stability

Accelerated Stability of The Multi-Drug Rapid Test Panel was evaluated using samples from three different batches; each batch will assembled in 10 in 1 panel. These were placed in an incubator with the temperature calibrated at 45℃ and 55℃, and relative humidity (RH) calibrated at 60%. A series of stability tests were performed at 0, 7, 14, 21, 28, 35, 42, 56, 77, 84 days. Tests were assayed using each testconcentration at drug free urine, -50% Cut-off and +50% Cut-off urine specimens. Testing at each specific time interval consisted of triplicates for each specimen. The tests were performed according to the package insert. Results are presented in tables below.

Table: Timeline for Accelerate Stability Study

Day / Temp.	0 day	7 days	14 days	21 days	28 days	35 days	42 days	56 days	77 days	84 days
45℃	√	√	√	√	√	√	√	√	√	√
55℃	√	√	√	√	√	√	√	X	X	X

Table: Accelerated Stability Summary

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			AMP14030001p-T	AMP14030002p-T	AMP14030003p-T
AMP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			AMP14030001p-T	AMP14030002p-T	AMP14030003p-T
AMP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			AMP14030007p-T	AMP14030008p-T	AMP14030009p-T
AMP 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			AMP14030007p-T	AMP14030008p-T	AMP14030009p-T
AMP 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			AMP14030013p-T	AMP14030014p-T	AMP14030015p-T
AMP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			AMP14030013p-T	AMP14030014p-T	AMP14030015p-T
AMP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			BAR14030001p-T	BAR14030002p-T	BAR14030003p-T
BAR 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			BAR14030001p-T	BAR14030002p-T	BAR14030003p-T
BAR 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			BAR14030007p-T	BAR14030008p-T	BAR14030009p-T
BAR 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			BAR14030007p-T	BAR14030008p-T	BAR14030009p-T
BAR 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			BUP14030001p-T	BUP14030002p-T	BUP14030003p-T
BUP 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			BUP14030001p-T	BUP14030002p-T	BUP14030003p-T
BUP 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃			
Product	**Day**	**Specimen**	**Batch No.**

			BUP14030007p-T	BUP14030008p-T	BUP14030009p-T
BUP 5	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			BUP14030007p-T	BUP14030008p-T	BUP14030009p-T
BUP 5	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			BZO14030019p-T	BZO14030020p-T	BZO14030021p-T
BZO 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			BZO14030019p-T	BZO14030020p-T	BZO14030021p-T
BZO 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			BZO14030001p-T	BZO14030002p-T	BZO14030003p-T
BZO 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			BZO14030001p-T	BZO14030002p-T	BZO14030003p-T
BZO 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			BZO14030007p-T	BZO14030008p-T	BZO14030009p-T

BZO 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			BZO14030007p-T	BZO14030008p-T	BZO14030009p-T
BZO 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			BZO14030013p-T	BZO14030014p-T	BZO14030015p-T
BZO 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					

Product	Day	Specimen	Batch No.		
			BZO14030013p-T	BZO14030014p-T	BZO14030015p-T
BZO 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COC14030001p-T	COC14030002p-T	COC14030003p-T
COC 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			COC14030001p-T	COC14030002p-T	COC14030003p-T
COC 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			COC14030013p-T	COC14030014p-T	COC14030015p-T
COC 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COC14030013p-T	COC14030014p-T	COC14030015p-T
COC 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COC14030007p-T	COC14030008p-T	COC14030009p-T
COC 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			COC14030007p-T	COC14030008p-T	COC14030009p-T
COC 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COC14030019p-T	COC14030020p-T	COC14030021p-T
COC 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			COC14030019p-T	COC14030020p-T	COC14030021p-T
COC 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			KET14030001p-T	KET14030002p-T	KET14030003p-T
KET 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			KET14030001p-T	KET14030002p-T	KET14030003p-T
KET 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			KET14030007p-T	KET14030008p-T	KET14030009p-T
KET 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			KET14030007p-T	KET14030008p-T	KET14030009p-T
KET 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			KET14030013p-T	KET14030014p-T	KET14030015p-T
KET 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			KET14030013p-T	KET14030014p-T	KET14030015p-T
KET 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			KET14030019p-T	KET14030020p-T	KET14030021p-T
KET 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			KET14030019p-T	KET14030020p-T	KET14030021p-T
KET 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDMA14030007p-T	MDMA14030008p-T	MDMA14030009p-T
MDMA 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			MDMA14030007p-T	MDMA14030008p-T	MDMA14030009p-T
MDMA 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDMA14030001p-T	MDMA14030002p-T	MDMA14030003p-T
MDMA 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MDMA14030001p-T	MDMA14030002p-T	MDMA14030003p-T
MDMA 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			MDMA14030013p-T	MDMA14030014p-T	MDMA14030015p-T
MDMA 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MDMA14030013p-T	MDMA14030014p-T	MDMA14030015p-T
MDMA 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MET14030001p-T	MET14030002p-T	MET14030003p-T
MET 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			MET14030001p-T	MET14030002p-T	MET14030003p-T
MET 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MET14030007p-T	MET14030008p-T	MET14030009p-T
MET 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MET14030007p-T	MET14030008p-T	MET14030009p-T
MET 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MET14030013p-T	MET14030014p-T	MET14030015p-T
MET 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MET14030013p-T	MET14030014p-T	MET14030015p-T
MET 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MOP14030001p-T	MOP14030002p-T	MOP14030003p-T
MOP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			MOP14030001p-T	MOP14030002p-T	MOP14030003p-T
MOP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MOP14030013p-T	MOP14030014p-T	MOP14030015p-T
MOP 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MOP14030013p-T	MOP14030014p-T	MOP14030015p-T
MOP 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MOP14030007p-T	MOP14030008p-T	MOP14030009p-T
MOP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MOP14030007p-T	MOP14030008p-T	MOP14030009p-T
MOP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MQL14030001p-T	MQL14030002p-T	MQL14030003p-T
MQL 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			MQL14030001p-T	MQL14030002p-T	MQL14030003p-T
MQL 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MTD14030001p-T	MTD14030002p-T	MTD14030003p-T
MTD 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MTD14030001p-T	MTD14030002p-T	MTD14030003p-T
MTD 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			MTD14030007p-T	MTD14030008p-T	MTD14030009p-T
MTD 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MTD14030007p-T	MTD14030008p-T	MTD14030009p-T
MTD 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			OPI14030001p-T	OPI14030002p-T	OPI14030003p-T
OPI 2000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			OPI14030001p-T	OPI14030002p-T	OPI14030003p-T
OPI	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

2000		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			PCP14030001p-T	PCP14030002p-T	PCP14030003p-T
PCP 25	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			PCP14030001p-T	PCP14030002p-T	PCP14030003p-T
PCP 25	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			PPX14030001p-T	PPX14030002p-T	PPX14030003p-T
PPX 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			PPX14030001p-T	PPX14030002p-T	PPX14030003p-T
PPX 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030001p-T	THC14030002p-T	THC14030003p-T
THC 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030001p-T	THC14030002p-T	THC14030003p-T
THC 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030007p-T	THC14030008p-T	THC14030009p-T
THC 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030007p-T	THC14030008p-T	THC14030009p-T
THC 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030013p-T	THC14030014p-T	THC14030015p-T
THC 25	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			THC14030013p-T	THC14030014p-T	THC14030015p-T
THC 25	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030019p-T	THC14030020p-T	THC14030021p-T
THC 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030019p-T	THC14030020p-T	THC14030021p-T
THC 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030037p-T	THC14030038p-T	THC14030039p-T
THC 20	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030037p-T	THC14030038p-T	THC14030039p-T
THC 20	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030025p-T	THC14030026p-T	THC14030027p-T
THC 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			THC14030025p-T	THC14030026p-T	THC14030027p-T
THC 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030031p-T	THC14030032p-T	THC14030033p-T
THC 30	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			THC14030031p-T	THC14030032p-T	THC14030033p-T
THC 30	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TCA14030001p-T	TCA14030002p-T	TCA14030003p-T
TCA 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TCA14030001p-T	TCA14030002p-T	TCA14030003p-T
TCA 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TCA14030007p-T	TCA14030008p-T	TCA14030009p-T
TCA 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			TCA14030007p-T	TCA14030008p-T	TCA14030009p-T
TCA 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃			
Product	**Day**	**Specimen**	**Batch No.**

			TML14030001p-T	TML14030002p-T	TML14030003p-T
TML 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			TML14030001p-T	TML14030002p-T	TML14030003p-T
TML 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TML14030007p-T	TML14030008p-T	TML14030009p-T
TML 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TML14030007p-T	TML14030008p-T	TML14030009p-T
TML 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TML14030013p-T	TML14030014p-T	TML14030015p-T
TML 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			TML14030013p-T	TML14030014p-T	TML14030015p-T
TML 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			OXY14030001p-T	OXY14030002p-T	OXY14030003p-T
OXY100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			OXY14030001p-T	OXY14030002p-T	OXY14030003p-T
OXY100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			OXY14030007p-T	OXY14030008p-T	OXY14030009p-T
OXY300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			OXY14030007p-T	OXY14030008p-T	OXY14030009p-T
OXY300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COT14030001p-T	COT14030002p-T	COT14030003p-T
COT 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			COT14030001p-T	COT14030002p-T	COT14030003p-T
COT 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			COT14030007p-T	COT14030008p-T	COT14030009p-T
COT 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			COT14030007p-T	COT14030008p-T	COT14030009p-T
COT 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COT14030013p-T	COT14030014p-T	COT14030015p-T
COT 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COT14030013p-T	COT14030014p-T	COT14030015p-T
COT 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			COT14030019p-T	COT14030020p-T	COT14030021p-T
COT 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			COT14030019p-T	COT14030020p-T	COT14030021p-T
COT 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			COT14030025p-T	COT14030026p-T	COT14030027p-T
COT 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			COT14030025p-T	COT14030026p-T	COT14030027p-T
COT 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			EDDP14030001p-T	EDDP14030002p-T	EDDP14030003p-T
EDDP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			EDDP14030001p-T	EDDP14030002p-T	EDDP14030003p-T
EDDP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			EDDP14030007p-T	EDDP14030008p-T	EDDP14030009p-T
EDDP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			EDDP14030007p-T	EDDP14030008p-T	EDDP14030009p-T
EDDP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040001p-T	FYL18040002p-T	FYL18040003p-T
FYL 20	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040001p-T	FYL18040002p-T	FYL18040003p-T
FYL 20	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040007p-T	FYL18040008p-T	FYL18040009p-T
FYL 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040007p-T	FYL18040008p-T	FYL18040009p-T
FYL 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040013p-T	FYL18040014p-T	FYL18040015p-T
FYL 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040013p-T	FYL18040014p-T	FYL18040015p-T
FYL 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040019p-T	FYL18040020p-T	FYL18040021p-T
FYL 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			FYL18040019p-T	FYL18040020p-T	FYL18040021p-T
FYL 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			FYL18040025p-T	FYL18040026p-T	FYL18040027p-T
FYL 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			FYL18040025p-T	FYL18040026p-T	FYL18040027p-T
FYL 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			SMA14030001p-T	SMA14030002p-T	SMA14030003p-T
K2 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			SMA14030001p-T	SMA14030002p-T	SMA14030003p-T
K2 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			SMA14030007p-T	SMA14030008p-T	SMA14030009p-T
K2 30	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			SMA14030007p-T	SMA14030008p-T	SMA14030009p-T
K2 30	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			SMA14030013p-T	SMA14030014p-T	SMA14030015p-T
K2 25	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			SMA14030013p-T	SMA14030014p-T	SMA14030015p-T
K2 25	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MAM14030001p-T	MAM14030002p-T	MAM14030003p-T
6-MAM 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MAM14030001p-T	MAM14030002p-T	MAM14030003p-T
6-MAM 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			DIA14030001p-T	DIA14030002p-T	DIA14030003p-T
DIA300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			DIA14030001p-T	DIA14030002p-T	DIA14030003p-T
DIA300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			DIA14030007p-T	DIA14030008p-T	DIA14030009p-T
DIA 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			DIA14030007p-T	DIA14030008p-T	DIA14030009p-T
DIA 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**

			CLO14030001p-T	CLO14030002p-T	CLO14030003p-T
CLO 400	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			CLO14030001p-T	CLO14030002p-T	CLO14030003p-T
CLO 400	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CLO14030007p-T	CLO14030008p-T	CLO14030009p-T
CLO 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CLO14030007p-T	CLO14030008p-T	CLO14030009p-T
CLO 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDA14030001p-T	MDA14030002p-T	MDA14030003p-T
MDA 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MDA14030001p-T	MDA14030002p-T	MDA14030003p-T
MDA 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			ETG14030013p-T	ETG14030014p-T	ETG14030015p-T

ETG300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			ETG14030013p-T	ETG14030014p-T	ETG14030015p-T
ETG300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ETG14030001p-T	ETG14030002p-T	ETG14030003p-T
ETG500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					

Product	Day	Specimen	Batch No.		
			ETG14030001p-T	ETG14030002p-T	ETG14030003p-T
ETG500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ETG14030007p-T	ETG14030008p-T	ETG14030009p-T
ETG1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			ETG14030007p-T	ETG14030008p-T	ETG14030009p-T
ETG1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			ACE14060001p-T	ACE14060002p-T	ACE14060003p-T
ACE 5000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ACE14060001p-T	ACE14060002p-T	ACE14060003p-T
ACE 5000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			LSD14060013p-T	LSD14060014p-T	LSD14060015p-T
LSD10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			LSD14060013p-T	LSD14060014p-T	LSD14060015p-T
LSD10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			LSD14060001p-T	LSD14060002p-T	LSD14060003p-T
LSD20	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			LSD14060001p-T	LSD14060002p-T	LSD14060003p-T
LSD20	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			LSD14060007p-T	LSD14060008p-T	LSD14060009p-T
LSD50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			LSD14060007p-T	LSD14060008p-T	LSD14060009p-T
LSD50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDPV14080001 p-T	MDPV14080002 p-T	MDPV14080003p -T
MDPV 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDPV14080001 p-T	MDPV14080002 p-T	MDPV14080003p -T
MDPV 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDPV14080007 p-T	MDPV14080008 p-T	MDPV14080009p -T
MDPV 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDPV14080007 p-T	MDPV14080008 p-T	MDPV14080009p -T
MDPV500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MDPV15090001 p-T	MDPV15090002 p-T	MDPV15090003p -T
MDPV 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MDPV15090001p-T	MDPV15090002p-T	MDPV15090003p-T
MDPV300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MPD14060001p-T	MPD14060002p-T	MPD14060003p-T
MPD300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MPD14060001p-T	MPD14060002p-T	MPD14060003p-T
MPD300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MPD14060007p-T	MPD14060008p-T	MPD14060009p-T
MPD1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MPD14060007p-T	MPD14060008p-T	MPD14060009p-T
MPD1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ZOL15010001p-T	ZOL15010002p-T	ZOL15010003p-T
ZOL 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			ZOL15010001p-T	ZOL15010002p-T	ZOL15010003p-T
ZOL 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃			
Product	**Day**	**Specimen**	**Batch No.**

			ZOP16080001p-T	ZOP16080002p-T	ZOP16080003p-T
ZOP 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			ZOP16080001p-T	ZOP16080002p-T	ZOP16080003p-T
ZOP 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MCAT16080001p-T	MCAT16080002p-T	MCAT16080003p-T
MCAT 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MCAT16080001p-T	MCAT16080002p-T	MCAT16080003p-T

MCAT 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ACL14100001p-T	**ACL14100002p-T**	**ACL14100003p-T**
7-ACL 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			ACL14100001p-T	ACL14100002p-T	ACL14100003p-T
7-ACL 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+, +	+,+,+,+,+,+,+,+,+,+, +	+,+,+,+,+,+,+,+,+,+, +
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			ACL14100007p-T	ACL14100008p-T	ACL14100009p-T
7-ACL 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+, +	+,+,+,+,+,+,+,+,+,+, +	+,+,+,+,+,+,+,+,+,+, +
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			ACL14100007p-T	**ACL14100008p-T**	**ACL14100009p-T**
7-ACL 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ACL14100013p-T	**ACL14100014p-T**	**ACL14100015p-T**
7-ACL 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ACL14100013p-T	**ACL14100014p-T**	**ACL14100015p-T**
7-ACL 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CFYL15020001p-T	**CFYL15020002p-T**	**CFYL15020003p-T**
CFYL 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CFYL15020001p-T	**CFYL15020002p-T**	**CFYL15020003p-T**
CFYL 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CFYL15020007p-T	**CFYL15020008p-T**	**CFYL15020009p-T**
CFYL 250	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			CFYL15020007p-T	CFYL15020008p-T	CFYL15020009p-T
CFYL 250	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			CAF15080001p-T	CAF15080002p-T	CAF15080003p-T
CAF 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+

	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			CAF15080001p-T	**CAF15080002p-T**	**CAF15080003p-T**
CAF 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CAT14080001p-T	**CAT14080002p-T**	**CAT14080003p-T**
CAT 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CAT14080001p-T	**CAT14080002p-T**	**CAT14080003p-T**
CAT 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TRO15050001p-T	TRO15050002p-T	TRO15050003p-T
TRO 350	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			TRO15050001p-T	TRO15050002p-T	TRO15050003p-T
TRO 350	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			MEP15070001p-T	**MEP15070002p-T**	**MEP15070003p-T**
MEP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MEP15070001p-T	**MEP15070002p-T**	**MEP15070003p-T**
MEP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**

			ALP16040001p-T	ALP16040002p-T	ALP16040003p-T
ALP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			ALP16040004-T	ALP16040005-T	ALP16040006-T
ALP 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ABP17010001p-T	**ABP17010002p-T**	**ABP17010003p-T**
ABP 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			ABP17010001p-T	ABP17010002p-T	ABP17010003p-T
ABP 10	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			APVP15050001p-T	**APVP15050002p-T**	**APVP15050003p-T**
APVP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			APVP15050001p-T	APVP15050002p-T	APVP15050003p-T
APVP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			APVP15050007p-T	APVP15050008p-T	APVP15050009p-T
APVP 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			APVP15050007p-T	**APVP15050008p-T**	**APVP15050009p-T**
APVP 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			APVP15050013p-T	**APVP15050014p-T**	**APVP15050015p-T**
APVP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			APVP15050013p-T	**APVP15050014p-T**	**APVP15050015p-T**
APVP 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CNB18060001p-T	**CNB18060002p-T**	**CNB18060003p-T**
CNB 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CNB18060001p-T	**CNB18060002p-T**	**CNB18060003p-T**
CNB 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MPRD18060001p-T	**MPRD18060002p-T**	**MPRD18060003p-T**
MPRD 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MPRD18060001p-T	**MPRD18060002p-T**	**MPRD18060003p-T**
MPRD 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			PGB16010001p-T	PGB16010002p-T	PGB16010003p-T

PGB 50,000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			PGB16010001p-T	PGB16010002p-T	PGB16010003p-T
PGB 50,000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			PGB19040001p-T	PGB19040002p-T	PGB19040003p-T
PGB 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		

			PGB19040001p-T	PGB19040002p-T	PGB19040003p-T
PGB 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TZD18070001p-T	TZD18070002p-T	TZD18070003p-T
TZD 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			TZD18070001p-T	TZD18070002p-T	TZD18070003p-T
TZD 200	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			UR18060001p-T	**UR18060002p-T**	**UR18060003p-T**
UR-14425	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			UR18060001p-T	UR18060002p-T	UR18060003p-T
UR-14425	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ZAL18050001p-T	**ZAL18050002p-T**	**ZAL18050003p-T**
ZAL 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			ZAL18050001p-T	**ZAL18050002p-T**	**ZAL18050003p-T**
ZAL 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MES18060004-T	**MES18060005-T**	**MES18060006-T**
MES 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			MES18060004-T	MES18060005-T	MES18060006-T
MES 100	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			MES18090004-T	MES18090005-T	MES18090006-T
MES 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			MES18090004-T	**MES18090005-T**	**MES18090006-T**
MES 300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	**Day**	**Specimen**	**Batch No.**		
			GAB18070001p-T	GAB18070002p-T	GAB18070003p-T
GAB 2000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			GAB18070001p-T	GAB18070002p-T	GAB18070003p-T
GAB 2000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TLD18110004-T	**TLD18110005-T**	**TLD18110006-T**
TLD 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TLD18110004-T	**TLD18110005-T**	**TLD18110006-T**
TLD 50	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			QTP19030004-T	**QTP19030005-T**	**QTP19030006-T**
QTP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			QTP19030004-T	QTP19030005-T	QTP19030006-T
QTP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			PAP18110001p-T	PAP18110002p-T	PAP18110003p-T
PAP500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			PAP18110001p-T	**PAP18110002p-T**	**PAP18110003p-T**
PAP500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			KRA19040001p-T	**KRA19040002p-T**	**KRA19040003p-T**
KRA300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			KRA19040001p-T	**KRA19040002p-T**	**KRA19040003p-T**
KRA300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CAR18100001p-T	**CAR18100002p-T**	**CAR18100003p-T**
CAR 2000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CAR18100001p-T	**CAR18100002p-T**	**CAR18100003p-T**
CAR 2000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃																				
Product	**Day**	**Specimen**	**Batch No.**																	
			FLX18110001p-T					**FLX18110002p-T**					**FLX18110003p-T**							
FLX 500	0	Negative	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-			
			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-			
		-50%cutoff	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-			
			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-			
		+50%cutoff	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+			
			+	+	+	+	+	+	+	+	+	+	+	+	+	+	+			
	7	Negative	-		-		-		-		-		-		-					
		-50%cutoff	-		-		-		-		-		-		-					
		+50%cutoff	+		+		+		+		+		+		+					
	14	Negative	-		-		-		-		-		-		-					
		-50%cutoff	-		-		-		-		-		-		-					
		+50%cutoff	+		+		+		+		+		+		+					
	21	Negative	-		-		-		-		-		-		-					
		-50%cutoff	-		-		-		-		-		-		-					
		+50%cutoff	+		+		+		+		+		+		+					
	28	Negative	-		-		-		-		-		-		-					
		-50%cutoff	-		-		-		-		-		-		-					
		+50%cutoff	+		+		+		+		+		+		+					
	35	Negative	-		-		-		-		-		-		-					
		-50%cutoff	-		-		-		-		-		-		-					
		+50%cutoff	+		+		+		+		+		+		+					

	42	Negative	-	-	-	-	-	-	-	-	-
		-50%cutoff	-	-	-	-	-	-	-	-	-
		+50%cutoff	+	+	+	+	+	+	+	+	+
	56	Negative	-	-	-	-	-	-	-	-	-
		-50%cutoff	-	-	-	-	-	-	-	-	-
		+50%cutoff	+	+	+	+	+	+	+	+	+
	77	Negative	-	-	-	-	-	-	-	-	-
		-50%cutoff	-	-	-	-	-	-	-	-	-
		+50%cutoff	+	+	+	+	+	+	+	+	+
	84	Negative	-	-	-	-	-	-	-	-	-
		-50%cutoff	-	-	-	-	-	-	-	-	-
		+50%cutoff	+	+	+	+	+	+	+	+	+

	Accelerated Stability Study Results at 55℃																			
	Day	**Specimen**	**Batch No.**																	
			FLX18110001p-T					**FLX18110002p-T**					**FLX18110003p-T**							
FLX 500	0	Negative	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
		-50%cutoff	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
		+50%cutoff	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+
			+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+
	7	Negative	-	-	-	-	-	-	-	-	-									
		-50%cutoff	-	-	-	-	-	-	-	-	-									
		+50%cutoff	+	+	+	+	+	+	+	+	+									
	14	Negative	-	-	-	-	-	-	-	-	-									
		-50%cutoff	-	-	-	-	-	-	-	-	-									
		+50%cutoff	+	+	+	+	+	+	+	+	+									
	21	Negative	-	-	-	-	-	-	-	-	-									
		-50%cutoff	-	-	-	-	-	-	-	-	-									
		+50%cutoff	+	+	+	+	+	+	+	+	+									
	28	Negative	-	-	-	-	-	-	-	-	-									
		-50%cutoff	-	-	-	-	-	-	-	-	-									
		+50%cutoff	+	+	+	+	+	+	+	+	+									
	35	Negative	-	-	-	-	-	-	-	-	-									
		-50%cutoff	-	-	-	-	-	-	-	-	-									
		+50%cutoff	+	+	+	+	+	+	+	+	+									
	42	Negative	-	-	-	-	-	-	-	-	-									
		-50%cutoff	-	-	-	-	-	-	-	-	-									
		+50%cutoff	+	+	+	+	+	+	+	+	+									

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			RPD20090001p-T	**RPD20090002p-T**	**RPD20090003p-T**
RPD 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	**Day**	**Specimen**	**Batch No.**		
			RPD20090001p-T	**RPD20090002p-T**	**RPD20090003p-T**
RPD 150	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CIT20120001p-T	**CIT20120002p-T**	**CIT20120003p-T**
CIT 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			CIT20120001p-T	**CIT20120002p-T**	**CIT20120003p-T**
CIT 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FKE21010001p-T	**FKE21010002p-T**	**FKE21010003p-T**
FKET 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			FKE21010001p-T	**FKE21010002p-T**	**FKE21010003p-T**
FKET 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			TAP21090001p-T	**TAP21090002p-T**	**TAP21090003p-T**
TAP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃

Product	Day	Specimen	Batch No.		
			TAP21090001p-T	TAP21090002p-T	TAP21090003p-T
TAP 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃

Product	Day	Specimen	Batch No.		
			SCOP21100004-T	SCOP21100005-T	SCOP21100006-T
SCOP 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			SCOP21100004-T	**SCOP21100005-T**	**SCOP21100006-T**
SCOP 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-

		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MTZ21110007-T	**MTZ21110008-T**	**MTZ21110009-T**
MTZ 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			MTZ21110007-T	**MTZ21110008-T**	**MTZ21110009-T**
MTZ 500	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-

		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Accelerated Stability Study Results at 45℃					
Product	**Day**	**Specimen**	**Batch No.**		
			NND21120001p-T	**NND21120002p-T**	**NND21120003p-T**
NND 1000	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	56	Drug free urine	-,-,-	-,-,-	-,-,-

		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	77	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	84	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
Accelerated Stability Study Results at 55℃					
Product	**Day**	**Specimen**	**Batch No.**		
			NND21120001p-T	**NND21120002p-T**	**NND21120003p-T**
KRA300	0	Drug free urine	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		-50% Cut-off	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-	-,-,-,-,-,-,-,-,-,-
		+50% Cut-off	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+	+,+,+,+,+,+,+,+,+,+
	7	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	14	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	21	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	28	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	35	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+
	42	Drug free urine	-,-,-	-,-,-	-,-,-
		-50% Cut-off	-,-,-	-,-,-	-,-,-
		+50% Cut-off	+,+,+	+,+,+	+,+,+

Conclusion: Multi-Drug Rapid Test Panel is stable at 45℃ for 84 days or 55℃ for 42 days. These data were plotted on an Arrhenius Plot and the shelf life of this product was determined to be at least 24 months from the data of manufacture.

2.8 Precision

A study was conducted at three hospitals by untrained operators using three different lots of product which assemble in cup to demonstrate the within run, between run and between operator precision. An identical cup of coded specimen, containing drug free urine, -50% Cut-off, -25%Cut-off, +25% Cut-off and +50% Cut-off to each site.

Results:

AMP 1,000

Amphetamine	n	Site A	Site B	Site C

conc. (ng/mL)	per site	-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	8	2	9	1
1,250	10	1	9	2	8	2	8
1,500	10	0	10	0	10	0	10

AMP 500

Amphetamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	9	1	9	1
625	10	2	8	1	9	2	8
750	10	0	10	0	10	0	10

AMP 300

Amphetamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	8	2	8	2	8	2
375	10	2	8	2	8	2	8
450	10	0	10	0	10	0	10

BAR 300

Secobarbital conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	8	2	9	1
375	10	2	8	1	9	2	8
450	10	0	10	0	10	0	10

BAR 200

Secobarbital conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	9	1	9	1
250	10	1	9	1	9	1	9
300	10	0	10	0	10	0	10

BZO 500

Oxazepam	n	Site A	Site B	Site C

conc. (ng/mL)	per site	-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	9	1	8	2
625	10	1	9	2	8	1	9
750	10	0	10	0	10	0	10

BZO 300

Oxazepam conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

BZO 200

Oxazepam conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	8	2	9	1
250	10	1	9	1	9	2	8
300	10	0	10	0	10	0	10

BZO 100

Oxazepam conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	8	2	7	3
125	10	1	9	1	9	2	8
150	10	0	10	0	10	0	10

BUP10

Buprenorphine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
5	10	10	0	10	0	10	0
7.5	10	9	1	9	1	8	2
12.5	10	1	9	1	9	1	9
15	10	0	10	0	10	0	10

BUP5

Buprenorphine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
2.5	10	10	0	10	0	10	0
3.75	10	9	1	9	1	8	2
6.25	10	1	9	1	9	1	9
7.5	10	0	10	0	10	0	10

COC 300

Benzoylecgonine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

COC 100

Benzoylecgonine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	9	1
125	10	2	8	2	8	2	8
150	10	0	10	0	10	0	10

COC150

Cocaine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
75	10	10	0	10	0	10	0
112.5	10	9	1	9	1	9	1
187.5	10	2	8	2	8	2	8
225	10	0	10	0	10	0	10

COC200

Cocaine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	9	1	9	1
250	10	1	9	1	9	1	9
300	10	0	10	0	10	0	10

THC150

11-nor-∆9-COOH	n	Site A	Site B	Site C

conc. (ng/mL)	per site	-	+	-	+	-	+
0	10	10	0	10	0	10	0
75	10	10	0	10	0	10	0
112.5	10	9	1	9	1	9	1
187.5	10	2	8	1	9	1	9
225	10	0	10	0	10	0	1

THC50

11-nor-Δ9-COOH conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
25	10	10	0	10	0	10	0
37.5	10	9	1	8	2	9	1
62.5	10	1	9	1	9	2	8
75	10	0	10	0	10	0	1

THC25

11-nor-Δ9-COOH conc. (ng/mL)	n per	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
12.5	10	10	0	10	0	10	0
18.75	10	8	2	8	2	8	2
31.25	10	1	9	1	9	2	8
37.5	10	0	10	0	10	0	10

THC300

11-nor-Δ9-COOH conc (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	8	2	9	1	9	1
375	10	2	8	3	7	1	9
450	10	0	10	0	10	0	10

THC20

11-nor-Δ9-COOH conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	-	+	-
0	10	10	0	10	0	10	0
10	10	10	0	10	0	10	0
15	10	8	2	8	2	8	2
25	10	1	9	1	9	2	8
30	10	0	10	0	10	0	10

THC200

THC Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	9	1	9	1
250	10	2	8	1	9	1	9
300	10	0	10	0	10	0	10

THC30

THC Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
15	10	10	0	10	0	10	0
22.5	10	9	1	9	1	9	1
37.5	10	2	8	2	8	1	9
45	10	0	10	0	10	0	10

MTD300

Methadone conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	8	2	9	1
375	10	2	8	1	9	2	8
450	10	0	10	0	10	0	10

MTD200

Methadone conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	8	2	9	1

250	10	2	8	1	9	2	8
300	10	0	10	0	10	0	10

MET1, 000

Methamphetamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	9	1	9	1
1,250	10	1	9	2	8	1	9
1,500	10	0	10	0	10	0	10

MET 500

Methamphetamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	9	1	9	1
625	10	1	9	1	9	1	9
750	10	0	10	0	10	0	10

MET 300

Methamphetamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

MDMA1, 000

Methylenedioxymethamphetamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	9	1	8	2
1,250	10	1	9	1	9	1	9
1,500	10	0	10	0	10	0	10

MDMA 500

Methylenedioxymethamphetamine conc. (ng/mL)	n	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	9	1	9	1
625	10	1	9	1	9	1	9
750	10	0	10	0	1	0	10

MDMA 300

Methylenedioxymethamphetamine conc. (ng/mL)	n per	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	8	2	9	1	7	3
375	10	2	8	1	9	1	9
450	10	0	10	0	1	0	10

MOP 300

Morphine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	2	8
450	10	0	10	0	10	0	10

MOP 100

Morphine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	9	1
125	10	1	9	1	9	1	9
150	10	0	10	0	10	0	10

MOP 200

Morphine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	7	3	9	1	9	1

250	10	1	9	2	8	1	9
300	10	0	10	0	10	0	10

MQL 300

Methaqualone conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

OPI 2,000

Morphine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
1,000	10	10	0	10	0	10	0
1,500	10	9	1	9	1	9	1
2,500	10	1	9	1	9	1	9
3,000	10	0	10	0	10	0	10

PCP 25

Phencyclidine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
12.5	10	10	0	10	0	10	0
18.75	10	8	2	9	1	9	1
31.25	10	1	9	1	9	1	9
37.5	10	0	10	0	10	0	10

PPX 300

Propoxyphene conc. (ng/mL)	n per	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	8	2	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

TCA 1000

Nortriptyline conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+

0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	8	2	8	2
1,250	10	1	9	1	9	1	9
1,500	10	0	10	0	10	0	10

TCA 500

Nortriptyline conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	7	3	9	1	8	2
625	10	2	8	1	9	1	9
750	10	0	10	0	10	0	10

TML100

Tramadol conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	8	2
125	10	1	9	1	9	2	8
150	10	0	10	0	10	0	10

TML200

Tramadol conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	9	1	8	2
250	10	1	9	1	9	2	8
300	10	0	10	0	10	0	10

TML300

Tramadol conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0

150	10	10	0	10	0	10	0
225	10	9	1	9	1	8	2
375	10	1	9	1	9	2	8
450	10	0	10	0	10	0	10

KET1, 000

Ketamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	8	2	9	1
1,250	10	1	9	1	9	2	8
1,500	10	0	10	0	10	0	10

KET500

Ketamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	9	1	8	2
625	10	1	9	1	9	2	8
750	10	0	10	0	10	0	10

KET300

Ketamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

KET100

Ketamine conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	9	1

125	10	1	9	1	9	2	8
150	10	0	10	0	10	0	10

OXY100

Oxycodone Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	9	1
125	10	1	9	1	9	1	9
150	10	0	10	0	10	0	10

OXY300

Oxycodone Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

COT 200

Cotinine Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	9	1	9	1
250	10	1	9	1	9	2	8
300	10	0	10	0	10	0	10

COT 100

Cotinine Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	9	1
125	10	1	9	1	9	1	9
150	10	0	10	0	10	0	10

COT 500

Cotinine Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	7	3	9	1
625	10	2	8	1	9	2	8
750	10	0	10	0	10	0	10

COT 50

Cotinine Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
25	10	10	0	10	0	10	0
37.5	10	8	2	9	1	9	1
62.5	10	1	9	1	9	2	8
150	10	0	10	0	10	0	10

COT 10

Cotinine Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
5	10	10	0	10	0	10	0
7.5	10	9	1	9	1	9	1
12.5	10	1	9	1	9	2	8
15	10	0	10	0	10	0	10

EDDP 300

EDDP Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	2	8	1	9
450	10	0	10	0	10	0	10

EDDP 100

EDDP Concentration (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	9	1
125	10	1	9	1	9	1	9
150	10	0	10	0	10	0	10

FYL 20

FYL Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	10
10	10	10	0	10	10	10	10
15	10	9	1	9	1	9	1
25	10	1	9	1	9	1	9
30	10	0	10	0	10	10	0

FYL 10

FYL	n per	Site A		Site B		Site C	

Concentration (ng/mL)	Site	-	+	-	+	-	+
0	10	10	0	10	0	10	0
5	10	10	0	10	0	10	0
7.5	10	9	1	9	1	9	1
12.5	10	1	9	1	9	1	9
15	10	0	10	0	10	0	10

FYL 100

FYL Concentration (ng/mL)	**n per Site**	**Site A**		**Site B**		**Site C**	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	8	2	9	1
125	10	1	9	1	9	2	8
150	10	0	10	0	10	0	10

FYL 200

FYL Concentration (ng/mL)	**n per Site**	**Site A**		**Site B**		**Site C**	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	9	1	9	1
250	10	1	9	1	9	1	9
300	10	0	10	0	10	0	10

FYL 300

FYL Concentration (ng/mL)	**n per Site**	**Site A**		**Site B**		**Site C**	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	8	2	9	1	8	2
375	10	1	9	2	8	1	9
450	10	0	10	0	10	0	10

K2 50

K2 Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
25	10	10	0	10	0	10	0
37.5	10	8	2	8	2	9	1
67.5	10	1	9	2	8	2	8
75	10	0	10	0	10	0	10

K2 30

K2 Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
15	10	10	0	10	0	10	0
22.5	10	8	2	9	1	9	1
37.5	10	1	9	1	9	1	9
45	10	0	10	0	10	0	10

K2 25

K2 Concentration (ng/mL)	**n per Site**	**Site A**		**Site B**		**Site C**	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
12.5	10	10	0	10	0	10	0
17.5	10	7	3	8	2	8	2
31.25	10	1	9	1	9	2	8
37.5	10	0	10	0	10	0	10

6-MAM 10

6-MAM Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
5	10	10	0	10	0	10	0
7.5	10	9	1	9	1	9	1
12.5	10	1	9	1	9	1	9
15	10	0	10	0	10	0	10

CLO 400

Clonazepam conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
200	10	10	0	10	0	10	0

300	10	9	1	8	2	9	1
500	10	1	9	2	8	1	9
600	10	0	10	0	10	0	10

CLO 150

Clonazepam conc. (ng/mL)	n per site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
75	10	10	0	10	0	10	0
112	10	9	1	8	2	9	1
187	10	1	9	2	8	1	9
225	10	0	10	0	10	0	10

MDA 500

MDA Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	9	1	9	1
625	10	1	9	1	9	1	9
750	10	0	10	0	10	0	10

DIA300

DIA Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	1	9
450	10	0	10	0	10	0	10

DIA200

DIA Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+

0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	9	1	9	1	9	1
250	10	1	9	1	9	1	9
300	10	0	10	0	10	0	10

ETG300

ETG Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	7	3	8	2	9	1
375	10	1	9	2	8	1	9
450	10	0	10	0	10	0	10

ETG500

ETG Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	8	2	9	1
625	10	1	9	2	8	2	8
750	10	0	10	0	10	0	10

ETG1000

ETG Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	8	2	8	2	9	1

1250	10	1	9	2	8	2	8
1500	10	0	10	0	10	0	10

ACE 5000

ACE Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
2500	10	10	0	10	0	10	0
3750	10	9	1	9	1	8	2
6250	10	1	9	1	9	1	9
7500	10	0	10	0	10	0	10

LSD 10

LSD Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
5	10	10	0	10	0	10	0
7.5	10	9	1	9	1	9	1
12.5	10	1	9	1	9	1	9
15	10	0	10	0	10	0	10

LSD 20

LSD Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
10	10	10	0	10	0	10	0
15	10	9	1	9	1	9	1
25	10	1	9	1	9	1	9
30	10	0	10	0	10	0	10

LSD 50

LSD	n	Site A	Site B	Site C

Concentration (ng/mL)	per Site	-	+	-	+	-	+
0	10	10	0	10	0	10	0
25	10	10	0	10	0	10	0
37.5	10	9	1	9	1	9	1
62.5	10	1	9	1	9	1	9
75	10	0	10	0	10	0	10

MDPV1000

MDPV Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	9	1	8	2
1250	10	1	9	1	9	1	9
1500	10	0	10	0	10	0	10

MDPV500

MDPV Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	9	1	8	2
1250	10	2	8	1	9	1	9
1500	10	0	10	0	10	0	10

MDPV300

MDPV Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0

225	10	8	2	9	1	9	1
375	10	1	9	1	9	2	8
450	10	0	10	0	10	0	10

MPD300

MPD Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	8	2	9	1
375	10	1	9	2	8	1	9
450	10	0	10	0	10	0	10

MPD1000

MPD Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	9	1	9	1
1250	10	2	8	2	8	2	8
1500	10	0	10	0	10	0	10

ZOL 50

ZOL Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
25	10	9	1	10	0	10	0
75	10	0	10	1	9	0	10

ZOP 50

ZOP Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+

0	10	10	0	10	0	10	0
25	10	10	0	10	0	10	0
37.5	10	9	1	8	2	9	1
62.5	10	2	8	2	8	2	8
75	10	0	10	0	10	0	10

MCAT 500

MCAT Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	8	2	9	1
625	10	2	8	2	8	2	8
750	10	0	10	0	10	0	10

7-ACL 300

7-ACL Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	8	2	9	1	9	1
375	10	2	8	2	8	3	7
450	10	0	10	0	10	0	10

7-ACL 100

7-ACL Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	7	3	7	3	9	1

125	10	2	8	1	9	2	8
150	10	0	10	0	10	0	10

7-ACL 200

7-ACL Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	8	2	9	1	8	2
250	10	2	8	2	8	2	8
300	10	0	10	0	10	0	10

CFYL 500

CFYL Concentration (ng/mL)	**n per Site**	**Site A**		**Site B**		**Site C**	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	7	3	9	1	8	2
625	10	2	8	1	9	2	8
750	10	0	10	0	10	0	10

CFYL 250

CFYL Concentration (ng/mL)	**n per Site**	**Site A**		**Site B**		**Site C**	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
125	10	10	0	10	0	10	0
187.5	10	7	3	9	1	8	2
312.5	10	2	8	1	9	2	8
375	10	0	10	0	10	0	10

CAF 1000

CAF Concentration (ng/ml)	n per	Site A	Site B	Site C

	Site	-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	8	2	9	1
1250	10	2	8	2	8	2	8
1500	10	0	10	0	10	0	10

CAT 150

CAT Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
75	10	10	0	10	0	10	0
112.5	10	9	1	8	2	9	1
187.5	10	2	8	2	8	2	8
225	10	0	10	0	10	0	10

TRO 350

TRO Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
175	10	10	0	10	0	10	0
262.5	10	8	2	8	2	8	2
437.5	10	2	8	2	8	2	8
525	10	0	10	0	10	0	10

MEP 100

MEP Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0

75	10	9	1	8	2	9	1
125	10	2	8	2	8	2	8
150	10	0	10	0	10	0	10

ALP 100

ALP Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	8	2	9	1
125	10	2	8	2	8	2	8
150	10	0	10	0	10	0	10

ABP 10

AB-PINACA Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
5	10	10	0	10	0	10	0
7.5	10	8	2	8	2	9	1
12.5	10	2	8	3	7	1	9
15	10	0	10	0	10	0	10

APVP1000

α-PVP Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	8	2	9	1	9	1
1,250	10	2	8	3	7	1	9
1,500	10	0	10	0	10	0	10

APVP500

α-PVP Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	8	2	9	1
625	10	2	8	2	8	1	9
750	10	0	10	0	10	0	10

APVP300

α-PVP Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	7	3	8	2	9	1
375	10	2	8	1	9	1	9
450	10	0	10	0	10	0	10

CNB 500

CNB Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	9	1	9	1
625	10	2	8	1	9	1	9
750	10	0	10	0	10	0	10

MPRD 100

Normeperidine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0

50	10	10	0	10	0	10	0
75	10	8	2	9	1	9	1
125	10	2	8	3	7	1	9
150	10	0	10	0	10	0	10

PGB 50,000

PGB Concentration (ug/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
25	10	10	0	10	0	10	0
37.5	10	8	2	8	2	8	2
62.5	10	2	8	2	8	2	8
75	10	0	10	0	10	0	10

PGB 500

PGB Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	8	2	8	2
625	10	2	8	2	8	2	8
750	10	0	10	0	10	0	10

TZD 200

TZD Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
100	10	10	0	10	0	10	0
150	10	8	2	8	2	8	2
250	10	1	9	2	8	2	8

300	10	0	10	0	10	0	10

UR-144 25

UR-144 5-Pentanoic Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
12.5	10	10	0	10	0	10	0
18.75	10	9	1	8	2	9	1
31.25	10	1	9	2	8	2	8
62.5	10	0	10	0	10	0	10

ZAL 100

ZAL Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	10	0	10	0	10	0
125	10	2	8	1	9	1	9
150	10	0	10	0	10	0	10

MES 100

MES Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
50	10	10	0	10	0	10	0
75	10	9	1	9	1	9	1
125	10	2	8	1	9	1	9
150	10	0	10	0	10	0	10

MES 300

MES Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+

0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	2	8	2	8	1	9
450	10	0	10	0	10	0	10

GAB 2000

GAB Concentration (ng/ml)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
1000	10	10	0	10	0	10	0
1500	10	9	1	9	1	8	2
2500	10	2	8	2	8	2	8
3000	10	0	10	0	10	0	10

QTP 1000

QTP Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	9	1	9	1
1250	10	1	9	1	9	1	9
1500	10	0	10	0	10	0	10

PAP 500

PAP Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	9	1	9	1

625	10	2	8	2	8	2	8
750	10	0	10	0	10	0	10

KRA 300

KRA Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
150	10	10	0	10	0	10	0
225	10	9	1	9	1	9	1
375	10	1	9	1	9	2	8
450	10	0	10	0	10	0	10

CAR 2000

CAR Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
1000	10	10	0	10	0	10	0
1500	10	8	2	9	1	9	1
2500	10	2	8	2	8	2	8
3000	10	0	10	0	10	0	10

FLX 500

Fluoxetine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	9	1	8	2
625	10	2	8	2	8	2	8
750	10	0	10	0	10	0	10

RPD 150

Risperidone	n per Site	Site A	Site B	Site C

Concentration (ng/mL)		-	+	-	+	-	+
0	10	10	0	10	0	10	0
75	10	10	0	10	0	10	0
112.5	10	8	2	8	2	9	1
187.5	10	2	8	3	7	1	9
225	10	0	10	0	10	0	10

CIT 500

Citalopram Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	8	2	9	1
675	10	2	8	3	7	1	9
750	10	0	10	0	10	0	10

FKET 1000

F-Ketamine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	7	3	9	1
1250	10	2	8	2	8	1	9
1500	10	0	10	0	10	0	10

TAP 1000

Tapentadol Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0

500	10	10	0	10	0	10	0
750	10	9	1	7	3	9	1
1250	10	2	8	2	8	1	9
1500	10	0	10	0	10	0	10

SCOP 500

Scopolamine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	9	1	9	1	8	2
625	10	1	9	1	9	1	9
750	10	0	10	0	10	0	10

MTZ 500

Mirtazapine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
250	10	10	0	10	0	10	0
375	10	8	2	8	2	9	1
625	10	2	8	3	7	2	8
750	10	0	10	0	10	0	10

NND 1000

N,N-Dimethyltryptamine Concentration (ng/mL)	n per Site	Site A		Site B		Site C	
		-	+	-	+	-	+
0	10	10	0	10	0	10	0
500	10	10	0	10	0	10	0
750	10	9	1	8	2	9	1

1250	10	2	8	2	8	1	9
1500	10	0	10	0	10	0	10

Conclusion: Specimens determined to be negative, -50% cutoff, -25% cutoff, +25% cutoff and +50%cutoff values demonstrate high precision for Multi-Drug Rapid Test Panel.

2.9 **Effect of Urinary Specific Gravity**

The urine specific gravity was measured with Clinical Refractometer. Five urine samples with the specific gravity at each following range were used for the study: (1) Specific gravity >1.025. (2) Specific gravity 1.015 – 1.025. (3) Specific gravity < 1.015. Each urine sample was aliquoted to three parts: (1) Drug free urine (2) Urine samples were spiked with each drug which contained in the panelto-50% Cut-off; (3) Urine samples were spiked with each drug which contained in the panel to+50% Cut-off. The each drug test is assembled in the 10 in 1 cup. The Multi-Drug Rapid Test Panel will test in duplicates with above urine samples Read the result at 5 minutes after sample application.

Results:

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
AMP 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
AMP 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+

	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
AMP 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BAR 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BAR 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BUP 10	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BUP 5	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+

	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BZO 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BZO 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BZO 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
BZO 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COC 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+

	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COC 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COC 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+

	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COC 150	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
KET 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
KET 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+

	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
KET 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
KET 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+

	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MDMA 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MDMA 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MDMA 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+

	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MET 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MET 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+

	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MET 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MOP 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MOP 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+

	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MOP 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MQL 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+

	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MTD 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MTD 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine	-50% Cut-off	+50% Cut-off

OPI 2000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
PCP 25	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
PPX 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+

	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
THC300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
THC 150	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
THC 50	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
THC 25	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
THC 20	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+

	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
THC 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
THC 30	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TCA 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TCA 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TML100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+

	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TML200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TML300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+

	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
OXY100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
OXY300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COT 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+

	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COT 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COT 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+

	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COT 50	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
COT 10	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
EDDP 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+

	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
EDDP 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
FYL 20	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+

	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
FYL 10	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
FYL 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
FYL 200	1	1.010	2	-	-	-	-	+	+

	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
FYL 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
K2 50	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+

	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
K2 30	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
K2 25	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
6-MAM 10	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CLO400	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CLO150	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+

	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
DIA300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
DIA200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+

	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MDA 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ETG300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ETG500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+

	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ETG1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ACE 5000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+

	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
LSD10	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
LSD20	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
LSD50	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MDPV1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MDPV500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+

	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MDPV300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MPD300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+

	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MPD1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ZOL 50	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ZOP 50	1	1.010	2	-	-	-	-	+	+

	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MCAT 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
7-ACL 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+

	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
7-ACL 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
7-ACL 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CFYL 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CFYL 250	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CAF 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+

	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CAT 150	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TRO 350	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+

	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MEP 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ALP 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ABP 10	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+

	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
APVP 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
APVP 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+

	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
APVP 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CNB 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MPRD 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
PGB 50,000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
PGB 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+

	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TZD 200	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
UR-144 25	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+

	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
ZAL 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MES 100	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MES 300	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+

	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
GAB 2000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TLD 50	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+

	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
QTP 1000	1	1.018	2	-	-	-	-	+	+
	2	1.012	2	-	-	-	-	+	+
	3	1.008	2	-	-	-	-	+	+
	4	1.006	2	-	-	-	-	+	+
	5	1.022	2	-	-	-	-	+	+
	6	1.024	2	-	-	-	-	+	+
	7	1.017	2	-	-	-	-	+	+
	8	1.028	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.035	2	-	-	-	-	+	+
	11	1.031	2	-	-	-	-	+	+
	12	1.043	2	-	-	-	-	+	+
	13	1.037	2	-	-	-	-	+	+
	14	1.040	2	-	-	-	-	+	+
	15	1.033	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
PAP 500	1	1.018	2	-	-	-	-	+	+
	2	1.012	2	-	-	-	-	+	+
	3	1.008	2	-	-	-	-	+	+
	4	1.006	2	-	-	-	-	+	+
	5	1.022	2	-	-	-	-	+	+
	6	1.024	2	-	-	-	-	+	+
	7	1.017	2	-	-	-	-	+	+
	8	1.028	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.035	2	-	-	-	-	+	+
	11	1.031	2	-	-	-	-	+	+
	12	1.043	2	-	-	-	-	+	+
	13	1.037	2	-	-	-	-	+	+
	14	1.040	2	-	-	-	-	+	+
	15	1.033	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
KRA 300	1	1.018	2	-	-	-	-	+	+

	2	1.012	2	-	-	-	-	+	+
	3	1.008	2	-	-	-	-	+	+
	4	1.006	2	-	-	-	-	+	+
	5	1.022	2	-	-	-	-	+	+
	6	1.024	2	-	-	-	-	+	+
	7	1.017	2	-	-	-	-	+	+
	8	1.028	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.035	2	-	-	-	-	+	+
	11	1.031	2	-	-	-	-	+	+
	12	1.043	2	-	-	-	-	+	+
	13	1.037	2	-	-	-	-	+	+
	14	1.040	2	-	-	-	-	+	+
	15	1.033	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CAR 2000	1	1.018	2	-	-	-	-	+	+
	2	1.012	2	-	-	-	-	+	+
	3	1.008	2	-	-	-	-	+	+
	4	1.006	2	-	-	-	-	+	+
	5	1.022	2	-	-	-	-	+	+
	6	1.024	2	-	-	-	-	+	+
	7	1.017	2	-	-	-	-	+	+
	8	1.028	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.035	2	-	-	-	-	+	+
	11	1.031	2	-	-	-	-	+	+
	12	1.043	2	-	-	-	-	+	+
	13	1.037	2	-	-	-	-	+	+
	14	1.040	2	-	-	-	-	+	+
	15	1.033	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
FLX 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+

	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
RPD 150	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
CIT 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
FKET 1000	1	1.010	2	-	-	-	-	+	+

	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
TAP 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
SCOP 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+

	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
MTZ 500	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Product	Urine ID	Urine Specific Gravity	N*	Drug free urine		-50% Cut-off		+50% Cut-off	
NND 1000	1	1.010	2	-	-	-	-	+	+
	2	1.011	2	-	-	-	-	+	+
	3	1.007	2	-	-	-	-	+	+
	4	1.005	2	-	-	-	-	+	+
	5	1.012	2	-	-	-	-	+	+
	6	1.014	2	-	-	-	-	+	+
	7	1.015	2	-	-	-	-	+	+
	8	1.018	2	-	-	-	-	+	+
	9	1.017	2	-	-	-	-	+	+
	10	1.025	2	-	-	-	-	+	+
	11	1.036	2	-	-	-	-	+	+
	12	1.045	2	-	-	-	-	+	+
	13	1.036	2	-	-	-	-	+	+
	14	1.041	2	-	-	-	-	+	+
	15	1.043	2	-	-	-	-	+	+

Conclusion: The results demonstrate that varying ranges of urinary specific gravity do not affect the test results.

2.10 Effect of Urinary pH

Approximately 150 ml of negative urine pool were allocated into 5 test tubes. The pH of the urine was adjusted within the range of 5 to 9 in 1 pH unit increments using NaOH or HCl. Each pH-adjusted urine was split into three glass containers (10 ml/container): Multi-Drug Rapid Test Panel was added to two aliquots to bring the final sample concentration of -50% cut-off and +50% cut-off respectively. Each pH-adjusted urine was tested in duplicate and read 5 and 10 minutes after sample application.

Results:

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
AMP 1000 Lot#: AMP14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
AMP 500 Lot#: AMP14030007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
AMP 300 Lot#: AMP14030013p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BAR 300 Lot#: BAR14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BAR 200 Lot#: BAR14030007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BUP 10 Lot#: BUP14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BUP 5 Lot#: BUP14030007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+

	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BZO 500 Lot#:BZO14030019 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BZO 300 Lot#:BZO14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BZO 200 Lot#:BZO14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
BZO 100	5	-	-	-	-	+	+

Lot#:BZO14030001 3p-T	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COC 300 Lot#:COC14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COC 100 Lot#:COC14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COC 150 Lot#:COC14030013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+

	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COC 200 Lot#:COC14030019 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
KET 1000 Lot#:KET14030001p -T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
KET 500 Lot#:KET14030007p -T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
KET 300 Lot#:KET140300013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+

	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
KET 100 Lot#:KET140300019 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MDMA 1000 LOT#: MDMA14030007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MDMA 500 LOT#: MDMA14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine	-50% cutoff	+50% cutoff

		5min	10min	5min	10min	5min	10min
MDMA 300 LOT#: MDMA14030013p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MET 1000 LOT#: MET14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MET 500 LOT#: MET14030007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MET 300 LOT#:MET14030013p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+

	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MOP 300 LOT#: MOP14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MOP 200 LOT#: MOP14030013p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MOP 100 LOT#: MOP14030007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MQL 300 LOT#:MQL1403000	5	-	-	-	-	+	+
	5	-	-	-	-	+	+

1p-T	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MTD 300 LOT#:MTD1403000 1p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MTD 200 LOT#:MTD1403000 7p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
OPI 2000 LOT#:OPI14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
PCP 25 LOT#: PCP14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
PPX 300 LOT#:PPX14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
THC 150 LOT#:THC14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
THC 50 LOT#:THC14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+

	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
THC 25 LOT#:THC14030001 3p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
THC 300 LOT#:THC14030019 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
THC 20 LOT#:THC14030037 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min

THC 200 LOT#:THC14030025 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
THC 30 LOT#:THC14030031 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TCA 1000 LOT#:TCA14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TCA 500 LOT#:TCA14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+

	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TML 100 LOT#:TML14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TML 200 LOT#:TML14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TML300 LOT#:TML14030013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
OXY100 LOT#:OXY14030001	5	-	-	-	-	+	+
	5	-	-	-	-	+	+

p-T	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
OXY300 LOT#:OXY14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COT 200 LOT#:COT14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COT 100 LOT#:COT14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COT 500 LOT#:COT14030013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COT 50 LOT#:COT14030019 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
COT 10 LOT#:COT14030025 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
EDDP300 LOT#:EDDP14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+

	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
EDDP 100 LOT#:EDDP14030007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
FYL 20 LOT#:FYL18040001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
FYL 10 LOT#:FYL18040007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min

FYL 200 LOT#:FYL18040019 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
FYL 100 LOT#:FYL18040013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
FYL 300 LOT#:FYL18040025 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
K2 50 LOT#:SMA14030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+

	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
K2 30 LOT#:SMA1403000 7p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
K2 25 LOT#:SMA1403001 3p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
6-MAM 10 LOT#:MAM1403000 1p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CLO 400 LOT#:CLO14030001	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+

p-T	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CLO 150 LOT#:CLO14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
DIA300 LOT#:DIA14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
DIA 200 LOT#:DIA14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MDA 500 LOT#:MDA1403000 1p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ETG300 LOT#:ETG14030013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ETG500 LOT#:ETG14030001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ETG1000 LOT#:ETG14030007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+

	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ACE 5000 LOT#:ACE14060001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
LSD 10 LOT#:LSD14060013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
LSD 20 LOT#:LSD14060001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
LSD 50	5	-	-	-	-	+	+

LOT#:LSD14060007 p-T	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MDPV1000 LOT#:MDPV140800 01p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MDPV500 LOT#:MDPV140800 07p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MDPV300 LOT#:MDPV150900 01p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+

	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MPD300 LOT#:MPD14060001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MPD1000 LOT#:MPD14060007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ZOL 50 LOT#:ZOL15010001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ZOP 50 LOT#:ZOP16080001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+

	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MCAT 500 LOT#:MCAT16080001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
7-ACL 300 LOT#:ACL14100001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
7-ACL 100 LOT#:ACL14100007 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine	-50% cutoff	+50% cutoff

		5min	10min	5min	10min	5min	10min
7-ACL 200 LOT#:ACL14100013 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CFYL 500 LOT#:CFYL15020001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CFYL 250 LOT#:CFYL15020007p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CAF 1000 LOT#:CAF15080001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+

	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CAT 150 LOT#:CAT14080001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TRO 350 LOT#:TRO15050001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MEP 100 LOT#:MEP15070001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ALP 100	5	-	-	-	-	+	+

LOT#:ALP16040001 p-T	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
ABP 10 LOT#:ABP17010001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
APVP1000 LOT#:APVP1505000 1p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
APVP500 LOT#:APVP1505000 7p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+

	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
APVP300 LOT#:APVP1505001 3p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CNB 500 LOT#:CNB18060001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MPRD 100 LOT#:MPRD180600 01p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
PGB 50,000 LOT#:PGB16010001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+

	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
PGB 500 LOT#:PGB19040001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TZD 200 LOT#:TZD18070001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
UR-144 25 LOT#:UR18060001p -T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine	-50% cutoff	+50% cutoff

		5min	10min	5min	10min	5min	10min
ZAL 100 LOT#:ZAL18050001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MES 100 LOT#:MES18060001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MES 300 LOT#:MES18090001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
GAB 2000 LOT#:GAB18070001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+

	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+
Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
QTP 1000 LOT#: QTP19030001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TLD 50 LOT#: TLD18110001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
PAP 500 LOT#: PAP18110001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
KRA 300 LOT#: KRA19040001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+

	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
CAR 2000 LOT#: CAR18100001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
FLX 500 LOT#:FLX18110001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+
Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
RPD 150 LOT#RPD20090001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine	-50% cutoff	+50% cutoff

		5min	10min	5min	10min	5min	10min
CIT 500 LOT#: CIT20120001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
FKET 1000 LOT#:FLX18110001 p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
MTZ 500 LOT#: MTZ21110007-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
TAP 1000 LOT#: TAP21090001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+

	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
SCOP 500 LOT#: SCOP21100004-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Product	pH	Negative urine		-50% cutoff		+50% cutoff	
		5min	10min	5min	10min	5min	10min
NND 1000 LOT#: NND21120001p-T	5	-	-	-	-	+	+
	5	-	-	-	-	+	+
	6	-	-	-	-	+	+
	6	-	-	-	-	+	+
	7	-	-	-	-	+	+
	7	-	-	-	-	+	+
	8	-	-	-	-	+	+
	8	-	-	-	-	+	+
	9	-	-	-	-	+	+
	9	-	-	-	-	+	+

Conclusion: The results demonstrate that varying ranges of pH from 5-9 do not interfere with the performance of the Multi-Drug Rapid Test Panel.

2.11 Open Pouch Stability

Drug standards including negative urine, the -50% cutoff, + 50% cutoff and 300% cutoff were run in triplicate using the Multi-Drug rapid tests which were opened pouch after 10 minutes, 20 minutes, 30 minutes, 1 hour, 1.5 hours and 2 hours at room temperature. One lot test (Lot 1 for each) was used. All results were read as positive or negative at 5 and 10 minutes after sample application. Read the results by ALLTEST gold color card at 5 minutes after sample application, T line <=3.5 regard as positive, T line >=4 regard as negative.

AMP1000 Lot: AMP14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

AMP500 Lot: AMP14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

1.5 hours	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

AMP300 Lot: AMP14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

BAR300 Lot: BAR14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

20 minutes	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

BAR200 Lot: BAR14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
	1	-	-	-	-	+	+	+	+

2 hours	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+

BZO500 Lot: BZO14030019p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+

BZO300 Lot: BZO14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

30 minutes	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+

BZO200 Lot: BZO14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+

BZO100 Lot: BZO14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	+	+

BUP10 Lot: BUP14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

BUP5 Lot: BUP14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

COC300 Lot: COC14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**

10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

COC100 Lot: COC14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

COC150 Lot: COC14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	-	+
	5	-	-	-	-	-	-	-	-

COC200 Lot: COC14030019p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC150 Lot: THC14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC50 Lot: THC14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+/-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC25 Lot: THC14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+/-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC300 Lot: THC14030019p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC200 Lot: THC14030025p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC30 Lot: THC14030031p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

30 minutes	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC20 Lot: THC14030037p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-
	5	-	-	-	-	-	-	-	-

THC200 Lot: THC14030025p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

THC30 Lot: THC14030031p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

MTD300 Lot: MTD14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	-

MTD200 Lot: MTD14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**

10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MET1000 Lot: MET14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MET500 Lot: MET14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

MET300 Lot: MET14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	-

MDMA1000 Lot: MDMA14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MDMA500 Lot: MDMA14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MDMA300 Lot: MDMA14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

MOP300 Lot: MOP14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

MOP100 Lot: MOP14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MOP200 Lot: MOP14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

30 minutes	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MQL300 Lot: MQL14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

OPI2000 Lot: OPI14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

PCP25 Lot: PCP14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

1 hour	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+/-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

PPX300 Lot: PPX14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

TCA1000 Lot: TCA14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
	1	-	-	-	-	+	+	+	+

10 minutes	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

TCA500 Lot: TCA14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

TML100 Lot: TML14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

TML200 Lot: TML14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

TML300 Lot: TML14030013p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+/-	+	+
	5	-	-	-	-	-	-	-	+

KET1000 Lot: KET14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

KET500 Lot: KET14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

KET300 Lot: KET14030013p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

KET100 Lot: KET14030019p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

OXY100 Lot: OXY14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+/-	-	-	+

OXY300 Lot: OXY14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+

COT200 Lot: COT14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

COT100 Lot: COT14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

COT500 Lot: COT14030013p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

COT50 Lot: COT14030019p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	-	+	+
	5	-	-	-	-	-	-	-	+

COT10 Lot: COT14030025p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

30 minutes	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

EDDP300 Lot: EDDP14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

EDDP100 Lot: EDDP14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	+	-	+

FYL20 Lot: FYL18040001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

1 hour	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	+	+	+	+	+	+
	3	-	-	+	+	+	+	+	+
	4	-	-	+	+	+	+	+	+
	5	-	-	+	+	+	+	+	+
2 hours	1	-	-	+	+	+	+	+	+
	2	-	-	+	+	+	+	+	+
	3	-	-	+	+	+	+	+	+
	4	-	-	+	+	+	+	+	+
	5	-	-	+	+	+	+	+	+

FYL10 Lot: FYL18040007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	+	+	+	+	+	+
	3	-	-	+	+	+	+	+	+
	4	-	-	+	+	+	+	+	+
	5	-	-	+	+	+	+	+	+
2 hours	1	-	-	+	+	+	+	+	+
	2	-	-	+	+	+	+	+	+
	3	-	-	+	+	+	+	+	+
	4	-	-	+	+	+	+	+	+
	5	-	-	+	+	+	+	+	+
	5	-	-	-	-	-	+	-	+

FYL200 Lot: FYL18040019p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**

10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

FYL100 Lot: FYL18040013p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

FYL300 Lot: FYL18040025p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

K2-50 Lot: SMA14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

K2-30 Lot: SMA14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	+/-	+	+
	5	-	-	-	-	-	-	-	+

K2-25Lot: SMA14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

6-MAM 10 Lot: MAM14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	+	+	+
	5	-	-	-	-	-	-	-	+

CLO400 Lot: CLO14030001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	+	+	+
	5	-	-	-	-	-	-	-	+

CLO150 Lot: CLO14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

ETG500 Lot: ETG14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

30 minutes	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	+	+	+
	5	-	-	-	-	-	-	-	+

ETG1000 Lot: ETG14030007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

ETG300 Lot: ETG14030013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

DIA300 Lot: DIA14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

1 hour	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

DIA200 Lot: DIA14030007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

MDA500 Lot: MDA14030001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
	1	-	-	-	-	+	+	+	+

10 minutes	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

ACE5000 Lot: ACE14060001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+/-	+	+	+
	5	-	-	-	-	-	-	-	+

CAT150 Lot: CAT14080001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	-	+

LSD20 Lot: LSD14060001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

LSD50 Lot: LSD14060007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

LSD10 Lot: LSD14060013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	+	+	+
	5	-	-	-	-	-	-	-	+

ZOL50 Lot: ZOL15010001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	+	+	+
	5	-	-	-	-	-	-	-	+

MDPV1000 Lot: MDPV14080001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+

	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MDPV500 Lot: MDPV14080007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MDPV300 Lot: MDPV15090001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MPD300 Lot: MPD14060001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	**10min**	**5min**	**10min**	**5min**	**10min**	**5min**	**10min**
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

MPD1000 Lot: MPD14060007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

FLX500 Lot: FLX18110001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

1 hour	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	-	+

ABP10 Lot: ABP17010001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CFYL500 Lot: CFYL15020001p-T

Time after opened pouch	**N**	**Negative urine**		**CFYL –50% cutoff**		**CFYL + 50% cutoff**		**CFYL 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
	1	-	-	-	-	+	+	+	+

10 minutes	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	+	+	+	+	+	+
	3	-	-	+	+	+	+	+	+
	4	-	-	+	+	+	+	+	+
	5	-	-	+	+	+	+	+	+
2 hours	1	-	-	+	+	+	+	+	+
	2	-	-	+	+	+	+	+	+
	3	-	-	+	+	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CFYL250 Lot: CFYL15020007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

α-PVP 1000 Lot: APVP15050001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

α-PVP 500 Lot: APVP15050007p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

α-PVP 300 Lot: APVP15050013p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

7-ACL300 LOT#:ACL14100001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

7-ACL200 LOT#:ACL14100013p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

7-ACL100 LOT#:ACL14100007p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CNB 500 LOT#:CNB18060001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

MPRD 100 LOT#: MPRD18060001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

TZD 200 LOT#: TZD18070001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

ZAL 100 LOT#: ZAL18050001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

UR-144 25 LOT#: UR18060001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

MES 100 LOT#: MES18060001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

MES 300 LOT#: MES18090001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

GAB 2000 LOT#: GAB18070001p-T

Time after opened pouch	N	Negative urine		–50% cutoff		+ 50% cutoff		300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

TLD 50 LOT#:TLD18110001p-T

Time after opened pouch	**N**	**Negative urine**		**–50% cutoff**		**+ 50% cutoff**		**300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

QTP1000 LOT#: **QTP19030001p**-T

Time after opened pouch	**N**	**Negative urine**		**Quetiapine –50% cutoff**		**Quetiapine + 50% cutoff**		**Quetiapine 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

PAP 500 LOT#: PAP18110001p-T

Time after opened pouch	**N**	**Negative urine**		**Papaverine –50% cutoff**		**Papaverine + 50% cutoff**		**Papaverine 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

KRA300 LOT#:KRA19040001p-T

Time after opened pouch	**N**	**Negative urine**		**Mitragynine –50% cutoff**		**Mitragynine + 50% cutoff**		**Mitragynine 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+

1.5 hours	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CAR 2000 Lot1#: CAR18100001p-T

Time after opened pouch	N	Negative urine		Carisoprodol –50% cutoff		Carisoprodol + 50% cutoff		Carisoprodol 300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

ZOP 50 LOT#:ZOP16080001p-T

Time after opened pouch	N	Negative urine		Carisoprodol –50% cutoff		Carisoprodol + 50% cutoff		Carisoprodol 300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

20 minutes	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

MCAT 500 LOT#:MCAT16080001p-T

Time after opened pouch	**N**	**Negative urine**		**Carisoprodol –50% cutoff**		**Carisoprodol + 50% cutoff**		**Carisoprodol 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+

2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CFYL 500 LOT#:CFYL15020001p-T

Time after opened pouch	**N**	**Negative urine**		**Carisoprodol –50% cutoff**		**Carisoprodol + 50% cutoff**		**Carisoprodol 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CFYL 250 LOT#:CFYL15020007p-T

Time after opened pouch	**N**	**Negative urine**		**Carisoprodol –50% cutoff**		**Carisoprodol + 50% cutoff**		**Carisoprodol 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+

	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CAF 1000 LOT#:CAF15080001p-T

Time after opened pouch	**N**	**Negative urine**		**Carisoprodol –50% cutoff**		**Carisoprodol + 50% cutoff**		**Carisoprodol 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+

	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

TRO 350 LOT#:TRO15050001p-T

Time after opened pouch	N	Negative urine		Carisoprodol –50% cutoff		Carisoprodol + 50% cutoff		Carisoprodol 300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

MEP 100 LOT#:MEP15070001p-T

Time after opened pouch	N	Negative urine		Carisoprodol –50% cutoff		Carisoprodol + 50% cutoff		Carisoprodol 300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+

30 minutes	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

ALP 100 LOT#:ALP16040001p-T

Time after opened pouch	**N**	**Negative urine**		**Carisoprodol –50% cutoff**		**Carisoprodol + 50% cutoff**		**Carisoprodol 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

PGB 50,000 LOT#:PGB16010001p-T

Time after opened pouch	N	Negative urine		Carisoprodol –50% cutoff		Carisoprodol + 50% cutoff		Carisoprodol 300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

PGB 500 LOT#:PGB19040001p-T

Time after opened pouch	N	Negative urine		Carisoprodol –50% cutoff		Carisoprodol + 50% cutoff		Carisoprodol 300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CAR 2000 LOT#: CAR18100001p-T

Time after opened pouch	**N**	**Negative urine**		**Carisoprodol –50% cutoff**		**Carisoprodol + 50% cutoff**		**Carisoprodol 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

RPD 150 LOT#: RPD20090001p-T

Time after opened pouch	**N**	**Negative urine**	**Risperidone –50% cutoff**	**Risperidone + 50% cutoff**	**Risperidone 300% cutoff**

		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

CIT 500 LOT#: CIT20120001p-T

Time after opened pouch	**N**	**Negative urine**		**Citalopram –50% cutoff**		**Citalopram + 50% cutoff**		**Citalopram 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

1 hour	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

FKET 1000 LOT#: FKE21010001p-T

Time after opened pouch	**N**	**Negative urine**		**F-Ketamine –50% cutoff**		**F-Ketamine + 50% cutoff**		**F-Ketamine 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

TAP 1000 LOT#: TAP21090001p-T

Time after opened pouch	**N**	**Negative urine**	**Tapentadol –50% cutoff**	**Tapentadol + 50% cutoff**	**Tapentadol 300% cutoff**

		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

SCOP 500 LOT#: SCOP21100004-T

Time after opened pouch	**N**	**Negative urine**		**Scopolamine –50% cutoff**		**Scopolamine + 50% cutoff**		**Scopolamine 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
	1	-	-	-	-	+	+	+	+

1 hour	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

MTZ 500 LOT#: MTZ21110007-T

Time after opened pouch	**N**	**Negative urine**		**Mirtazapine –50% cutoff**		**Mirtazapine + 50% cutoff**		**Mirtazapine 300% cutoff**	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	-	-	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

NND 1000 LOT#: NND21120001p-T

Time after opened pouch	N	Negative urine		N,N-Dimethyltryptamine –50% cutoff		N,N-Dimethyltryptamine + 50% cutoff		N,N-Dimethyltryptamine 300% cutoff	
		5min	10min	5min	10min	5min	10min	5min	10min
10 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
20 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
30 minutes	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1 hour	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	+	+	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+
1.5 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	+	+	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	-	-	+	+
	5	-	-	-	-	-	-	+	+
2 hours	1	-	-	-	-	+	+	+	+
	2	-	-	-	-	-	-	+	+
	3	-	-	-	-	-	-	+	+
	4	-	-	-	-	+	+	+	+
	5	-	-	-	-	+	+	+	+

Conclusion: The products were stable at 1 hour after the pouch was opened.

Document History Summary

Version No.	Date	Description	Remark
01	2019.08.23	New Document	N/A
02	2022.02.17	Add new cutoff	N/A

Self test rapido multi-droga (Urina)
Foglio illustrativo
Per l'autotest

Test rapido per la rilevazione qualitativa di molteplici droghe e metaboliti di droghe nelle urine umane.
Per autotest, non per uso diagnostico in vitro.

【USO PREVISTO】

Il self test rapido multi-droga è un immunodosaggio cromatografico rapido per la rilevazione qualitativa di più farmaci e metaboliti di farmaci nelle urine alle seguenti concentrazioni limite:

Test	Calibratore	Valore limite (ng/mL)
Anfetamina (AMP)	d-Anfetamina	1.000
Cocaina (COC)	Benzoilecgonina	300
Marijuana (THC)	11-nor-Δ^9-THC-9 COOH	50
Metanfetamina (MET)	d-Metanfetamina	1.000
Metilendiossimetamfetamina (MDMA)	d,l-metilendiossimetanfetamina	500
Oppiacei (OPI)	Morfina	2.000
Ketamina (KET)	Ketamina	1.000

Il test fornisce solo risultati preliminari. Per ottenere un risultato analitico confermato, è necessario utilizzare un metodo chimico alternativo più specifico. GC/MS o LC/MS sono i metodi di conferma preferiti.

【RIEPILOGO】

Il self test rapido multi-droga è un test di screening rapido delle urine che può essere eseguito senza l'uso di uno strumento. Il test utilizza anticorpi monoclonali per rilevare selettivamente livelli elevati di droghe specifiche nelle urine.

【PRECAUZIONI】

- Leggere l'intero foglio illustrativo prima di eseguire il test.
- Per autotest, non per uso diagnostico *in vitro.*
- Solo per uso esterno.
- Non utilizzare il test dopo la data di scadenza riportata sulla confezione.
- Non utilizzare il test se la busta di alluminio è strappata o danneggiata.
- Monouso. Smaltire dopo il primo utilizzo.
- Il dispositivo di prova deve rimanere nella confezione sigillata fino al momento dell'uso.
- Un campione di urina contaminato o contaminato può dare risultati falsi.
- Tenere fuori dalla portata dei bambini.

【MATERIALI FORNITI】

- Pannello di test (sigillato in un sacchetto di alluminio con un essiccante)
- Foglio illustrativo

【MATERIALI RICHIESTI MA NON FORNITI】

- Orologio, orologio o timer
- Contenitore per la raccolta delle urine

【CONSERVAZIONE E STABILITÀ】

Conservare nella confezione originale sigillata a 2-30 °C fino alla data di scadenza. **Tenere lontano dalla luce solare diretta, dall'umidità e dal calore. NON CONGELARE.**

【ISTRUZIONI PER L'USO】

Lasciare che il test raggiunga la temperatura ambiente (15-30 °C) prima del test.
Lavatevi accuratamente le mani.

Passo 1: Raccogliere il campione di urina


- Raccogliere l'urina in un contenitore pulito e asciutto.

Passo 2: Test

- Estrarre il pannello di prova dalla busta sigillata e utilizzarlo immediatamente.
- Rimuovere il tappo del pannello.
- Immergere il pannello di test verticalmente nel campione di urina **per almeno 10-15 secondi.**
- **Nota:** immergere il pannello dell'astina di livello almeno fino al livello delle linee ondulate, ma non oltre la freccia sul pannello di prova.
- Riposizionare il tappo e posizionare il pannello di prova su una superficie piana e non assorbente.
- Avvia un timer.

Passo 3: leggere i risultati

- **Leggere i risultati del test dopo 5 minuti,** non leggere i risultati dopo 10 minuti.



【INTERPRETAZIONE DEI RISULTATI】

	**NEGATIVO: Appaiono due linee colorate.** Una riga appare accanto alla lettera "C" e l'altra riga appare accanto alla lettera "T".
	**POSITIVO: Accanto alla lettera "C" appare solo una linea colorata. Accanto alla lettera "T" non appare alcuna linea colorata.**
	**NON VALIDO: La linea di controllo non viene visualizzata.**

IMPORTANTE: il risultato ottenuto è definito preliminare per un motivo. Il campione deve essere testato in laboratorio per confermare il risultato.

【DOMANDE E RISPOSTE】

1. Come faccio a sapere se il test ha funzionato bene?
Quando appare la linea di controllo (C), significa che l'unità di prova funziona correttamente.

2. Quanto tempo ci vuole prima che io possa leggere i risultati?
È possibile leggere i risultati dopo 5 minuti, a condizione che sia comparsa una linea rossa o rosa accanto alla zona di controllo (C); non leggere i risultati dopo 10 minuti.

3. Qual è il momento migliore per eseguire il test?
Puoi eseguire il test in qualsiasi momento della giornata.

4. Come leggere il test se il colore e l'intensità delle linee sono diversi?
Il colore e l'intensità delle linee non hanno importanza per l'interpretazione del risultato. Il test dovrebbe essere considerato negativo qualunque sia l'intensità del colore della linea di test (T).

5. Che cosa è un test falso positivo?
Un risultato falso positivo del test significa che il farmaco non è presente ma risulta rilevato dal dispositivo. Le cause più comuni di un test falso positivo sono i cross-reattivi.

6. Cos'è un test falso negativo?
Un test falso negativo significa che il farmaco è presente ma non viene rilevato dal dispositivo. Se il campione è diluito o contaminato, ciò può causare un risultato falso negativo.

【LIMITAZIONI】

1. Il test è destinato esclusivamente all'uso con urina umana.
2. Il test fornisce solo un risultato preliminare.
3. Sostanze come candeggina e/o allume nei campioni di urina possono produrre risultati non corretti. Se si sospetta un'adulterazione, il test deve essere ripetuto con un altro campione di urina.
4. Il test non distingue tra droghe d'abuso e determinati farmaci.
5. Un risultato positivo al test può essere ottenuto con determinati alimenti o integratori alimentari.

Indice dei simboli

	Consultare le istruzioni per l'uso o consultare le istruzioni per l'uso elettroniche		Contiene sufficiente per <n> test	2°C 30°C	Limite di temperatura
	Produttore	LOT	Codice batch	REF	Catalogare numero
	Attenzione		Data di scadenza		Non riutilizzare
	Non utilizzare se la confezione è danneggiata e consultare le istruzioni per l'uso	EU REP	Rappresentante autorizzato europeo		

Hangzhou AllTest Biotech Co.,Ltd.
#550,Yinhai Street
Hangzhou Economic & Technological Development Area
Hangzhou, 310018 P.R. China
Web: www.alltests.com.cn Email: info@alltests.com.cn

EU REP
VidaQuick Biotech S.L.
No.132, Rosello Street, Barcelona,
Barcelona Province, 08036, Spain
E-mail: info@vidaquick.com

Importato da:
Verify SRL
Torino - Italy
info@verifytest.it
www.verifytest.it

Numero : 14603512700
Data di revisione : 2026-01-09

Multi-Drug Rapid Test Panel (Urine) Package Insert For Self-testing

A rapid test for the qualitative detection of multiple drugs and drug metabolites in human urine.
For self-testing, not for in vitro diagnostic use.

【INTENDED USE】

The Multi-Drug Rapid Test Panel is a rapid chromatographic immunoassay for the qualitative detection of multiple drugs and drug metabolites in urine at the following cut-off concentrations:

Test	Calibrator	Cut-off (ng/mL)
Amphetamine (AMP)	d-Amphetamine	1,000
Cocaine (COC)	Benzoylecgonine	300
Marijuana (THC)	11-nor-Δ^9-THC-9 COOH	50
Methamphetamine (MET)	d-Methamphetamine	1,000
Methylenedioxymethamphetamine (MDMA)	d,l-Methylenedioxymethamphetamine	500
Opiate (OPI)	Morphine	2,000
Ketamine (KET)	Ketamine	1,000

The test provides only preliminary test results. A more specific alternative chemical method must be used in order to obtain a confirmed analytical result. GC/MS or LC/MS is the preferred confirmatory method.

【SUMMARY】

The Multi-Drug Rapid Test Panel is a rapid urine screening test that can be performed without the use of an instrument. The test utilizes monoclonal antibodies to selectively detect elevated levels of specific drugs in urine.

【PRECAUTIONS】

- Read the entire package insert prior to performing test.
- For self-testing, not for *in vitro* diagnostic use.
- For external use only.
- Do not use the test after expiration date printed on the package.
- Do not use the test if its foil pouch is torn or damaged.
- For single use. Discard after first use.
- The test device should remain in the sealed pouch until use.
- Contaminated or tainted urine sample may give false results.
- Keep out of reach of children.

【MATERIALS PROVIDED】

- Test Panel (sealed in foil pouch with a desiccant)
- Package Insert

【MATERIALS REQUIRED BUT NOT PROVIDED】

- Watch, clock or timer
- Urine collection cup

【STORAGE AND STABILITY】

Store as packaged in the sealed pouch at 2-30 °C up to the date of expiration. **Keep away from direct sunlight, moisture and heat. DO NOT FREEZE**.

【DIRECTIONS FOR USE】

Allow the test to reach room temperature (15-30 °C) before testing.
Wash your hands thoroughly.

Step 1: Collect Urine Sample


- Collect urine in a clean and dry container.


Step 2: Testing

- Remove the test panel from the sealed pouch and use it immediately.
- Remove the cap of the panel.
- Immerse the test panel vertically in the urine specimen for at least **10 to 15 seconds**.
- **Note:** Immerse the dipstick panel to at least the level of the wavy lines, but not above the arrow on the test panel.
- Replace the cap and place the test panel on a non-absorbent flat surface.
- Start a timer.

Step 3: Read Results

- **Read the test results at 5 minutes,** do not read results after 10 minutes.





【INTERPRETATION OF RESULTS】

	**NEGATIVE: Two colored lines appear.** One line appears next to the letter 'C' and the other line appears next to the letter 'T'.
	**POSITIVE: Only one colored line appears next to the letter 'C'. No colored line appears next to the letter 'T'.**
	**INVALID: Control line fails to appear.**

IMPORTANT: The result you obtained is called preliminary for a reason. The sample must be tested by laboratory in order to confirm the result.

【Q&A】

1. How do I know if the test worked well?

When the control line(C) appears, it means that the test unit is working well.

2. How soon can I read my results?

You can read your results at 5 minutes as long as a colored line has appeared next to the Control region(C), do not read results after 10 minutes.

3. When is the best time to run the test?

You can run the test at any time of the day.

4. How to read the test if the color and the intensity of the lines are different?

The color and intensity of the lines have no importance for result interpretation. The test should be considered as negative whatever the color intensity of the test line (T) is.

5. What is a false positive test?

A false positive test result means the drug is not present but shows detected by the device. The most common causes of a false positive test are cross reactants.

6. What is a false negative test?

A false negative test means the drug is present but is not detected by the test. If the sample is diluted, or the sample is contaminated that may cause a false negative result.

【LIMITATIONS】

1. The test is for use with human urine only.
2. The test only provides a preliminary result.
3. Substances, such as bleach and/or alum, in urine specimens may produce incorrect results. If adulteration is suspected, the test should be repeated with another urine specimen.
4. Test does not distinguish between drugs of abuse and certain medications.
5. A positive test result may be obtained from certain foods or food supplements.

Index of Symbols

	Consult instructions for use or consult electronic instructions for use	Σ	Contains sufficient for <n> tests	2°C 30°C	Temperature limit
	Manufacturer	LOT	Batch code	REF	Catalogue number
	Caution		Use-by date		Do not re-use
	Do not use if package is damaged and consult instructions for use	EU REP	European Authorized Representative		

Hangzhou AllTest Biotech Co.,Ltd.
#550,Yinhai Street
Hangzhou Economic & Technological Development Area
Hangzhou, 310018 P.R. China
Web: www.alltests.com.cn Email: info@alltests.com.cn

EU REP
VidaQuick Biotech S.L.
No.132, Rosello Street, Barcelona,
Barcelona Province, 08036, Spain
E-mail: info@vidaquick.com

Imported by:
Verify SRL
Torino - Italy
info@verifytest.it
www.verifytest.it

Number: 14603512700
Revision date:2026-01-09

Multi-Drogen-Schnelltest-Panel
(Urin)
Packungsbeilage
Zum Selbsttest

Ein Schnelltest zum qualitativen Nachweis mehrerer Drogen und Drogenmetaboliten im menschlichen Urin.
Zur Selbsttestung, nicht für die In-vitro-Diagnostik.

【BESTIMMTER VERWENDUNGSZWECK】

Das Multi-Drogen-Schnelltest-panel ist ein schneller chromatographischer Immunoassay zum qualitativen Nachweis mehrerer Drogen und Drogenmetaboliten im Urin bei den folgenden Grenzwertkonzentrationen:

Test	Kalibrator	Abschaltung (ng/mL)
Amphetamin (AMP)	d-Amphetamin	1,000
Kokain (COC)	Benzoylecgonin	300
Marihuana (THC)	11-Nor-Δ9-THC-9 COOH	50
Methamphetamin (MET)	d-Methamphetamin	1,000
Methylendioxymethamphetamin (MDMA)	d,l-Methylendioxymethamphetamin	500
Opiat (OPI)	Morphin	2,000
Ketamin (KET)	Ketamin	1,000

Der Test liefert lediglich vorläufige Ergebnisse. Um ein bestätigtes Analyseergebnis zu erhalten, muss eine spezifischere chemische Alternativmethode angewendet werden. GC/MS oder LC/MS sind die bevorzugten Bestätigungsmethoden.

【ZUSAMMENFASSUNG】

Der Multi-Drug-Schnelltest ist ein Urin-Schnelltest, der ohne Gerät durchgeführt werden kann. Der Test nutzt monoklonale Antikörper, um erhöhte Konzentrationen bestimmter Drogen im Urin gezielt nachzuweisen.

【VORSICHTSMASSNAHMEN】

- Lesen Sie die gesamte Packungsbeilage, bevor Sie den Test durchführen.
- Für Zum Selbsttest, nicht für *die In-vitro*-Diagnostik.
- Nur zur äußerlichen Anwendung.
- Verwenden Sie den Test nicht mehr nach Ablauf des auf der Verpackung aufgedruckten Verfallsdatums.
- Verwenden Sie den Test nicht, wenn der Folienbeutel eingerissen oder beschädigt ist.
- Zum einmaligen Gebrauch. Nach dem ersten Gebrauch entsorgen..
- Das Testgerät sollte bis zur Verwendung im versiegelten Beutel verbleiben.
- Eine verunreinigte oder verfälschte Urinprobe kann zu falschen Ergebnissen führen.
- Außerhalb der Reichweite von Kindern aufbewahren.

【MITGESTELLTE MATERIALIEN】

- Testpanel (versiegelt in einem Folienbeutel mit Trockenmittel)
- Packungsbeilage

【BENÖTIGTE, ABER NICHT BEREITGESTELLTE MATERIALIEN】

- Uhr, Wecker oder Timer
- Urinauffangbecher

【LAGERUNG UND STABILITÄT】

Lagern Sie das Produkt in der verschlossenen Verpackung bei 2-30 °C bis zum Verfallsdatum. **Vor direkter Sonneneinstrahlung, Feuchtigkeit und Hitze schützen. Nicht einfrieren.**

【GEBRAUCHSANWEISUNG】

Lassen Sie den Test vor der Durchführung auf Raumtemperatur (15-30 °C) kommen.
Waschen Sie sich gründlich die Hände.

Schritt 1: Urinprobe entnehmen


- Sammeln Sie den Urin in einem sauberen und trockenen Behälter.

Schritt 2: Testen

- Nehmen Sie das Testpanel aus dem versiegelten Beutel und verwenden Sie es sofort.
- Entfernen Sie die Abdeckung der Abdeckung.
- Tauchen Sie das Testpanel mindestens **10 bis 15 Sekunden lang senkrecht in die Urinprobe ein.**
- **Hinweis:** Tauchen Sie den Peilstab mindestens bis zur Höhe der Wellenlinien ein, jedoch nicht über den Pfeil auf dem Testfeld hinaus.
- Setzen Sie die Kappe wieder auf und legen Sie das Testfeld auf eine nicht saugfähige, ebene Fläche.
- Starte einen Timer.

Schritt 3: Ergebnisse lesen

- **Lesen Sie die Testergebnisse nach 5 Minuten ab,** lesen Sie die Ergebnisse nicht mehr nach 10 Minuten ab.



【INTERPRETATION DER ERGEBNISSE】

	**NEGATIV: Es erscheinen zwei farbige Linien.** Neben dem Buchstaben 'C' erscheint ein Strich, neben dem Buchstaben 'T' ein anderer.
	**POSITIV: Neben dem Buchstaben 'C' erscheint nur eine farbige Linie. Neben dem Buchstaben 'T' erscheint keine farbige Linie.**
	**UNGÜLTIG: Die Steuerlinie wird nicht angezeigt.**

WICHTIG: Das Ergebnis ist aus gutem Grund vorläufig. Die Probe muss im Labor untersucht werden, um das Ergebnis zu bestätigen.

【FRAGEN UND ANTWORTEN】

1. Woher weiß ich, ob die Hat der Test gut funktioniert?
Wenn die Kontrolllinie (C) erscheint, bedeutet dies, dass das Testgerät ordnungsgemäß funktioniert.

2. Wie schnell kann ich meine Ergebnisse einsehen?
Sie können Ihre Ergebnisse nach 5 Minuten ablesen, sobald neben dem Kontrollbereich (C) eine farbige Linie erscheint. Lesen Sie die Ergebnisse nicht nach 10 Minuten ab.

3. Wann ist der beste Zeitpunkt, um den Test durchzuführen?
Sie können den Test zu jeder Tageszeit durchführen.

4. Wie ist der Test zu interpretieren, wenn Farbe und Intensität der Linien unterschiedlich sind?
Farbe und Intensität der Linien sind für die Ergebnisinterpretation irrelevant. Der Test ist unabhängig von der Farbintensität der Testlinie (T) als negativ zu werten.

5. Was ist ein/e falsch positiver Test ?
Ein falsch positives Testergebnis bedeutet, dass der Wirkstoff nicht vorhanden ist, aber vom Gerät fälschlicherweise als nachgewiesen angezeigt wird. Die häufigste Ursache für ein falsch positives Testergebnis sind Kreuzreaktionen.

6. Was ist ein/e Falsch-negatives Testergebnis ? Ein falsch-negatives
Testergebnis bedeutet, dass die Droge vorhanden ist, aber vom Test nicht erkannt wird. Eine Verdünnung oder Verunreinigung der Probe kann zu einem falsch-negativen Ergebnis führen.

【EINSCHRÄNKUNGEN】

1. Der Test ist nur zur Verwendung mit menschlichem Urin bestimmt.
2. Der Test liefert lediglich ein vorläufiges Ergebnis.
3. Substanzen wie Bleichmittel und/oder Alaun in Urinproben können zu falschen Ergebnissen führen. Bei Verdacht auf Verfälschung sollte der Test mit einer weiteren Urinprobe wiederholt werden.
4. Der Test unterscheidet nicht zwischen Drogenmissbrauch und bestimmten Medikamenten.
5. Bestimmte Lebensmittel oder Nahrungsergänzungsmittel können zu einem positiven Testergebnis führen.

Symbolverzeichnis

	Beachten Sie die Gebrauchsanweisung oder die elektronische Gebrauchsanweisung.	Σ	Enthält ausreichend für <n> Tests	2°C 30°C	Temperaturgrenze
	Hersteller	LOT	Batch-Code	REF	Katalog Nummer
	Vorsicht		Mindesthaltbarkeitsdatum		Nicht wiederverwenden
	Nicht verwenden, wenn die Verpackung beschädigt ist. Gebrauchsanweisung beachten.	EU REP	Europäischer Bevollmächtigter		

Hangzhou AllTest Biotech Co.,Ltd.
#550,Yinhai Street
Hangzhou Economic & Technological Development Area
Hangzhou, 310018 P.R. China
Web: www.alltests.com.cn Email: info@alltests.com.cn

EU REP
VidaQuick Biotech S.L.
No.132, Rosello Street, Barcelona,
Barcelona Province, 08036, Spain
E-mail: info@vidaquick.com

Importiert von:
Verify SRL
Torino-Italy
info@verifytest.it
www.verifytest.it

Nummer: 14603512700
Revisionsdatum: 2026-01-09